

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072623\
 Data File : BF134498.D
 Acq On : 27 Jul 2023 00:00
 Operator : CG\JU
 Sample : 03668-01DL 5X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jul 27 01:59:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 01:46:04 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	51160	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	192215	20.000	ng	#-0.01
39) Acenaphthene-d10	9.869	164	100614	20.000	ng	0.00
64) Phenanthrene-d10	11.363	188	169369	20.000	ng	# 0.00
76) Chrysene-d12	14.010	240	125494	20.000	ng	-0.02
86) Perylene-d12	15.504	264	137694	20.000	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.457	112	59702	19.280	ng	-0.01
7) Phenol-d6	6.469	99	72067	18.996	ng	-0.01
23) Nitrobenzene-d5	7.387	82	45595	12.001	ng	-0.02
42) 2,4,6-Tribromophenol	10.669	330	15298	10.701	ng	0.00
45) 2-Fluorobiphenyl	9.181	172	90255	11.776	ng	-0.01
79) Terphenyl-d14	12.945	244	81674	11.361	ng	-0.02
Target Compounds						Qvalue
4) n-Nitrosodimethylamine	3.299	42	5664	3.251	ng	# 9
9) Benzaldehyde	6.381	77	11258	5.541	ng	# 1
18) Hexachloroethane	7.351	117	65646	49.222	ng	# 1
19) n-Nitroso-di-n-propyla...	7.240	70	10354	4.384	ng	# 72
22) Acetophenone	7.216	105	166057	34.737	ng	# 50
24) Nitrobenzene	7.387	77	11671	3.089	ng	# 31
25) Isophorone	7.628	82	21568	3.406	ng	# 1
31) Naphthalene	8.128	128	32085	3.354	ng	# 42
35) Caprolactam	8.557	113	4923	5.873	ng	# 1
37) 2-Methylnaphthalene	8.822	142	236006	35.507	ng	97
38) 1-Methylnaphthalene	8.922	142	219089	35.530	ng	94
41) Hexachlorocyclopentadiene	8.804	237	77	6.856	ng	# 31
46) 1,1'-Biphenyl	9.286	154	24821	2.993	ng	89
49) Acenaphthylene	9.728	152	24888	2.410	ng	# 1
51) 2,6-Dinitrotoluene	9.634	165	4031	2.463	ng	88
52) Acenaphthene	9.904	154	39614	6.493	ng	# 65
53) 3-Nitroaniline	9.845	138	4295	2.289	ng	# 53
54) 2,4-Dinitrophenol	9.939	184	771	7.393	ng	# 1
55) Dibenzofuran	10.075	168	34164	3.638	ng	# 1
56) 4-Nitrophenol	10.016	139	8835	8.829	ng	# 47
57) 2,4-Dinitrotoluene	10.075	165	22505	10.238	ng	# 62
58) Fluorene	10.422	166	109133	14.406	ng	# 94
66) n-Nitrosodiphenylamine	10.528	169	152656	28.330	ng	# 48
69) Atrazine	11.075	200	4576	3.046	ng	# 52
71) Phenanthrene	11.392	178	296426	33.497	ng	# 96
72) Anthracene	11.445	178	29863	3.270	ng	# 20
78) Pyrene	12.810	202	61480	6.161	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ClientSampleId :


Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.828 min Scan# 807
 Delta R.T. -0.005 min
 Lab File: BF134498.D
 Acq: 27 Jul 2023 00:00


Instrument :
 BNA_F
 ClientSampleId :





Tgt Ion:152 Resp: 51160
 Ion Ratio Lower Upper
 152 100
 150 155.1 128.3 192.5
 115 59.8 52.4 78.6




#4
 n-Nitrosodimethylamine
 Concen: 3.251 ng
 RT: 3.299 min Scan# 206
 Delta R.T. -0.035 min
 Lab File: BF134498.D
 Acq: 27 Jul 2023 00:00



Tgt Ion: 42 Resp: 5664
 Ion Ratio Lower Upper
 42 100
 74 0.0 72.2 108.4#
 44 16.0 5.4 8.0#









#5
2-Fluorophenol
Concen: 19.280 ng
RT: 5.457 min Scan# 51
Delta R.T. -0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:112 Resp: 59702
Ion Ratio Lower Upper
112 100
64 66.5 44.9 67.3
63 38.8 25.4 38.0#



#7
Phenol-d6
Concen: 18.996 ng
RT: 6.469 min Scan# 745
Delta R.T. -0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion: 99 Resp: 72067
Ion Ratio Lower Upper
99 100
42 26.7 18.7 28.1
71 45.7 36.4 54.6









#9
Benzaldehyde
Concen: 5.541 ng
RT: 6.381 min Scan# 71
Delta R.T. -0.011 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00




Instrument :
BNA_F
ClientSampleId :


Tgt Ion: 77 Resp: 11258
Ion Ratio Lower Upper
77 100
105 760.0 77.1 117.1#
106 61.1 71.4 111.4#



#18
Hexachloroethane
Concen: 49.222 ng
RT: 7.351 min Scan# 895
Delta R.T. 0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion: 117 Resp: 65646
Ion Ratio Lower Upper
117 100
119 495.9 75.8 113.8#
201 0.0 73.6 110.4#





#19

n-Nitroso-di-n-propylamine

Concen: 4.384 ng

RT: 7.240 min Scan# 81

Delta R.T. -0.005 min


Lab File: BF134498.D

Acq: 27 Jul 2023 00:00

Instrument :

BNA_F

ClientSampleId :



Tgt Ion: 70 Resp: 10354




Ion Ratio Lower Upper

70 100

42 49.2 54.7 82.1#

101 0.0 8.2 12.2#

130 0.0 15.1 22.7#



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.104 min Scan# 1023

Delta R.T. -0.012 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00

Tgt Ion: 136 Resp: 192215




Ion Ratio Lower Upper


136 100

137 11.5 8.4 12.6

54 11.5 5.2 7.8#


68 8.2 2.9 4.3#








#22
Acetophenone
Concen: 34.737 ng
RT: 7.216 min Scan# 81
Delta R.T. -0.029 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:105 Resp: 166057
Ion Ratio Lower Upper
105 100
71 6.0 7.4 11.0#
51 5.6 36.4 54.6#
120 0.3 16.6 25.0#



#23
Nitrobenzene-d5
Concen: 12.001 ng
RT: 7.387 min Scan# 901
Delta R.T. -0.017 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion: 82 Resp: 45595
Ion Ratio Lower Upper
82 100
128 43.3 34.3 51.5
54 54.5 43.8 65.6









#24
 Nitrobenzene
 Concen: 3.089 ng
 RT: 7.387 min Scan# 907
 Delta R.T. -0.035 min
 Lab File: BF134498.D
 Acq: 27 Jul 2023 00:00




Instrument :
 BNA_F
 ClientSampleId :


Tgt Ion: 77 Resp: 11671
 Ion Ratio Lower Upper
 77 100
 123 21.3 35.5 53.3#
 65 85.3 11.9 17.9#



#25
 Isophorone
 Concen: 3.406 ng
 RT: 7.628 min Scan# 942
 Delta R.T. -0.029 min
 Lab File: BF134498.D
 Acq: 27 Jul 2023 00:00

Tgt Ion: 82 Resp: 21568
 Ion Ratio Lower Upper
 82 100
 95 168.9 6.0 9.0#
 138 10.4 14.6 22.0#









#31
Naphthalene
Concen: 3.354 ng
RT: 8.128 min Scan# 1100
Delta R.T. -0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:128 Resp: 32085
Ion Ratio Lower Upper
128 100
129 44.2 8.4 12.6#
127 27.0 10.7 16.1#



#35
Caprolactam
Concen: 5.873 ng
RT: 8.557 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion:113 Resp: 4923
Ion Ratio Lower Upper
113 100
55 1063.6 161.5 201.5#
56 407.8 117.8 157.8#





#37

2-Methylnaphthalene

Concen: 35.507 ng

RT: 8.822 min Scan# 1146

Delta R.T. -0.005 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00

Instrument :

BNA_F





ClientSampleId :

Tgt Ion:142 Resp: 236006

Ion Ratio Lower Upper

142 100

141 91.1 70.8 106.2



#38

1-Methylnaphthalene

Concen: 35.530 ng

RT: 8.922 min Scan# 1162

Delta R.T. -0.005 min

Lab File: BF134498.D




Acq: 27 Jul 2023 00:00


Tgt Ion:142 Resp: 219089

Ion Ratio Lower Upper

142 100


141 94.9 71.4 107.2








#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.869 min Scan# 1142
Delta R.T. -0.006 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:164 Resp: 100614
Ion Ratio Lower Upper
164 100
162 103.6 82.6 123.8
160 50.7 37.2 55.8




#41
Hexachlorocyclopentadiene
Concen: 6.856 ng
RT: 8.804 min Scan# 1142
Delta R.T. -0.171 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00




Tgt Ion:237 Resp: 77
Ion Ratio Lower Upper
237 100
235 0.0 37.0 77.0#
272 0.0 0.0 32.7








#42
2,4,6-Tribromophenol
Concen: 10.701 ng
RT: 10.669 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00


Instrument :
BNA_F
ClientSampleId :





Tgt Ion:330 Resp: 15298
Ion Ratio Lower Upper
330 100
332 96.3 78.8 118.2
141 41.8 25.4 38.2#




#45
2-Fluorobiphenyl
Concen: 11.776 ng
RT: 9.181 min Scan# 1206
Delta R.T. -0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00



Tgt Ion:172 Resp: 90255
Ion Ratio Lower Upper
172 100
171 35.1 28.8 43.2
170 22.7 18.7 28.1









#46
1,1'-Biphenyl
Concen: 2.993 ng
RT: 9.286 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:154 Resp: 24821
Ion Ratio Lower Upper
154 100
153 46.8 20.8 60.8
76 16.5 0.0 30.2



#49
Acenaphthylene
Concen: 2.410 ng
RT: 9.728 min Scan# 1299
Delta R.T. -0.005 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion:152 Resp: 24888
Ion Ratio Lower Upper
152 100
151 59.7 16.4 24.6#
153 96.0 10.9 16.3#





#51

2,6-Dinitrotoluene

Concen: 2.463 ng

RT: 9.634 min Scan# 1

Delta R.T. -0.029 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:165 Resp: 4031

Ion Ratio Lower Upper

165 100

63 79.5 57.8 86.6

89 52.3 51.8 77.8



#52

Acenaphthene

Concen: 6.493 ng

RT: 9.904 min Scan# 1329

Delta R.T. -0.006 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00




Tgt Ion:154 Resp: 39614


Ion Ratio Lower Upper

154 100

153 139.0 90.0 135.0#

152 93.0 42.4 63.6#









#53
3-Nitroaniline
Concen: 2.289 ng
RT: 9.845 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00




Instrument :
BNA_F
ClientSampleId :


Tgt Ion:138 Resp: 4295
Ion Ratio Lower Upper
138 100
108 78.0 9.3 13.9#
92 72.9 86.6 129.8#



#54
2,4-Dinitrophenol
Concen: 7.393 ng
RT: 9.939 min Scan# 1335
Delta R.T. -0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00


Tgt Ion:184 Resp: 771
Ion Ratio Lower Upper
184 100
63 976.7 57.1 85.7#
154 813.7 59.1 88.7#








#55
Dibenzenofuran
Concen: 3.638 ng
RT: 10.075 min Scan# 1358
Delta R.T. -0.005 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:168 Resp: 34164
Ion Ratio Lower Upper
168 100
139 110.5 33.4 50.2#
169 241.1 10.6 15.8#



#56
4-Nitrophenol
Concen: 8.829 ng
RT: 10.016 min Scan# 1348
Delta R.T. 0.012 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00


Tgt Ion:139 Resp: 8835
Ion Ratio Lower Upper
139 100
109 160.7 76.9 116.9#
65 70.0 98.5 138.5#








#57
2,4-Dinitrotoluene
Concen: 10.238 ng
RT: 10.075 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:165 Resp: 22505
Ion Ratio Lower Upper
165 100
63 86.9 48.9 73.3#
89 40.9 61.7 92.5#
182 10.8 2.0 3.0#



#58
Fluorene
Concen: 14.406 ng
RT: 10.422 min Scan# 1417
Delta R.T. -0.005 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion:166 Resp: 109133
Ion Ratio Lower Upper
166 100
165 96.7 74.5 111.7
167 21.7 10.6 16.0#





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.363 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00

Instrument :

BNA_F

ClientSampleId :





Tgt Ion:188 Resp: 169369

Ion Ratio Lower Upper

188 100

94 9.7 5.1 7.7#

80 11.3 6.6 9.8#



#66

n-Nitrosodiphenylamine

Concen: 28.330 ng

RT: 10.528 min Scan# 1435

Delta R.T. -0.012 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00




Tgt Ion:169 Resp: 152656


Ion Ratio Lower Upper

169 100

168 19.7 52.6 78.8#

167 12.2 29.8 44.8#









#69
 Atrazine
 Concen: 3.046 ng
 RT: 11.075 min Scan# 1528
 Delta R.T. 0.006 min
 Lab File: BF134498.D
 Acq: 27 Jul 2023 00:00




Instrument :
 BNA_F
 ClientSampleId :


Tgt Ion:200 Resp: 4576
 Ion Ratio Lower Upper
 200 100
 173 90.4 6.1 46.1#
 215 55.1 29.7 69.7



#71
 Phenanthrene
 Concen: 33.497 ng
 RT: 11.392 min Scan# 1582
 Delta R.T. 0.000 min
 Lab File: BF134498.D
 Acq: 27 Jul 2023 00:00


Tgt Ion:178 Resp: 296426
 Ion Ratio Lower Upper
 178 100
 176 20.0 15.8 23.6
 179 19.4 12.7 19.1#








#72
Anthracene
Concen: 3.270 ng
RT: 11.445 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Instrument :
BNA_F
ClientSampleId :







Tgt Ion:178 Resp: 29863
Ion Ratio Lower Upper
178 100
176 19.9 14.6 21.8
179 86.7 12.3 18.5#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.010 min Scan# 2027
Delta R.T. -0.023 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Tgt Ion:240 Resp: 125494
Ion Ratio Lower Upper
240 100
120 8.0 6.6 10.0
236 24.1 20.2 30.2





#78

Pyrene

Concen: 6.161 ng

RT: 12.810 min Scan# 1823

Delta R.T. -0.012 min

Lab File: BF134498.D





Acq: 27 Jul 2023 00:00

Instrument :

BNA_F

ClientSampleId :

Tgt Ion:	202	Resp:	61480
Ion Ratio	Lower	Upper	
202	100		
200	19.4	15.8	23.8
203	20.7	13.5	20.3



#79

Terphenyl-d14

Concen: 11.361 ng




RT: 12.945 min Scan# 1846


Delta R.T. -0.017 min

Lab File: BF134498.D

Acq: 27 Jul 2023 00:00


Tgt Ion:	244	Resp:	81674
Ion Ratio	Lower	Upper	
244	100		
212	8.1	5.6	8.4
122	10.5	8.0	12.0





#86
Perylene-d12
Concen: 20.000 ng
RT: 15.504 min Scan# 21
Delta R.T. -0.029 min
Lab File: BF134498.D
Acq: 27 Jul 2023 00:00

Instrument :
BNA_F
ClientSampleId :



Tgt Ion:264 Resp: 137694

Ion	Ratio	Lower	Upper
264	100		
260	21.6	18.4	27.6
265	20.5	15.8	23.8

