

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072623\
 Data File : BF134495.D
 Acq On : 26 Jul 2023 22:23
 Operator : CG\JU
 Sample : 03738-04
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 80002

Quant Time: Jul 27 01:57:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 01:46:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	84084	20.000	ng	-0.01
21) Naphthalene-d8	8.098	136	319346	20.000	ng	#-0.02
39) Acenaphthene-d10	9.857	164	132059	20.000	ng	-0.02
64) Phenanthrene-d10	11.357	188	200327	20.000	ng	#-0.01
76) Chrysene-d12	14.039	240	214567	20.000	ng	# 0.00
86) Perylene-d12	15.527	264	243468	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.457	112	435364	85.546	ng	-0.01
7) Phenol-d6	6.463	99	521443	83.627	ng	-0.02
23) Nitrobenzene-d5	7.387	82	351770	55.731	ng	-0.02
42) 2,4,6-Tribromophenol	10.657	330	146580	78.121	ng	-0.01
45) 2-Fluorobiphenyl	9.175	172	702833	69.868	ng	-0.02
79) Terphenyl-d14	12.969	244	465581	37.878	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.575	163	33543	3.426	ng	Qvalue 100
84) Bis(2-ethylhexyl)phtha...	13.998	149	29423	3.544	ng	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072623\
Data File : BF134495.D
Acq On : 26 Jul 2023 22:23
Operator : CG\JU
Sample : 03738-04
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
80002

Quant Time: Jul 27 01:57:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 01:46:04 2023
Response via : Initial Calibration

