

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072718\
 Data File : BF107780.D
 Acq On : 28 Jul 2018 7:23
 Operator : JU/SJ
 Sample : J4184-01
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-3B

Quant Time: Jul 30 03:14:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF072018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 20 16:29:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.97	152	153452	20.00	ng	-0.01
21) Naphthalene-d8	8.25	136	478576	20.00	ng	0.00
38) Acenaphthene-d10	10.00	164	220765	20.00	ng	-0.02
63) Phenanthrene-d10	11.50	188	402297	20.00	ng	-0.02
75) Chrysene-d12	14.15	240	429169	20.00	ng	-0.02
86) Perylene-d12	15.68	264	337639	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	498822	52.26	ng	-0.01
7) Phenol-d6	6.60	99	379629	33.13	ng	-0.01
23) Nitrobenzene-d5	7.53	82	896929	126.48	ng	-0.02
41) 2,4,6-Tribromophenol	10.80	330	349725	139.31	ng	-0.02
44) 2-Fluorobiphenyl	9.32	172	1652646	167.80	ng	-0.02
78) Terphenyl-d14	13.08	244	1479144	88.23	ng	-0.02
Target Compounds						
27) 2,4-Dimethylphenol	7.91	122	29495	4.543	ng	90
31) Naphthalene	8.30	128	8371615	397.796	ng	# 90
37) 2-Methylnaphthalene	8.95	142	103658	7.108	ng	97
45) 1,1'-Biphenyl	9.42	154	320876	16.685	ng	96
48) Acenaphthylene	9.86	152	109384	4.954	ng	# 93
49) Dimethylphthalate	9.71	163	61914	3.721	ng	98
51) Acenaphthene	10.04	154	721477	52.147	ng	100
57) Fluorene	10.55	166	175614	12.085	ng	93
70) Phenanthrene	11.52	178	226251	11.541	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072718\
Data File : BF107780.D
Acq On : 28 Jul 2018 7:23
Operator : JU/SJ
Sample : J4184-01
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-3B

Quant Time: Jul 30 03:14:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF072018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 20 16:29:27 2018
Response via : Initial Calibration

