

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072719\
 Data File : BF115825.D
 Acq On : 27 Jul 2019 12:42
 Operator : HP/JU
 Sample : K4011-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CONCRETE-PIPE

Manual Integrations
 APPROVED

mohammad
 7/29/2019 5:29:47 PM

Quant Time: Jul 29 01:54:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 25 13:16:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	141281	20.00	ng	0.00
21) Naphthalene-d8	8.15	136	538960	20.00	ng	0.00
39) Acenaphthene-d10	9.90	164	296371	20.00	ng	0.00
64) Phenanthrene-d10	11.39	188	520256	20.00	ng	0.00
76) Chrysene-d12	14.03	240	389770	20.00	ng	0.00
87) Perylene-d12	15.50	264	376430	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	930532	107.73	ng	0.00
7) Phenol-d6	6.50	99	1273030	116.16	ng	0.00
23) Nitrobenzene-d5	7.43	82	834824	90.07	ng	0.00
42) 2,4,6-Tribromophenol	10.69	330	282300	81.60	ng	0.00
45) 2-Fluorobiphenyl	9.23	172	1518162	89.66	ng	0.00
79) Terphenyl-d14	12.97	244	1726254	81.72	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.86	142	68008	3.931	ng	Qvalue 100
38) 1-Methylnaphthalene	8.96	142	46409m	2.824	ng	
50) Dimethylphthalate	9.62	163	185155	8.304	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF072719\
Data File : BF115825.D
Acq On : 27 Jul 2019 12:42
Operator : HP/JU
Sample : K4011-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CONCRETE-PIPE

Manual Integrations
APPROVED

mohammad
7/29/2019 5:29:47 PM

Quant Time: Jul 29 01:54:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF071219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 25 13:16:07 2019
Response via : Initial Calibration

