

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072723\
 Data File : BF134508.D
 Acq On : 27 Jul 2023 16:33
 Operator : CG\JU
 Sample : 03732-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CONC-WC-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/28/2023
 Supervised By :mohammad ahmed 07/28/2023

Quant Time: Jul 27 17:11:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 27 12:16:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.822	152	88080	20.000	ng	0.00
21) Naphthalene-d8	8.104	136	278164	20.000	ng	0.00
39) Acenaphthene-d10	9.857	164	142432	20.000	ng	0.00
64) Phenanthrene-d10	11.357	188	212136	20.000	ng	0.00
76) Chrysene-d12	14.015	240	164371	20.000	ng	0.00
86) Perylene-d12	15.515	264	132622	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	375963	70.523	ng	0.00
7) Phenol-d6	6.469	99	466204	71.376	ng	0.00
23) Nitrobenzene-d5	7.386	82	318523	57.935	ng	0.00
42) 2,4,6-Tribromophenol	10.657	330	102900	50.847	ng	0.00
45) 2-Fluorobiphenyl	9.180	172	596615	54.990	ng	0.00
79) Terphenyl-d14	12.951	244	554923	58.934	ng	0.00
Target Compounds						
						Qvalue
50) Dimethylphthalate	9.575	163	27513	2.606	ng	98
75) Fluoranthene	12.574	202	48147	3.727	ng	98
78) Pyrene	12.804	202	42659	3.264	ng	96
88) Benzo(b)fluoranthene	15.074	252	23790m	2.867	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072723\
Data File : BF134508.D
Acq On : 27 Jul 2023 16:33
Operator : CG\JU
Sample : 03732-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CONC-WC-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/28/2023
Supervised By :mohammad ahmed 07/28/2023

Quant Time: Jul 27 17:11:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 27 12:16:03 2023
Response via : Initial Calibration

