

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072823\
 Data File : BF134531.D
 Acq On : 29 Jul 2023 00:10
 Operator : CG\JU
 Sample : 03645-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-07-(1-3)

Quant Time: Jul 29 01:29:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 27 12:16:03 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.828	152	99614	20.000	ng	0.00
21) Naphthalene-d8	8.110	136	368068	20.000	ng	# 0.00
39) Acenaphthene-d10	9.863	164	164170	20.000	ng	0.00
64) Phenanthrene-d10	11.357	188	266517	20.000	ng	# 0.00
76) Chrysene-d12	14.016	240	176024	20.000	ng	# 0.00
86) Perylene-d12	15.515	264	152338	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	480258	79.655	ng	0.01
7) Phenol-d6	6.475	99	574170	77.727	ng	0.00
23) Nitrobenzene-d5	7.393	82	379320	52.141	ng	0.00
42) 2,4,6-Tribromophenol	10.663	330	153935	65.994	ng	0.00
45) 2-Fluorobiphenyl	9.187	172	663691	53.072	ng	0.00
79) Terphenyl-d14	12.951	244	577061	57.228	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072823\
Data File : BF134531.D
Acq On : 29 Jul 2023 00:10
Operator : CG\JU
Sample : 03645-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-07-(1-3)

Quant Time: Jul 29 01:29:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 27 12:16:03 2023
Response via : Initial Calibration

