

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF073018\
 Data File : BF107801.D
 Acq On : 30 Jul 2018 15:49
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 30 18:21:32 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.515	0.532	-3.3	101	0.00
3	Pyridine	1.258	1.208	4.0	96	0.00
4	n-Nitrosodimethylamine	0.599	0.519	13.4	94	0.00
5 S	2-Fluorophenol	1.177	1.043	11.4	85	0.00
6	Aniline	1.613	1.568	2.8	94	0.00
7 S	Phenol-d6	1.551	1.418	8.6	98	0.00
8	2-Chlorophenol	1.478	1.380	6.6	98	0.00
9	Benzaldehyde	0.936	0.859	8.2	96	0.00
10 C	Phenol	1.850	1.481	19.9	96	0.00
11	bis(2-Chloroethyl)ether	1.645	1.626	1.2	105	0.00
12	1,3-Dichlorobenzene	1.550	1.465	5.5	97	0.00
13 C	1,4-Dichlorobenzene	1.726	1.678	2.8	98	0.00
14	1,2-Dichlorobenzene	1.602	1.493	6.8	99	0.00
15	Benzyl Alcohol	0.964	0.940	2.5	96	0.00
16	2,2'-oxybis(1-Chloropropane	2.365	2.173	8.1	96	0.00
17	2-Methylphenol	1.091	1.051	3.7	96	0.00
18	Hexachloroethane	0.620	0.564	9.0	97	0.00
19 P	n-Nitroso-di-n-propylamine	0.984	0.959	2.5	98	0.00
20	3+4-Methylphenols	1.340	1.377	-2.8	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.481	0.439	8.7	98	0.00
23 S	Nitrobenzene-d5	0.347	0.329	5.2	101	0.00
24	Nitrobenzene	0.364	0.340	6.6	98	0.00
25	Isophorone	0.570	0.531	6.8	95	0.00
26 C	2-Nitrophenol	0.183	0.180	1.6	97	0.00
27	2,4-Dimethylphenol	0.245	0.231	5.7	99	0.00
28	bis(2-Chloroethoxy)methane	0.376	0.370	1.6	97	0.00
29 C	2,4-Dichlorophenol	0.277	0.273	1.4	98	0.00
30	1,2,4-Trichlorobenzene	0.316	0.305	3.5	98	0.00
31	Naphthalene	0.930	0.912	1.9	96	0.00
32	Benzoic acid	0.157	0.131	16.6	76	0.00
33	4-Chloroaniline	0.397	0.399	-0.5	102	0.00
34 C	Hexachlorobutadiene	0.195	0.182	6.7	98	0.00
35	Caprolactam	0.083	0.081	2.4	99	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.301	-0.3	98	0.00
37	2-Methylnaphthalene	0.584	0.589	-0.9	98	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
39	1,2,4,5-Tetrachlorobenzene	0.682	0.663	2.8	97	0.00
40 P	Hexachlorocyclopentadiene	0.243	0.240	1.2	96	0.00
41 S	2,4,6-Tribromophenol	0.272	0.265	2.6	99	0.00
42 C	2,4,6-Trichlorophenol	0.377	0.384	-1.9	98	0.00
43	2,4,5-Trichlorophenol	0.454	0.456	-0.4	97	0.00
44 S	2-Fluorobiphenyl	1.435	1.384	3.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF073018\
 Data File : BF107801.D
 Acq On : 30 Jul 2018 15:49
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 30 18:21:32 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.820	1.763	3.1	98	0.00
46	2-Chloronaphthalene	1.373	1.328	3.3	98	0.00
47	2-Nitroaniline	0.446	0.438	1.8	98	0.00
48	Acenaphthylene	2.128	1.981	6.9	99	0.00
49	Dimethylphthalate	1.502	1.465	2.5	100	0.00
50	2,6-Dinitrotoluene	0.326	0.323	0.9	102	0.00
51 C	Acenaphthene	1.236	1.129	8.7	100	0.00
52	3-Nitroaniline	0.411	0.396	3.6	98	0.00
53 P	2,4-Dinitrophenol	0.140	0.138	1.4	95	0.00
54	Dibenzofuran	1.911	1.795	6.1	98	0.00
55 P	4-Nitrophenol	0.150	0.147	2.0	89	0.00
56	2,4-Dinitrotoluene	0.429	0.404	5.8	95	0.00
57	Fluorene	1.485	1.387	6.6	99	0.00
58	2,3,4,6-Tetrachlorophenol	0.417	0.403	3.4	93	0.00
59	Diethylphthalate	1.615	1.524	5.6	100	0.00
60	4-Chlorophenyl-phenylether	0.809	0.762	5.8	100	0.00
61	4-Nitroaniline	0.365	0.356	2.5	101	0.00
62	Azobenzene	1.504	1.407	6.4	99	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
64	4,6-Dinitro-2-methylphenol	0.109	0.110	-0.9	102	0.00
65 c	n-Nitrosodiphenylamine	0.659	0.616	6.5	100	0.00
66	4-Bromophenyl-phenylether	0.233	0.210	9.9	98	0.00
67	Hexachlorobenzene	0.258	0.243	5.8	99	0.00
68	Atrazine	0.195	0.189	3.1	101	0.00
69 C	Pentachlorophenol	0.127	0.128	-0.8	99	0.00
70	Phenanthrene	0.950	0.887	6.6	98	0.00
71	Anthracene	1.095	1.004	8.3	98	0.00
72	Carbazole	1.082	1.032	4.6	100	0.00
73	Di-n-butylphthalate	1.181	1.089	7.8	98	0.00
74 C	Fluoranthene	1.116	1.026	8.1	98	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
76	Benzidine	0.647	0.633	2.2	98	0.00
77	Pyrene	1.440	1.349	6.3	100	0.00
78 S	Terphenyl-d14	0.922	0.868	5.9	102	0.00
79	Butylbenzylphthalate	0.631	0.598	5.2	98	0.00
80	Benzo(a)anthracene	1.133	1.048	7.5	96	0.00
81	3,3'-Dichlorobenzidine	0.439	0.408	7.1	98	0.00
82	Chrysene	1.222	1.150	5.9	101	0.00
83	Bis(2-ethylhexyl)phthalate	0.793	0.742	6.4	99	0.00
84 c	Di-n-octyl phthalate	1.273	1.215	4.6	99	0.00
85	Indeno(1,2,3-cd)pyrene	0.929	0.776	16.5	93	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Benzo(b)fluoranthene	0.986	0.956	3.0	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF073018\
Data File : BF107801.D
Acq On : 30 Jul 2018 15:49
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jul 30 18:21:32 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 17:48:53 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.211	1.113	8.1	95	0.00
89 C	Benzo(a)pyrene	1.018	0.928	8.8	99	0.00
90	Dibenzo(a,h)anthracene	0.888	0.770	13.3	96	0.00
91	Benzo(a,h,i)perylene	0.859	0.725	15.6	93	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0