


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073018\
 Data File : BF107800.D
 Acq On : 30 Jul 2018 15:22
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 30 17:48:53 2018



AutoFind: Scans 1595, 1596, 1597; Background Corrected with Scan 1590

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.5	68690	PASS
68	69	0.00	2	1.2	673	PASS
69	198	0.00	100	34.8	57626	PASS
70	69	0.00	2	0.6	344	PASS
127	198	10	80	47.5	78525	PASS
197	198	0.00	2	0.6	1000	PASS
198	198	100	100	100.0	165424	PASS
199	198	5	9	6.9	11337	PASS
275	198	10	60	28.1	46501	PASS
365	198	1	100	3.4	5691	PASS
441	198	0.01	100	15.3	25263	PASS
442	198	50	100	97.7	161576	PASS
443	442	15	24	19.7	31810	PASS