


# Hexachlorocyclopentadiene

Response Ratio




Resp Ratio =  $2.59 \times 10^{-1} \times \text{Amt} - 7.68 \times 10^{-2}$   
Coef of Det ( $r^2$ ) = 0.999    Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF073019.M  
Calibration Table Last Updated: Wed Jul 31 15:31:51 2019

# 2,4-Dinitrophenol

Response Ratio



Resp Ratio =  $1.67 \times 10^{-1} \times \text{Amt} - 6.52 \times 10^{-2}$   
Coef of Det ( $r^2$ ) = 0.998    Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF073019.M  
Calibration Table Last Updated: Wed Jul 31 15:31:51 2019