

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF080218\
 Data File : BF107923.D
 Acq On : 2 Aug 2018 21:33
 Operator : JU/SJ
 Sample : J4285-05MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD

Quant Time: Aug 03 00:57:20 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	133135	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	513931	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	188846	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	334296	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	333614	20.00	ng	-0.01
86) Perylene-d12	15.68	264	287092	20.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.60	112	724699	92.51	ng	0.00
7) Phenol-d6	6.61	99	926646	89.75	ng	0.00
23) Nitrobenzene-d5	7.52	82	593514	66.53	ng	-0.01
41) 2,4,6-Tribromophenol	10.80	330	215951	84.16	ng	0.00
44) 2-Fluorobiphenyl	9.31	172	1094891	80.79	ng	-0.01
78) Terphenyl-d14	13.07	244	1005086	65.33	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.85	88	90048	27.337	ng	# 81
3) Pyridine	3.66	79	203934	25.542	ng	87
4) n-Nitrosodimethylamine	3.64	42	33529	8.412	ng	# 96
6) Aniline	6.64	93	264679	24.648	ng	# 54
8) 2-Chlorophenol	6.75	128	312941	31.803	ng	92
9) Benzaldehyde	6.52	77	70778	11.363	ng	89
10) Phenol	6.62	94	410177	33.316	ng	88
11) bis(2-Chloroethyl)ether	6.69	93	339922	31.049	ng	97
12) 1,3-Dichlorobenzene	6.90	146	316445	30.662	ng	97
13) 1,4-Dichlorobenzene	6.97	146	369698	32.180	ng	98
14) 1,2-Dichlorobenzene	7.12	146	322393	30.231	ng	98
15) Benzyl Alcohol	7.11	79	188027	29.301	ng	93
16) 2,2'-oxybis(1-Chloropropan	7.22	45	498690	31.683	ng	65
17) 2-Methylphenol	7.22	107	232251	31.973	ng	# 72
18) Hexachloroethane	7.45	117	118780	28.764	ng	94
19) n-Nitroso-di-n-propylamine	7.36	70	209733	32.024	ng	100
20) 3+4-Methylphenols	7.38	107	271815	30.479	ng	# 65
22) Acetophenone	7.37	105	446078	36.077	ng	# 95
24) Nitrobenzene	7.54	77	276641	29.574	ng	97
25) Isophorone	7.77	82	571977	39.060	ng	99
26) 2-Nitrophenol	7.86	139	105836	22.545	ng	98
27) 2,4-Dimethylphenol	7.90	122	252692	40.204	ng	98
28) bis(2-Chloroethoxy)methane	7.98	93	356632	36.956	ng	99
29) 2,4-Dichlorophenol	8.11	162	238300	33.519	ng	98
30) 1,2,4-Trichlorobenzene	8.18	180	271229	33.442	ng	98
31) Naphthalene	8.26	128	887925	37.168	ng	100
32) Benzoic acid	7.90	122	252692	59.047	ng	# 14
33) 4-Chloroaniline	8.32	127	220280	21.587	ng	97
34) Hexachlorobutadiene	8.36	225	163820	32.686	ng	98
35) Caprolactam	8.67	113	61350	28.665	ng	# 74
36) 4-Chloro-3-methylphenol	8.81	107	229934	29.780	ng	97
37) 2-Methylnaphthalene	8.95	142	534028	35.587	ng	99
39) 1,2,4,5-Tetrachlorobenzene	9.11	216	258273	40.113	ng	# 96
40) Hexachlorocyclopentadiene	9.09	237	38839	20.129	ng	97


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF080218\
 Data File : BF107923.D
 Acq On : 2 Aug 2018 21:33
 Operator : JU/SJ
 Sample : J4285-05MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD

Quant Time: Aug 03 00:57:20 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.24	196	131731	37.008	ng	99
43) 2,4,5-Trichlorophenol	9.29	196	151387	35.328	ng	95
45) 1,1'-Biphenyl	9.41	154	671821	39.091	ng	97
46) 2-Chloronaphthalene	9.44	162	482802	37.253	ng	97
47) 2-Nitroaniline	9.55	65	143281	33.989	ng	88
48) Acenaphthylene	9.86	152	752665	37.461	ng	99
49) Dimethylphthalate	9.71	163	683224	48.174	ng	100
50) 2,6-Dinitrotoluene	9.78	165	108055	35.059	ng	89
51) Acenaphthene	10.02	154	433090	37.107	ng	100
52) 3-Nitroaniline	9.98	138	100445	25.863	ng	95
53) 2,4-Dinitrophenol	10.15	184	6260	11.887	ng #	23
54) Dibenzofuran	10.20	168	607629	33.683	ng	97
55) 4-Nitrophenol	10.20	139	362690	184.636	ng #	19
56) 2,4-Dinitrotoluene	10.20	165	124153	30.648	ng #	82
57) Fluorene	10.55	166	478648	34.141	ng	95
58) 2,3,4,6-Tetrachlorophenol	10.32	232	126065	32.029	ng #	77
59) Diethylphthalate	10.40	149	540437	35.442	ng	100
60) 4-Chlorophenyl-phenylether	10.53	204	239941	31.417	ng	93
61) 4-Nitroaniline	10.62	138	90038	26.123	ng	89
62) Azobenzene	10.70	77	452154	31.838	ng	94
64) 4,6-Dinitro-2-methylphenol	10.61	198	11221	6.170	ng #	44
65) n-Nitrosodiphenylamine	10.65	169	411052	37.343	ng	99
66) 4-Bromophenyl-phenylether	11.02	248	140319	36.013	ng	89
67) Hexachlorobenzene	11.09	284	143667	33.251	ng #	90
68) Atrazine	11.18	200	116091	35.634	ng	95
69) Pentachlorophenol	11.30	266	119669	56.262	ng	99
70) Phenanthrene	11.51	178	576689	36.331	ng	100
71) Anthracene	11.57	178	713866	38.990	ng	99
72) Carbazole	11.74	167	644463	35.619	ng	97
73) Di-n-butylphthalate	12.03	149	738099	37.393	ng	99
74) Fluoranthene	12.71	202	745276	39.965	ng	97
76) Benzidine	12.85	184	519724	48.182	ng	98
77) Pyrene	12.94	202	751557	31.285	ng	97
79) Butylbenzylphthalate	13.54	149	335341	31.877	ng	95
80) Benzo(a)anthracene	14.13	228	678791	35.919	ng	99
81) 3,3'-Dichlorobenzidine	14.10	252	257942	35.211	ng #	97
82) Chrysene	14.17	228	764055	37.484	ng	98
83) Bis(2-ethylhexyl)phthalate	14.09	149	463918	35.079	ng	98
84) Di-n-octyl phthalate	14.71	149	860526	40.538	ng	97
85) Indeno(1,2,3-cd)pyrene	17.25	276	498837	32.174	ng #	92
87) Benzo(b)fluoranthene	15.21	252	530933	37.497	ng	98
88) Benzo(k)fluoranthene	15.25	252	681447	39.199	ng	98
89) Benzo(a)pyrene	15.61	252	542637	37.123	ng	99
90) Dibenzo(a,h)anthracene	17.26	278	427556	33.532	ng	96
91) Benzo(g,h,i)perylene	17.73	276	400381	32.453	ng #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.96 min Scan# 786
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD

Tgt Ion:152 Resp: 133135
Ion Ratio Lower Upper
152 100
150 156.0 124.6 186.8
115 56.3 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 27.337 ng
RT: 2.85 min Scan# 87
Delta R.T. 0.03 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion: 88 Resp: 90048
Ion Ratio Lower Upper
88 100
58 100.3 62.6 93.8#
43 29.7 24.6 37.0



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1





#3

Pyridine

Concen: 25.542 ng

RT: 3.66 min Scan# 225

Delta R.T. 0.05 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion: 79 Resp: 203934

Ion Ratio Lower Upper




79 100

52 87.4 59.8 89.6

51 43.5 29.8 44.8



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 8.412 ng

RT: 3.64 min Scan# 221

Delta R.T. 0.05 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion: 42 Resp: 33529

Ion Ratio Lower Upper



42 100


74 135.7 104.9 157.3

44 11.2 6.9 10.3#



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1









#5
 2-Fluorophenol
 Concen: 92.512 ng
 RT: 5.60 min Scan# 555
 Delta R.T. 0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD

Tgt Ion: 112 Resp: 724699
 Ion Ratio Lower Upper
 112 100
 64 63.4 47.9 71.9
 63 32.2 23.7 35.5




Abundance Ion 112.00 (111.70 to 112.70): BF1
 Ion 64.00 (63.70 to 64.70): BF1
 Ion 63.00 (62.70 to 63.70): BF1






#6
 Aniline
 Concen: 24.648 ng
 RT: 6.64 min Scan# 731
 Delta R.T. 0.00 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion: 93 Resp: 264679
 Ion Ratio Lower Upper
 93 100
 66 69.6 32.2 48.2#
 65 40.9 16.4 24.6#



Abundance Ion 93.00 (92.70 to 93.70): BF1
 Ion 66.00 (65.70 to 66.70): BF1
 Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 89.755 ng

RT: 6.61 min Scan# 726

Delta R.T. 0.00 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion: 99 Resp: 926646

Ion Ratio Lower Upper

99 100

42 18.1 14.5 21.7




71 33.3 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 31.803 ng

RT: 6.75 min Scan# 750

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion: 128 Resp: 312941

Ion Ratio Lower Upper

128 100

130 31.5 13.0 53.0



64 56.9 29.4 69.4




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 11.363 ng

RT: 6.52 min Scan# 712

Delta R.T. 0.00 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion: 77 Resp: 70778

Ion Ratio Lower Upper

77 100

105 111.5 81.1 121.1




106 105.6 74.5 114.5



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 33.316 ng

RT: 6.62 min Scan# 728

Delta R.T. 0.00 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion: 94 Resp: 410177

Ion Ratio Lower Upper

94 100

65 31.7 7.9 47.9



66 45.6 16.2 56.2




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#11
bis(2-Chloroethyl)ether
Concen: 31.049 ng
RT: 6.69 min Scan# 740
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion: 93 Resp: 339922
Ion Ratio Lower Upper
93 100
63 81.6 58.6 98.6
95 29.8 11.8 51.8



#12
1,3-Dichlorobenzene
Concen: 30.662 ng
RT: 6.90 min Scan# 775
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion: 146 Resp: 316445
Ion Ratio Lower Upper
146 100
148 62.4 51.8 77.8
75 31.1 24.2 36.4





#13

1,4-Dichlorobenzene

Concen: 32.180 ng

RT: 6.97 min Scan# 788

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:146 Resp: 369698

Ion Ratio Lower Upper

146 100

148 61.6 50.8 76.2

111 42.4 34.2 51.2






Abundance

Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E



#14

1,2-Dichlorobenzene

Concen: 30.231 ng

RT: 7.12 min Scan# 814

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:146 Resp: 322393

Ion Ratio Lower Upper

146 100

148 65.1 50.6 75.8

111 44.0 36.0 54.0






Abundance

Ion 146.00 (145.70 to 146.70): E

Ion 148.00 (147.70 to 148.70): E

Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 29.301 ng

RT: 7.11 min Scan# 811

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD





Tgt Ion: 79 Resp: 188027

Ion Ratio Lower Upper

79 100

108 74.9 65.1 97.7

77 69.1 50.6 75.8



#16

2,2'-oxybis(1-Chloropropane)

Concen: 31.683 ng

RT: 7.22 min Scan# 830

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33




Tgt Ion: 45 Resp: 498690


Ion Ratio Lower Upper

45 100

77 25.7 0.0 32.9

79 23.5 0.0 29.6









#17
2-Methylphenol
Concen: 31.973 ng
RT: 7.22 min Scan# 830
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion	Ratio	Lower	Upper
107	100		
108	111.9	91.7	137.5
77	85.5	33.8	50.8#
79	78.4	33.8	50.8#



#18
Hexachloroethane
Concen: 28.764 ng
RT: 7.45 min Scan# 870
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion	Ratio	Lower	Upper
117	100		
119	93.1	75.8	113.8
201	85.1	75.7	113.5









#19
n-Nitroso-di-n-propylamine
Concen: 32.024 ng
RT: 7.36 min Scan# 854
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion:	70	Resp:	209733
Ion	Ratio	Lower	Upper
70	100		
42	59.1	47.5	71.3
101	9.2	7.4	11.2
130	20.7	16.6	25.0



#20
3+4-Methylphenols
Concen: 30.479 ng
RT: 7.38 min Scan# 857
Delta R.T. 0.00 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion:	107	Resp:	271815
Ion	Ratio	Lower	Upper
107	100		
108	90.2	68.2	108.2
77	124.0	26.7	66.7#
79	28.3	6.3	46.3





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.24 min Scan# 1003

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD

Tgt Ion:136 Resp: 513931



Ion Ratio Lower Upper

136 100

137 10.8 8.9 13.3

54 9.5 6.6 9.8

68 5.9 5.0 7.4





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 36.077 ng

RT: 7.37 min Scan# 855

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Tgt Ion:105 Resp: 446078



Ion Ratio Lower Upper

105 100

71 8.3 4.4 6.6#

51 32.2 22.3 33.5

120 20.9 16.9 25.3





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 66.531 ng
 RT: 7.52 min Scan# 882
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion	Ratio	Lower	Upper
82	100		
128	46.6	36.1	54.1
54	60.3	41.4	62.2



#24
 Nitrobenzene
 Concen: 29.574 ng
 RT: 7.54 min Scan# 885
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion	Ratio	Lower	Upper
77	100		
123	45.1	37.9	56.9
65	14.1	11.0	16.4





#25

Isophorone

Concen: 39.060 ng

RT: 7.77 min Scan# 924

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion: 82 Resp: 571977

Ion Ratio Lower Upper

82 100

95 6.5 5.2 7.8




138 19.0 14.9 22.3



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 22.545 ng

RT: 7.86 min Scan# 939

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion: 139 Resp: 105836

Ion Ratio Lower Upper

139 100

109 26.8 22.7 34.1



65 51.1 40.5 60.7




Abundance Ion 139.00 (138.70 to 139.70): BF1

Ion 109.00 (108.70 to 109.70): BF1

Ion 65.00 (64.70 to 65.70): BF1









#27
 2,4-Dimethylphenol
 Concen: 40.204 ng
 RT: 7.90 min Scan# 945
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:122 Resp: 252692
 Ion Ratio Lower Upper
 122 100
 107 116.1 91.2 136.8
 121 57.9 47.2 70.8



#28
 bis(2-Chloroethoxy)methane
 Concen: 36.956 ng
 RT: 7.98 min Scan# 959
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion: 93 Resp: 356632
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.3 39.5
 123 11.1 9.8 14.6









#29
2,4-Dichlorophenol
Concen: 33.519 ng
RT: 8.11 min Scan# 981
Delta R.T. 0.00 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion	Ratio	Lower	Upper
162	100		
164	61.8	42.7	82.7
98	36.5	18.1	58.1



#30
1,2,4-Trichlorobenzene
Concen: 33.442 ng
RT: 8.18 min Scan# 993
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion	Ratio	Lower	Upper
180	100		
182	95.6	76.8	115.2
145	29.8	26.5	39.7









#31
Naphthalene
Concen: 37.168 ng
RT: 8.26 min Scan# 1007
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	13.6	10.9	16.3




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 59.047 ng
RT: 7.90 min Scan# 945
Delta R.T. -0.14 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion	Ratio	Lower	Upper
122	100		
105	4.2	101.1	141.1#
77	33.4	70.7	110.7#



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 21.587 ng
RT: 8.32 min Scan# 1018
Delta R.T. 0.00 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD

Tgt Ion:	127	Resp:	220280
Ion	Ratio	Lower	Upper
127	100		
129	34.7	25.9	38.9
65	32.6	25.0	37.4
92	18.6	15.2	22.8




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BF1
Ion 92.00 (91.70 to 92.70): BF1






#34
Hexachlorobutadiene
Concen: 32.686 ng
RT: 8.36 min Scan# 1024
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion:	225	Resp:	163820
Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.6	76.0
227	62.4	51.9	77.9



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 28.665 ng
 RT: 8.67 min Scan# 1077
 Delta R.T. -0.04 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:113 Resp: 61350
 Ion Ratio Lower Upper
 113 100
 55 226.3 163.3 203.3#
 56 160.3 115.7 155.7#



#36
 4-Chloro-3-methylphenol
 Concen: 29.780 ng
 RT: 8.81 min Scan# 1100
 Delta R.T. 0.00 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:107 Resp: 229934
 Ion Ratio Lower Upper
 107 100
 144 26.0 20.5 30.7
 142 80.7 62.0 93.0









#37
2-Methylnaphthalene
Concen: 35.587 ng
RT: 8.95 min Scan# 1124
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion:142 Resp: 534028
Ion Ratio Lower Upper
142 100
141 89.7 70.8 106.2
115 31.6 26.1 39.1



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.00 min Scan# 1302
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion:164 Resp: 188846
Ion Ratio Lower Upper
164 100
162 106.0 86.1 129.1
160 48.2 38.3 57.5





#39

1,2,4,5-Tetrachlorobenzene

Concen: 40.113 ng

RT: 9.11 min Scan# 1152

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD

Tgt Ion:216 Resp: 258273


Ion Ratio Lower Upper

216 100

214 78.1 63.0 94.4

179 19.8 18.1 27.1

108 14.8 17.2 25.8#






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 20.129 ng

RT: 9.09 min Scan# 1148

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:237 Resp: 38839

Ion Ratio Lower Upper

237 100

235 62.7 44.8 84.8



272 12.5 0.0 33.3




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 84.157 ng
 RT: 10.80 min Scan# 1438
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:330 Resp: 215951
 Ion Ratio Lower Upper
 330 100
 332 97.6 77.0 115.6
 141 28.3 29.9 44.9#



#42
 2,4,6-Trichlorophenol
 Concen: 37.008 ng
 RT: 9.24 min Scan# 1173
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:196 Resp: 131731
 Ion Ratio Lower Upper
 196 100
 198 94.9 75.7 113.5
 200 28.7 24.5 36.7





#43

2,4,5-Trichlorophenol

Concen: 35.328 ng

RT: 9.29 min Scan# 1182

Delta R.T. 0.00 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD

Tgt Ion:196 Resp: 151387


Ion Ratio Lower Upper

196 100

198 95.2 74.6 112.0

97 46.1 42.9 64.3

132 30.2 26.3 39.5






Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 97.00 (96.70 to 97.70): BF1

Ion 132.00 (131.70 to 132.70): E



#44

2-Fluorobiphenyl

Concen: 80.792 ng

RT: 9.31 min Scan# 1185

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:172 Resp: 1094891

Ion Ratio Lower Upper

172 100

171 36.9 29.7 44.5



170 24.2 19.6 29.4




Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 39.091 ng
 RT: 9.41 min Scan# 1203
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD

Tgt Ion:154 Resp: 671821
 Ion Ratio Lower Upper
 154 100
 153 40.0 21.3 61.3
 76 11.8 0.0 34.0




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF1






#46
 2-Chloronaphthalene
 Concen: 37.253 ng
 RT: 9.44 min Scan# 1208
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:162 Resp: 482802
 Ion Ratio Lower Upper
 162 100
 127 39.6 33.9 50.9
 164 32.5 26.5 39.7



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
 2-Nitroaniline
 Concen: 33.989 ng
 RT: 9.55 min Scan# 1227
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion: 65 Resp: 143281
 Ion Ratio Lower Upper
 65 100
 92 58.1 55.8 83.6
 138 102.8 91.2 136.8



#48
 Acenaphthylene
 Concen: 37.461 ng
 RT: 9.86 min Scan# 1279
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion: 152 Resp: 752665
 Ion Ratio Lower Upper
 152 100
 151 19.8 16.3 24.5
 153 13.0 10.7 16.1









#49
Dimethylphthalate
Concen: 48.174 ng
RT: 9.71 min Scan# 1253
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD

Tgt Ion:163 Resp: 683224
Ion Ratio Lower Upper
163 100
194 3.7 2.7 4.1
164 10.1 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#50
2,6-Dinitrotoluene
Concen: 35.059 ng
RT: 9.78 min Scan# 1266
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion:165 Resp: 108055
Ion Ratio Lower Upper
165 100
63 65.7 44.7 67.1
89 48.8 44.0 66.0



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1





#51

Acenaphthene

Concen: 37.107 ng

RT: 10.02 min Scan# 1307

Delta R.T. -0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:154 Resp: 433090

Ion Ratio Lower Upper

154 100

153 114.1 91.2 136.8




152 54.7 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 25.863 ng

RT: 9.98 min Scan# 1299

Delta R.T. 0.00 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:138 Resp: 100445

Ion Ratio Lower Upper

138 100

108 12.0 9.4 14.0



92 106.0 89.4 134.2




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF1









#53
2,4-Dinitrophenol
Concen: 11.887 ng
RT: 10.15 min Scan# 1328
Delta R.T. 0.06 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD

Tgt Ion:184 Resp: 6260
Ion Ratio Lower Upper
184 100
63 282.5 54.2 81.2#
154 193.9 184.0 276.0




Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E






#54
Dibenzofuran
Concen: 33.683 ng
RT: 10.20 min Scan# 1337
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion:168 Resp: 607629
Ion Ratio Lower Upper
168 100
139 40.2 30.1 45.1
169 13.5 10.9 16.3



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E





#55

4-Nitrophenol

Concen: 184.636 ng

RT: 10.20 min Scan# 1337

Delta R.T. 0.04 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:139 Resp: 362690

Ion Ratio Lower Upper

139 100

109 6.6 48.4 88.4#




65 12.1 72.6 112.6#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1



#56

2,4-Dinitrotoluene

Concen: 30.648 ng

RT: 10.20 min Scan# 1336

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Tgt Ion:165 Resp: 124153


Ion Ratio Lower Upper

165 100

63 73.1 36.4 54.6#

89 75.4 57.8 86.6

182 2.2 1.6 2.4






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E









#57
 Fluorene
 Concen: 34.141 ng
 RT: 10.55 min Scan# 1396
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:166 Resp: 478648
 Ion Ratio Lower Upper
 166 100
 165 93.7 79.8 119.8
 167 12.9 10.6 16.0



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 32.029 ng
 RT: 10.32 min Scan# 1358
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:232 Resp: 126065
 Ion Ratio Lower Upper
 232 100
 131 35.6 43.8 65.8#
 130 1.7 2.1 3.1#
 166 23.7 27.8 41.6#









#59
Diethylphthalate
Concen: 35.442 ng
RT: 10.40 min Scan# 1371
Delta R.T. -0.02 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion:149 Resp: 540437
Ion Ratio Lower Upper
149 100
177 20.2 16.1 24.1
150 12.7 10.1 15.1



#60
4-Chlorophenyl-phenylether
Concen: 31.417 ng
RT: 10.53 min Scan# 1393
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion:204 Resp: 239941
Ion Ratio Lower Upper
204 100
206 33.6 26.5 39.7
141 59.6 54.6 81.8





#61

4-Nitroaniline

Concen: 26.123 ng

RT: 10.62 min Scan# 1408

Delta R.T. 0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:138 Resp: 90038

Ion Ratio Lower Upper

138 100

92 49.2 30.2 70.2




108 58.0 52.5 92.5



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF1

Ion 108.00 (107.70 to 108.70): E



#62

Azobenzene

Concen: 31.838 ng

RT: 10.70 min Scan# 1421

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Tgt Ion: 77 Resp: 452154


Ion Ratio Lower Upper

77 100

182 25.7 1.7 41.7

105 25.1 3.0 43.0

51 32.3 9.9 49.9






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF1









#63
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.49 min Scan# 1556
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:188 Resp: 334296
 Ion Ratio Lower Upper
 188 100
 94 9.4 8.5 12.7
 80 10.7 9.2 13.8



#64
 4,6-Dinitro-2-methylphenol
 Concen: 6.170 ng
 RT: 10.61 min Scan# 1406
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:198 Resp: 11221
 Ion Ratio Lower Upper
 198 100
 51 127.2 37.7 77.7#
 105 68.8 36.3 76.3





#65

n-Nitrosodiphenylamine

Concen: 37.343 ng

RT: 10.65 min Scan# 1414

Delta R.T. -0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:169 Resp: 411052

Ion Ratio Lower Upper

169 100

168 67.1 54.4 81.6




167 36.3 29.8 44.6



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 36.013 ng

RT: 11.02 min Scan# 1477

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:248 Resp: 140319

Ion Ratio Lower Upper

248 100

250 100.0 76.9 115.3



141 74.5 74.4 111.6




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 33.251 ng

RT: 11.09 min Scan# 1488

Delta R.T. -0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:284 Resp: 143667

Ion Ratio Lower Upper

284 100

142 33.7 34.6 52.0#




249 29.0 24.8 37.2



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 35.634 ng

RT: 11.18 min Scan# 1503

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:200 Resp: 116091

Ion Ratio Lower Upper

200 100

173 29.5 8.6 48.6



215 49.6 25.1 65.1




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 56.262 ng
 RT: 11.30 min Scan# 1523
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:266 Resp: 119669
 Ion Ratio Lower Upper
 266 100
 268 62.6 51.1 76.7
 264 62.8 49.5 74.3



#70
 Phenanthrene
 Concen: 36.331 ng
 RT: 11.51 min Scan# 1560
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:178 Resp: 576689
 Ion Ratio Lower Upper
 178 100
 176 19.8 15.8 23.8
 179 15.9 13.0 19.6









#71
 Anthracene
 Concen: 38.990 ng
 RT: 11.57 min Scan# 1569
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:178 Resp: 713866
 Ion Ratio Lower Upper
 178 100
 176 18.8 15.4 23.2
 179 15.2 12.6 19.0



#72
 Carbazole
 Concen: 35.619 ng
 RT: 11.74 min Scan# 1598
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:167 Resp: 644463
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.6 26.4
 139 11.6 10.9 16.3





#73

Di-n-butylphthalate

Concen: 37.393 ng

RT: 12.03 min Scan# 1648

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD





Tgt Ion:149 Resp: 738099

Ion Ratio Lower Upper

149 100

150 9.2 7.8 11.6

104 4.4 3.8 5.8



#74

Fluoranthene

Concen: 39.965 ng

RT: 12.71 min Scan# 1763

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33




Tgt Ion:202 Resp: 745276


Ion Ratio Lower Upper

202 100

101 11.8 0.0 34.1

203 18.1 0.0 38.1





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.14 min Scan# 2007

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:240 Resp: 333614

Ion Ratio Lower Upper

240 100

120 11.8 9.5 14.3




236 25.0 20.7 31.1



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 48.182 ng

RT: 12.85 min Scan# 1787

Delta R.T. 0.00 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:184 Resp: 519724

Ion Ratio Lower Upper

184 100

185 14.5 12.6 18.8



183 11.0 9.3 13.9




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 31.285 ng

RT: 12.94 min Scan# 1803

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:202 Resp: 751557

Ion Ratio Lower Upper

202 100

200 20.0 17.4 26.2




203 17.4 14.3 21.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 65.329 ng

RT: 13.07 min Scan# 1825

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:244 Resp: 1005086

Ion Ratio Lower Upper

244 100

212 6.5 5.4 8.0



122 10.6 11.0 16.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 31.877 ng

RT: 13.54 min Scan# 1904

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:149 Resp: 335341

Ion Ratio Lower Upper

149 100

91 67.1 56.8 85.2




206 19.2 14.1 21.1



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BF1

Ion 206.00 (205.70 to 206.70): E



#80

Benzo(a)anthracene

Concen: 35.919 ng

RT: 14.13 min Scan# 2005

Delta R.T. -0.01 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:228 Resp: 678791

Ion Ratio Lower Upper

228 100

226 28.1 22.6 33.8



229 20.4 15.8 23.6




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 35.211 ng
 RT: 14.10 min Scan# 1999
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:252 Resp: 257942
 Ion Ratio Lower Upper
 252 100
 254 64.7 50.4 75.6
 126 11.2 11.3 16.9#



#82
 Chrysene
 Concen: 37.484 ng
 RT: 14.17 min Scan# 2012
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:228 Resp: 764055
 Ion Ratio Lower Upper
 228 100
 226 29.7 25.0 37.6
 229 19.4 16.2 24.4









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 35.079 ng
 RT: 14.09 min Scan# 1998
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:149 Resp: 463918
 Ion Ratio Lower Upper
 149 100
 167 27.5 21.3 31.9
 279 4.2 3.0 4.4



#84
 Di-n-octyl phthalate
 Concen: 40.538 ng
 RT: 14.71 min Scan# 2103
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:149 Resp: 860526
 Ion Ratio Lower Upper
 149 100
 167 1.8 1.2 1.8
 43 9.2 8.4 12.6









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 32.174 ng
 RT: 17.25 min Scan# 2536
 Delta R.T. -0.02 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33




Instrument :
 BNA_F
 ClientSampleId :
 KG-PIT-2MSD


Tgt Ion:276 Resp: 498837
 Ion Ratio Lower Upper
 276 100
 138 23.9 25.0 37.6#
 277 24.9 20.1 30.1



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.68 min Scan# 2268
 Delta R.T. -0.01 min
 Lab File: BF107923.D
 Acq: 2 Aug 2018 21:33

Tgt Ion:264 Resp: 287092
 Ion Ratio Lower Upper
 264 100
 260 23.6 19.2 28.8
 265 21.7 17.5 26.3









#87
Benzo(b)fluoranthene
Concen: 37.497 ng
RT: 15.21 min Scan# 2189
Delta R.T. -0.02 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33




Instrument :
BNA_F
ClientSampleId :
KG-PIT-2MSD


Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.0	25.6
125	10.5	9.0	13.6



#88
Benzo(k)fluoranthene
Concen: 39.199 ng
RT: 15.25 min Scan# 2195
Delta R.T. -0.01 min
Lab File: BF107923.D
Acq: 2 Aug 2018 21:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.9	26.9
125	11.0	10.2	15.2





#89

Benzo(a)pyrene

Concen: 37.123 ng

RT: 15.61 min Scan# 2256

Delta R.T. -0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD


Tgt Ion:252 Resp: 542637

Ion Ratio Lower Upper

252 100

253 22.1 17.1 25.7




125 12.4 9.8 14.6



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 33.532 ng

RT: 17.26 min Scan# 2537

Delta R.T. -0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33


Tgt Ion:278 Resp: 427556

Ion Ratio Lower Upper

278 100

139 17.0 15.9 23.9



279 23.0 19.0 28.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 32.453 ng

RT: 17.73 min Scan# 2617

Delta R.T. -0.02 min

Lab File: BF107923.D

Acq: 2 Aug 2018 21:33

Instrument :

BNA_F

ClientSampleId :

KG-PIT-2MSD



Tgt Ion: 276 Resp: 400381

Ion Ratio Lower Upper

276 100

277 24.1 17.9 26.9

138 20.0 21.8 32.8#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

