

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080218\
 Data File : BF107917.D
 Acq On : 2 Aug 2018 18:50
 Operator : JU/SJ
 Sample : J4267-11
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 UST-N

Quant Time: Aug 03 02:26:59 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	146818	20.00	ng	-0.01
21) Naphthalene-d8	8.24	136	600723	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	226627	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	411616	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	395592	20.00	ng	-0.01
86) Perylene-d12	15.68	264	303143	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	965403	111.75	ng	0.00
7) Phenol-d6	6.60	99	1139674	100.10	ng	0.00
23) Nitrobenzene-d5	7.52	82	756328	72.53	ng	-0.01
41) 2,4,6-Tribromophenol	10.79	330	271948	88.31	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	1351197	83.08	ng	-0.01
78) Terphenyl-d14	13.07	244	1203257	65.96	ng	-0.01
Target Compounds						
10) Phenol	6.62	94	52747	3.885	ng	# 31
49) Dimethylphthalate	9.70	163	150047	8.816	ng	# 99
74) Fluoranthene	12.71	202	50885	2.216	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080218\
Data File : BF107917.D
Acq On : 2 Aug 2018 18:50
Operator : JU/SJ
Sample : J4267-11
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
UST-N

Quant Time: Aug 03 02:26:59 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 17:48:53 2018
Response via : Initial Calibration

