

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080218\
 Data File : BF107926.D
 Acq On : 2 Aug 2018 22:56
 Operator : JU/SJ
 Sample : J4284-02 100X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OILY-WATER

Quant Time: Aug 03 04:45:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	84263	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	552941	20.00	ng	-0.01
38) Acenaphthene-d10	9.99	164	214274	20.00	ng	-0.02
63) Phenanthrene-d10	11.49	188	397729	20.00	ng	-0.01
75) Chrysene-d12	14.15	240	336560	20.00	ng	0.00
86) Perylene-d12	15.68	264	271605	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0	0.00	ng	
7) Phenol-d6	0.00	99	0	0.00	ng	
23) Nitrobenzene-d5	0.00	82	0	0.00	ng	
41) 2,4,6-Tribromophenol	0.00	330	0	0.00	ng	
44) 2-Fluorobiphenyl	9.32	172	1080	0.07	ng	0.00
78) Terphenyl-d14	13.08	244	3163	0.20	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080218\
Data File : BF107926.D
Acq On : 2 Aug 2018 22:56
Operator : JU/SJ
Sample : J4284-02 100X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OILY-WATER

Quant Time: Aug 03 04:45:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 17:48:53 2018
Response via : Initial Calibration

