

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080218\
 Data File : BF107930.D
 Acq On : 3 Aug 2018 00:45
 Operator : JU/SJ
 Sample : J4287-01
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 A4742

Quant Time: Aug 03 05:18:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 17:48:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	126021	20.00	ng	0.00
21) Naphthalene-d8	8.24	136	519571	20.00	ng	-0.01
38) Acenaphthene-d10	10.00	164	209917	20.00	ng	-0.01
63) Phenanthrene-d10	11.49	188	378725	20.00	ng	-0.01
75) Chrysene-d12	14.14	240	359361	20.00	ng	-0.01
86) Perylene-d12	15.68	264	284611	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.60	112	305113	41.15	ng	0.00
7) Phenol-d6	6.61	99	281885	28.84	ng	0.00
23) Nitrobenzene-d5	7.52	82	710560	78.79	ng	-0.01
41) 2,4,6-Tribromophenol	10.79	330	267053	93.63	ng	-0.01
44) 2-Fluorobiphenyl	9.31	172	1275591	84.68	ng	-0.01
78) Terphenyl-d14	13.07	244	1215082	73.32	ng	-0.01
Target Compounds						
10) Phenol	6.62	94	76867	6.596	ng	84
49) Dimethylphthalate	9.70	163	102666	6.512	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080218\
Data File : BF107930.D
Acq On : 3 Aug 2018 00:45
Operator : JU/SJ
Sample : J4287-01
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
A4742

Quant Time: Aug 03 05:18:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 17:48:53 2018
Response via : Initial Calibration

