


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080218\
 Data File : BF107905.D
 Acq On : 2 Aug 2018 13:19
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jul 30 17:48:53 2018



AutoFind: Scans 1594, 1595, 1596; Background Corrected with Scan 1589

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	41402	PASS
68	69	0.00	2	1.3	456	PASS
69	198	0.00	100	32.6	34724	PASS
70	69	0.00	2	0.3	102	PASS
127	198	10	80	46.1	49120	PASS
197	198	0.00	2	0.5	544	PASS
198	198	100	100	100.0	106573	PASS
199	198	5	9	6.6	7028	PASS
275	198	10	60	28.0	29848	PASS
365	198	1	100	3.5	3721	PASS
441	198	0.01	100	15.8	16865	PASS
442	198	50	100	98.8	105280	PASS
443	442	15	24	18.9	19869	PASS