

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/3/2023	BNA_F	<u>BF134599.D</u>
Compound Name	Response	Retention Time
DDT	650317	13.492
DDD	24207	13.186
DDE	265	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
24472	674789	3.63