

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080423\
 Data File : BF134625.D
 Acq On : 04 Aug 2023 11:11
 Operator : CG\JU
 Sample : 03756-04
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RPTP-10-CS-10

Quant Time: Aug 05 00:04:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	164443	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	662704	20.000	ng	# 0.00
39) Acenaphthene-d10	9.828	164	324017	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	521720	20.000	ng	# 0.00
76) Chrysene-d12	13.963	240	326868	20.000	ng	#-0.01
86) Perylene-d12	15.433	264	402203	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	877933	81.156	ng	0.00
7) Phenol-d6	6.428	99	1122259	81.155	ng	-0.01
23) Nitrobenzene-d5	7.357	82	642421	60.580	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	265358	80.337	ng	-0.01
45) 2-Fluorobiphenyl	9.151	172	1222442	62.730	ng	-0.01
79) Terphenyl-d14	12.910	244	997287	53.786	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080423\
Data File : BF134625.D
Acq On : 04 Aug 2023 11:11
Operator : CG\JU
Sample : 03756-04
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RPTP-10-CS-10

Quant Time: Aug 05 00:04:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

