

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080423\
 Data File : BF134640.D
 Acq On : 04 Aug 2023 19:07
 Operator : CG\JU
 Sample : 03775-04DL 25X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 BUILDING-A-6-(5-10)DL

Quant Time: Aug 05 00:09:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/07/2023
 Supervised By :mohammad ahmed 08/07/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	163797	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	572414	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	226426	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	367884	20.000	ng	# 0.00
76) Chrysene-d12	13.974	240	333131	20.000	ng	# 0.00
86) Perylene-d12	15.439	264	264497	20.000	ng	# 0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	42350	3.930	ng	-0.01
7) Phenol-d6	6.422	99	55309	4.015	ng	-0.02
23) Nitrobenzene-d5	7.351	82	26286	2.870	ng	-0.02
42) 2,4,6-Tribromophenol	10.616	330	6018	2.607	ng	-0.01
45) 2-Fluorobiphenyl	9.151	172	50786	3.729	ng	-0.01
79) Terphenyl-d14	12.910	244	57994	3.069	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	8.116	128	5646858	196.286	ng	96
37) 2-Methylnaphthalene	8.798	142	1730258	90.738	ng	99
38) 1-Methylnaphthalene	8.892	142	1284547	72.442	ng	97
46) 1,1'-Biphenyl	9.251	154	514680	29.231	ng	97
49) Acenaphthylene	9.692	152	800907	38.619	ng	98
52) Acenaphthene	9.863	154	372262	27.699	ng	99
55) Dibenzofuran	10.033	168	350852	18.339	ng	# 95
58) Fluorene	10.380	166	852411	61.169	ng	95
71) Phenanthrene	11.357	178	3296964	168.725	ng	99
72) Anthracene	11.398	178	1239428	62.097	ng	98
73) Carbazole	11.551	167	368230	20.662	ng	99
75) Fluoranthene	12.551	202	2924179	141.669	ng	97
78) Pyrene	12.780	202	3151205	109.317	ng	95
81) Benzo(a)anthracene	13.968	228	1815811	76.229	ng	96
83) Chrysene	14.004	228	1525393	65.729	ng	97
87) Indeno(1,2,3-cd)pyrene	16.927	276	450632	28.969	ng	# 96
88) Benzo(b)fluoranthene	15.015	252	1335233m	82.772	ng	
89) Benzo(k)fluoranthene	15.039	252	407756m	25.374	ng	
90) Benzo(a)pyrene	15.380	252	997749	66.337	ng	# 94
91) Dibenzo(a,h)anthracene	16.939	278	125369	9.973	ng	# 91
92) Benzo(g,h,i)perylene	17.374	276	426314	33.244	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080423\
Data File : BF134640.D
Acq On : 04 Aug 2023 19:07
Operator : CG\JU
Sample : 03775-04DL 25X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL

Quant Time: Aug 05 00:09:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.798 min Scan# 800
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL



Tgt Ion:152 Resp: 16379
Ion Ratio Lower Upper
152 100
150 156.5 128.3 192.5
115 63.9 52.4 78.6




Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023



#5
2-Fluorophenol
Concen: 3.930 ng
RT: 5.416 min Scan# 566
Delta R.T. -0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07


Tgt Ion:112 Resp: 42350
Ion Ratio Lower Upper
112 100
64 60.2 44.9 67.3
63 42.7 25.4 38.0#





#7
Phenol-d6
Concen: 4.015 ng
RT: 6.422 min Scan# 71
Delta R.T. -0.018 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL



Tgt Ion: 99 Resp: 55309
Ion Ratio Lower Upper
99 100
42 17.8 18.7 28.1#
71 29.7 36.4 54.6#




Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.081 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07


Tgt Ion:136 Resp: 572414
Ion Ratio Lower Upper
136 100
137 11.4 8.4 12.6
54 8.3 5.2 7.8#
68 6.3 2.9 4.3#





#23
Nitrobenzene-d5
Concen: 2.870 ng
RT: 7.351 min Scan# 896
Delta R.T. -0.018 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL




Tgt Ion: 82 Resp: 26280
Ion Ratio Lower Upper
82 100
128 43.3 34.3 51.5
54 59.6 43.8 65.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023




#31
Naphthalene
Concen: 196.286 ng
RT: 8.116 min Scan# 1025
Delta R.T. 0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07



Tgt Ion:128 Resp: 5646858
Ion Ratio Lower Upper
128 100
129 12.1 8.4 12.6
127 14.8 10.7 16.1





#37

2-Methylnaphthalene

Concen: 90.738 ng

RT: 8.798 min Scan# 1141

Delta R.T. 0.006 min

Lab File: BF134640.D

Acq: 04 Aug 2023 19:07

Instrument :

BNA_F

ClientSampleId :

BUILDING-A-6-(5-10)DL

Tgt Ion:142 Resp: 1730258

Ion Ratio Lower Upper

142 100





141 89.2 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/07/2023

Supervised By :mohammad ahmed 08/07/2023



#38

1-Methylnaphthalene

Concen: 72.442 ng

RT: 8.892 min Scan# 1157

Delta R.T. 0.000 min

Lab File: BF134640.D




Acq: 04 Aug 2023 19:07


Tgt Ion:142 Resp: 1284547

Ion Ratio Lower Upper

142 100


141 92.1 71.4 107.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.833 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL



Tgt Ion:164 Resp: 226420
Ion Ratio Lower Upper
164 100
162 101.2 82.6 123.8
160 45.6 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023



#42
2,4,6-Tribromophenol
Concen: 2.607 ng
RT: 10.616 min Scan# 1450
Delta R.T. -0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07


Tgt Ion:330 Resp: 6018
Ion Ratio Lower Upper
330 100
332 100.6 78.8 118.2
141 63.9 25.4 38.2#





#45
2-Fluorobiphenyl
Concen: 3.729 ng
RT: 9.151 min Scan# 1201
Delta R.T. -0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL




Tgt Ion:172 Resp: 50780
Ion Ratio Lower Upper
172 100
171 35.3 28.8 43.2
170 22.4 18.7 28.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023




#46
1,1'-Biphenyl
Concen: 29.231 ng
RT: 9.251 min Scan# 1218
Delta R.T. -0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Tgt Ion:154 Resp: 514680
Ion Ratio Lower Upper
154 100
153 39.8 20.8 60.8
76 12.9 0.0 30.2





#49
Acenaphthylene
Concen: 38.619 ng
RT: 9.692 min Scan# 1293
Delta R.T. -0.006 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL




Tgt Ion:152 Resp: 800907
Ion Ratio Lower Upper
152 100
151 20.6 16.4 24.6
153 15.6 10.9 16.3

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023




#52
Acenaphthene
Concen: 27.699 ng
RT: 9.863 min Scan# 1322
Delta R.T. -0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Tgt Ion:154 Resp: 372262
Ion Ratio Lower Upper
154 100
153 112.8 90.0 135.0
152 52.1 42.4 63.6





#55
Dibenzenofuran
Concen: 18.339 ng
RT: 10.033 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL




Tgt Ion:168 Resp: 350852
Ion Ratio Lower Upper
168 100
139 39.7 33.4 50.2
169 16.8 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023




#58
Fluorene
Concen: 61.169 ng
RT: 10.380 min Scan# 1410
Delta R.T. -0.006 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Tgt Ion:166 Resp: 852411
Ion Ratio Lower Upper
166 100
165 98.5 74.5 111.7
167 14.1 10.6 16.0





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.322 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL




Tgt Ion:188 Resp: 367884
Ion Ratio Lower Upper
188 100
94 10.1 5.1 7.7
80 10.5 6.6 9.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023




#71
Phenanthrene
Concen: 168.725 ng
RT: 11.357 min Scan# 1576
Delta R.T. 0.006 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Tgt Ion:178 Resp: 3296964
Ion Ratio Lower Upper
178 100
176 20.8 15.8 23.6
179 16.1 12.7 19.1





#72
 Anthracene
 Concen: 62.097 ng
 RT: 11.398 min Scan# 1139
 Delta R.T. -0.006 min
 Lab File: BF134640.D
 Acq: 04 Aug 2023 19:07




Instrument :
 BNA_F
 ClientSampleId :
 BUILDING-A-6-(5-10)DL




Tgt Ion:178 Resp: 1239428
 Ion Ratio Lower Upper
 178 100
 176 18.8 14.6 21.8
 179 16.2 12.3 18.5

Manual Integrations
 APPROVED



Reviewed By :Yogesh Patel 08/07/2023
 Supervised By :mohammad ahmed 08/07/2023




#73
 Carbazole
 Concen: 20.662 ng
 RT: 11.551 min Scan# 1609
 Delta R.T. -0.006 min
 Lab File: BF134640.D
 Acq: 04 Aug 2023 19:07




Tgt Ion:167 Resp: 368230
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.5 26.3
 139 13.3 11.0 16.4





#75
Fluoranthene
Concen: 141.669 ng
RT: 12.551 min Scan# 11
Delta R.T. 0.012 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL



Tgt Ion:202 Resp: 2924179
Ion Ratio Lower Upper
202 100
101 10.6 0.0 28.5
203 18.5 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.974 min Scan# 2021
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07

Tgt Ion:240 Resp: 333131
Ion Ratio Lower Upper
240 100
120 11.9 6.6 10.0#
236 25.3 20.2 30.2





#78

Pyrene

Concen: 109.317 ng

RT: 12.780 min Scan# 1818

Delta R.T. 0.012 min

Lab File: BF134640.D


Acq: 04 Aug 2023 19:07

Instrument :

BNA_F

ClientSampleId :

BUILDING-A-6-(5-10)DL



Tgt Ion:202 Resp: 315120

Ion Ratio Lower Upper

202 100

200 22.4 15.8 23.8




203 18.9 13.5 20.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/07/2023

Supervised By :mohammad ahmed 08/07/2023



#79

Terphenyl-d14

Concen: 3.069 ng

RT: 12.910 min Scan# 1840

Delta R.T. -0.012 min

Lab File: BF134640.D

Acq: 04 Aug 2023 19:07




Tgt Ion:244 Resp: 57994


Ion Ratio Lower Upper

244 100

212 7.3 5.6 8.4


122 12.9 8.0 12.0#





#81
Benzo(a)anthracene
Concen: 76.229 ng
RT: 13.968 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL



Tgt Ion:228 Resp: 181581
Ion Ratio Lower Upper
228 100
226 30.2 22.3 33.5
229 21.4 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023



#83
Chrysene
Concen: 65.729 ng
RT: 14.004 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07


Tgt Ion:228 Resp: 1525393
Ion Ratio Lower Upper
228 100
226 30.9 24.5 36.7
229 23.2 15.8 23.8





#86
Perylene-d12
Concen: 20.000 ng
RT: 15.439 min Scan# 21
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL



Tgt Ion:264 Resp: 264497
Ion Ratio Lower Upper
264 100
260 23.1 18.4 27.6
265 25.4 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023



#87
Indeno(1,2,3-cd)pyrene
Concen: 28.969 ng
RT: 16.927 min Scan# 2523
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07


Tgt Ion:276 Resp: 450632
Ion Ratio Lower Upper
276 100
138 21.7 14.2 21.4#
277 25.1 19.9 29.9





#88
Benzo(b)fluoranthene
Concen: 82.772 ng m
RT: 15.015 min Scan# 2194
Delta R.T. 0.006 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07




Instrument :
BNA_F
ClientSampleId :
BUILDING-A-6-(5-10)DL




Tgt Ion:252 Resp: 133523
Ion Ratio Lower Upper
252 100
253 23.1 17.3 25.9
125 11.5 6.0 9.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/07/2023
Supervised By :mohammad ahmed 08/07/2023




#89
Benzo(k)fluoranthene
Concen: 25.374 ng m
RT: 15.039 min Scan# 2202
Delta R.T. 0.000 min
Lab File: BF134640.D
Acq: 04 Aug 2023 19:07



Tgt Ion:252 Resp: 407756
Ion Ratio Lower Upper
252 100
253 24.1 17.9 26.9
125 10.5 6.3 9.5





#90

Benzo(a)pyrene

Concen: 66.337 ng

RT: 15.380 min Scan# 21

Delta R.T. 0.000 min

Lab File: BF134640.D


Acq: 04 Aug 2023 19:07

Instrument :

BNA_F

ClientSampleId :

BUILDING-A-6-(5-10)DL



Tgt Ion:252 Resp: 997749

Ion Ratio Lower Upper

252 100

253 23.2 17.0 25.4




125 11.2 6.2 9.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/07/2023

Supervised By :mohammad ahmed 08/07/2023



#91

Dibenzo(a,h)anthracene

Concen: 9.973 ng

RT: 16.939 min Scan# 2525

Delta R.T. -0.012 min

Lab File: BF134640.D

Acq: 04 Aug 2023 19:07




Tgt Ion:278 Resp: 125369


Ion Ratio Lower Upper

278 100

139 15.6 10.0 15.0


279 29.1 19.7 29.5





#92
 Benzo(g,h,i)perylene
 Concen: 33.244 ng
 RT: 17.374 min Scan# 2599
 Delta R.T. 0.000 min
 Lab File: BF134640.D
 Acq: 04 Aug 2023 19:07

Instrument :
 BNA_F
 ClientSampleId :
 BUILDING-A-6-(5-10)DL



Tgt Ion:276 Resp: 426314
 Ion Ratio Lower Upper
 276 100
 277 24.7 19.0 28.6
 138 21.5 12.2 18.2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/07/2023
 Supervised By :mohammad ahmed 08/07/2023

