

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080519\  
 Data File : BF115984.D  
 Acq On : 5 Aug 2019 17:21  
 Operator : HP/JU  
 Sample : K4156-07  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 RT3380

Quant Time: Aug 06 08:07:37 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jul 31 15:31:51 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)      |
|-----------------------------|-------|------|----------|--------|-------|---------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.85  | 152  | 132824   | 20.00  | ng    | 0.00          |
| 21) Naphthalene-d8          | 8.12  | 136  | 490303   | 20.00  | ng    | -0.01         |
| 39) Acenaphthene-d10        | 9.87  | 164  | 278821   | 20.00  | ng    | -0.02         |
| 64) Phenanthrene-d10        | 11.36 | 188  | 503391   | 20.00  | ng    | -0.02         |
| 76) Chrysene-d12            | 14.00 | 240  | 417106   | 20.00  | ng    | -0.02         |
| 87) Perylene-d12            | 15.47 | 264  | 305010   | 20.00  | ng    | -0.02         |
| System Monitoring Compounds |       |      |          |        |       |               |
| 5) 2-Fluorophenol           | 5.47  | 112  | 641622   | 80.87  | ng    | 0.00          |
| 7) Phenol-d6                | 6.48  | 99   | 1192741  | 119.15 | ng    | 0.00          |
| 23) Nitrobenzene-d5         | 7.40  | 82   | 740265   | 87.15  | ng    | -0.01         |
| 42) 2,4,6-Tribromophenol    | 10.67 | 330  | 282882   | 104.65 | ng    | -0.02         |
| 45) 2-Fluorobiphenyl        | 9.20  | 172  | 1051595  | 70.64  | ng    | -0.02         |
| 79) Terphenyl-d14           | 12.94 | 244  | 1150727  | 58.13  | ng    | -0.02         |
| Target Compounds            |       |      |          |        |       |               |
| 50) Dimethylphthalate       | 9.59  | 163  | 69960    | 3.685  | ng    | Qvalue<br>100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080519\  
Data File : BF115984.D  
Acq On : 5 Aug 2019 17:21  
Operator : HP/JU  
Sample : K4156-07  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
RT3380

Quant Time: Aug 06 08:07:37 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jul 31 15:31:51 2019  
Response via : Initial Calibration

