

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080519\
 Data File : BF115985.D
 Acq On : 5 Aug 2019 17:47
 Operator : HP/JU
 Sample : K4153-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RT-4221

Quant Time: Aug 06 02:31:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	144016	20.00	ng	-0.01
21) Naphthalene-d8	8.12	136	553584	20.00	ng	-0.01
39) Acenaphthene-d10	9.87	164	293393	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	535416	20.00	ng	-0.02
76) Chrysene-d12	14.00	240	426589	20.00	ng	-0.02
87) Perylene-d12	15.46	264	375402	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	1100228	127.89	ng	0.00
7) Phenol-d6	6.48	99	1461098	134.61	ng	0.00
23) Nitrobenzene-d5	7.40	82	912982	95.20	ng	-0.01
42) 2,4,6-Tribromophenol	10.67	330	367967	129.36	ng	-0.02
45) 2-Fluorobiphenyl	9.20	172	1486021	94.87	ng	-0.02
79) Terphenyl-d14	12.95	244	1666927	82.34	ng	-0.02
Target Compounds						
50) Dimethylphthalate	9.59	163	103966	5.204	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080519\
Data File : BF115985.D
Acq On : 5 Aug 2019 17:47
Operator : HP/JU
Sample : K4153-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RT-4221

Quant Time: Aug 06 02:31:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:31:51 2019
Response via : Initial Calibration

