

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080522\
 Data File : BF129608.D
 Acq On : 05 Aug 2022 19:32
 Operator : CG\JU
 Sample : N4023-08
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP3-1-6

Quant Time: Aug 08 02:00:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 05 16:27:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	173845	20.000	ng	0.00
21) Naphthalene-d8	8.128	136	696826	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	369701	20.000	ng	-0.01
64) Phenanthrene-d10	11.374	188	618232	20.000	ng	0.00
76) Chrysene-d12	14.016	240	347382	20.000	ng	-0.01
86) Perylene-d12	15.492	264	289667	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	1281477	120.080	ng	0.02
7) Phenol-d6	6.498	99	1419415	105.816	ng	0.00
23) Nitrobenzene-d5	7.416	82	1177678	92.258	ng	0.00
42) 2,4,6-Tribromophenol	10.680	330	528532	148.698	ng	0.00
45) 2-Fluorobiphenyl	9.204	172	2053339	85.918	ng	0.00
79) Terphenyl-d14	12.963	244	2042528	110.927	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080522\
Data File : BF129608.D
Acq On : 05 Aug 2022 19:32
Operator : CG\JU
Sample : N4023-08
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP3-1-6

Quant Time: Aug 08 02:00:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 05 16:27:34 2022
Response via : Initial Calibration

