

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080619\
 Data File : BF116016.D
 Acq On : 6 Aug 2019 16:37
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 07 07:29:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.604	0.593	1.8	108	0.00
3	Pyridine	1.686	1.611	4.4	104	0.00
4	n-Nitrosodimethylamine	0.665	0.658	1.1	107	-0.02
5 S	2-Fluorophenol	1.195	1.245	-4.2	110	0.00
6	Aniline	2.159	2.185	-1.2	107	0.00
7 S	Phenol-d6	1.507	1.577	-4.6	112	0.00
8	2-Chlorophenol	1.321	1.425	-7.9	114	0.00
9	Benzaldehyde	1.040	0.967	7.0	98	0.00
10 C	Phenol	1.775	1.848	-4.1	111	0.00
11	bis(2-Chloroethyl)ether	1.305	1.416	-8.5	116	0.00
12	1,3-Dichlorobenzene	1.527	1.590	-4.1	111	0.00
13 C	1,4-Dichlorobenzene	1.529	1.601	-4.7	111	0.00
14	1,2-Dichlorobenzene	1.408	1.476	-4.8	109	0.00
15	Benzyl Alcohol	1.144	1.224	-7.0	111	0.00
16	2,2'-oxybis(1-Chloropropane	1.780	1.951	-9.6	115	0.00
17	2-Methylphenol	1.119	1.200	-7.2	116	0.00
18	Hexachloroethane	0.563	0.604	-7.3	113	0.00
19 P	n-Nitroso-di-n-propylamine	0.934	1.023	-9.5	118	0.00
20	3+4-Methylphenols	1.324	1.417	-7.0	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
22	Acetophenone	0.439	0.457	-4.1	115	0.00
23 S	Nitrobenzene-d5	0.346	0.358	-3.5	114	0.00
24	Nitrobenzene	0.374	0.393	-5.1	116	0.00
25	Isophorone	0.663	0.715	-7.8	120	0.00
26 C	2-Nitrophenol	0.176	0.194	-10.2	121	0.00
27	2,4-Dimethylphenol	0.265	0.275	-3.8	114	0.00
28	bis(2-Chloroethoxy)methane	0.411	0.445	-8.3	120	0.00
29 C	2,4-Dichlorophenol	0.280	0.302	-7.9	117	0.00
30	1,2,4-Trichlorobenzene	0.319	0.335	-5.0	115	0.00
31	Naphthalene	0.941	0.996	-5.8	115	0.00
32	Benzoic acid	0.187	0.197	-5.3	128	-0.01
33	4-Chloroaniline	0.421	0.439	-4.3	114	0.00
34 C	Hexachlorobutadiene	0.184	0.195	-6.0	116	0.00
35	Caprolactam	0.091	0.094	-3.3	114	-0.01
36 C	4-Chloro-3-methylphenol	0.291	0.318	-9.3	120	0.00
37	2-Methylnaphthalene	0.620	0.663	-6.9	116	0.00
38	1-Methylnaphthalene	0.586	0.635	-8.4	118	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
40	1,2,4,5-Tetrachlorobenzene	0.535	0.565	-5.6	118	0.00
41 P	Hexachlorocyclopentadiene	0.209	0.224	-7.2	118	0.00
42 S	2,4,6-Tribromophenol	0.194	0.203	-4.6	117	0.00
43 C	2,4,6-Trichlorophenol	0.368	0.371	-0.8	113	0.00
44	2,4,5-Trichlorophenol	0.380	0.389	-2.4	115	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080619\
 Data File : BF116016.D
 Acq On : 6 Aug 2019 16:37
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 07 07:29:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 31 15:31:51 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.068	1.095	-2.5	114	0.00
46	1,1'-Biphenyl	1.412	1.464	-3.7	116	0.00
47	2-Chloronaphthalene	1.175	1.201	-2.2	115	0.00
48	2-Nitroaniline	0.388	0.396	-2.1	115	0.00
49	Acenaphthylene	1.715	1.792	-4.5	116	0.00
50	Dimethylphthalate	1.362	1.447	-6.2	120	0.00
51	2,6-Dinitrotoluene	0.296	0.312	-5.4	118	0.00
52 C	Acenaphthene	1.052	1.083	-2.9	114	0.00
53	3-Nitroaniline	0.366	0.365	0.3	113	0.00
54 P	2,4-Dinitrophenol	0.126	0.130	-3.2	114	0.00
55	Dibenzofuran	1.530	1.594	-4.2	115	0.00
56 P	4-Nitrophenol	0.234	0.218	6.8	104	0.00
57	2,4-Dinitrotoluene	0.394	0.408	-3.6	114	0.00
58	Fluorene	1.177	1.241	-5.4	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.320	0.336	-5.0	118	0.00
60	Diethylphthalate	1.271	1.358	-6.8	119	0.00
61	4-Chlorophenyl-phenylether	0.596	0.638	-7.0	118	0.00
62	4-Nitroaniline	0.378	0.375	0.8	111	0.00
63	Azobenzene	1.308	1.401	-7.1	120	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.118	-11.3	120	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.640	-6.3	117	0.00
67	4-Bromophenyl-phenylether	0.214	0.232	-8.4	118	0.00
68	Hexachlorobenzene	0.248	0.264	-6.5	116	0.00
69	Atrazine	0.198	0.160	19.2	89	0.00
70 C	Pentachlorophenol	0.120	0.112	6.7	100	0.00
71	Phenanthrene	1.022	1.067	-4.4	114	0.00
72	Anthracene	1.035	1.078	-4.2	114	0.00
73	Carbazole	0.939	0.984	-4.8	114	0.00
74	Di-n-butylphthalate	1.095	1.238	-13.1	120	0.00
75 C	Fluoranthene	1.070	1.114	-4.1	114	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
77	Benzidine	0.676	0.705	-4.3	103	0.00
78	Pyrene	1.508	1.584	-5.0	114	0.00
79 S	Terphenyl-d14	0.949	0.997	-5.1	114	0.00
80	Butylbenzylphthalate	0.638	0.733	-14.9	121	0.00
81	Benzo(a)anthracene	1.278	1.385	-8.4	115	0.00
82	3,3'-Dichlorobenzidine	0.444	0.500	-12.6	116	0.00
83	Chrysene	1.241	1.330	-7.2	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.829	1.009	-21.7	127	0.00
85 c	Di-n-octyl phthalate	1.316	1.564	-18.8	122	0.00
86	Indeno(1,2,3-cd)pyrene	1.042	1.055	-1.2	112	0.00
87 I	Perylene-d12	1.000	1.000	0.0	114	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080619\
Data File : BF116016.D
Acq On : 6 Aug 2019 16:37
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 07 07:29:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 31 15:31:51 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.156	1.192	-3.1	112	0.00
89	Benzo(k)fluoranthene	1.126	1.181	-4.9	124	0.00
90 C	Benzo(a)pyrene	1.034	1.069	-3.4	117	0.00
91	Dibenzo(a,h)anthracene	0.895	0.863	3.6	113	0.00
92	Benzo(q,h,i)perylene	0.879	0.806	8.3	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0