

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080723\
 Data File : BF134644.D
 Acq On : 07 Aug 2023 09:48
 Operator : CG\JU
 Sample : 03891-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-1

Quant Time: Aug 08 01:19:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	192864	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	770504	20.000	ng	# 0.00
39) Acenaphthene-d10	9.828	164	391745	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	686165	20.000	ng	# 0.00
76) Chrysene-d12	13.963	240	414347	20.000	ng	#-0.01
86) Perylene-d12	15.433	264	379236	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	1416560	111.649	ng	0.00
7) Phenol-d6	6.434	99	1812944	111.782	ng	0.00
23) Nitrobenzene-d5	7.357	82	1113819	90.338	ng	-0.01
42) 2,4,6-Tribromophenol	10.622	330	496589	124.350	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1875237	79.591	ng	0.00
79) Terphenyl-d14	12.922	244	2089187	88.886	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080723\
Data File : BF134644.D
Acq On : 07 Aug 2023 09:48
Operator : CG\JU
Sample : 03891-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-1

Quant Time: Aug 08 01:19:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

