

Data File : BF116078.D
Acq On : 8 Aug 2019 20:18
Operator : HP/JU
Sample : K4131-31
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
1S-E1-E4-COMP-(0-1)

Quant Time: Aug 12 04:00:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 11:13:18 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	131393	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	508034	20.00	ng	-0.01
39) Acenaphthene-d10	9.87	164	280844	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	547522	20.00	ng	-0.01
76) Chrysene-d12	14.00	240	508956	20.00	ng	-0.01
87) Perylene-d12	15.46	264	478712	20.00	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.47	112	912892	116.31	ng	0.00
7) Phenol-d6	6.47	99	1062064	107.25	ng	-0.01
23) Nitrobenzene-d5	7.40	82	872792	99.17	ng	-0.02
42) 2,4,6-Tribromophenol	10.66	330	415184	152.48	ng	-0.01
45) 2-Fluorobiphenyl	9.19	172	1457973	97.24	ng	-0.01
79) Terphenyl-d14	12.94	244	2005384	83.03	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BF116078.D
Acq On : 8 Aug 2019 20:18
Operator : HP/JU
Sample : K4131-31
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
1S-E1-E4-COMP-(0-1)

Quant Time: Aug 12 04:00:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 11:13:18 2019
Response via : Initial Calibration

