

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080819\
 Data File : BF116097.D
 Acq On : 9 Aug 2019 4:44
 Operator : HP/JU
 Sample : K4154-05
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PS-5

Manual Integrations
 APPROVED

mohammad
 8/14/2019 9:21:59 AM

Quant Time: Aug 09 17:15:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 11:13:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	134402	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	539275	20.00	ng	-0.01
39) Acenaphthene-d10	9.87	164	267206	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	389660	20.00	ng	-0.01
76) Chrysene-d12	14.00	240	295404	20.00	ng	-0.01
87) Perylene-d12	15.46	264	278887	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	651674	81.17	ng	0.00
7) Phenol-d6	6.47	99	856518	84.56	ng	-0.01
23) Nitrobenzene-d5	7.40	82	490083	52.46	ng	-0.02
42) 2,4,6-Tribromophenol	10.66	330	157268	60.71	ng	-0.01
45) 2-Fluorobiphenyl	9.19	172	723195	50.69	ng	-0.01
79) Terphenyl-d14	12.94	244	539115	38.46	ng	-0.01
Target Compounds						Qvalue
50) Dimethylphthalate	9.59	163	227544	12.507	ng	98
75) Fluoranthene	12.57	202	68503	3.285	ng	100
78) Pyrene	12.80	202	86638	3.891	ng	98
81) Benzo(a)anthracene	13.99	228	40263	2.132	ng	97
83) Chrysene	14.03	228	44703	2.439	ng	99
88) Benzo(b)fluoranthene	15.04	252	38055m	2.360	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF080819\
Data File : BF116097.D
Acq On : 9 Aug 2019 4:44
Operator : HP/JU
Sample : K4154-05
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PS-5

Manual Integrations
APPROVED

mohammad
8/14/2019 9:21:59 AM

Quant Time: Aug 09 17:15:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 11:13:18 2019
Response via : Initial Calibration

