

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080823\
 Data File : BF134669.D
 Acq On : 08 Aug 2023 13:44
 Operator : CG\JU
 Sample : 03903-11
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-11

Quant Time: Aug 09 02:17:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	186905	20.000	ng	0.00
21) Naphthalene-d8	8.080	136	682722	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	301453	20.000	ng	0.00
64) Phenanthrene-d10	11.321	188	494197	20.000	ng	# 0.00
76) Chrysene-d12	13.974	240	342912	20.000	ng	0.00
86) Perylene-d12	15.445	264	290664	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	954892	77.661	ng	0.00
7) Phenol-d6	6.434	99	1177658	74.926	ng	0.00
23) Nitrobenzene-d5	7.363	82	663077	60.694	ng	0.00
42) 2,4,6-Tribromophenol	10.621	330	237436	77.265	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1066472	58.822	ng	0.00
79) Terphenyl-d14	12.915	244	938391	48.241	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080823\
Data File : BF134669.D
Acq On : 08 Aug 2023 13:44
Operator : CG\JU
Sample : 03903-11
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-11

Quant Time: Aug 09 02:17:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

