

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081018\
 Data File : BF108085.D
 Acq On : 10 Aug 2018 15:49
 Operator : JU/SJ
 Sample : J4409-05 5X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-10C

Manual Integrations
 APPROVED

Sohil
 8/13/2018 3:51:54 PM

Quant Time: Aug 10 17:33:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.02	152	256486	20.00	ng	0.00
21) Naphthalene-d8	8.30	136	947994	20.00	ng	0.00
38) Acenaphthene-d10	10.07	164	438790	20.00	ng	0.00
63) Phenanthrene-d10	11.56	188	755068	20.00	ng	0.00
75) Chrysene-d12	14.22	240	629233	20.00	ng	0.00
86) Perylene-d12	15.79	264	662469	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	357198	25.21	ng	0.00
7) Phenol-d6	6.62	99	477016	25.34	ng	-0.01
23) Nitrobenzene-d5	7.57	82	299689	17.64	ng	0.00
41) 2,4,6-Tribromophenol	10.85	330	91755	20.96	ng	0.00
44) 2-Fluorobiphenyl	9.37	172	498449	18.24	ng	0.00
78) Terphenyl-d14	13.15	244	435040	17.58	ng	0.00
Target Compounds						
48) Acenaphthylene	9.92	152	99457	2.460	ng	99
49) Dimethylphthalate	9.77	163	79316	2.595	ng	# 98
51) Acenaphthene	10.10	154	93180	3.763	ng	98
57) Fluorene	10.61	166	76982	2.711	ng	97
70) Phenanthrene	11.59	178	848590	23.828	ng	98
71) Anthracene	11.64	178	456089	12.495	ng	99
74) Fluoranthene	12.79	202	1385631	35.649	ng	95
77) Pyrene	13.02	202	1629455	40.903	ng	97
80) Benzo(a)anthracene	14.21	228	946521m	26.211	ng	
82) Chrysene	14.25	228	776177	23.445	ng	99
85) Indeno(1,2,3-cd)pyrene	17.41	276	336994m	11.748	ng	
87) Benzo(b)fluoranthene	15.32	252	929578m	23.483	ng	
88) Benzo(k)fluoranthene	15.34	252	276857m	7.608	ng	
89) Benzo(a)pyrene	15.72	252	861864	24.314	ng	97
90) Dibenzo(a,h)anthracene	17.41	278	88161	3.006	ng	93
91) Benzo(g,h,i)perylene	17.90	276	307166m	10.636	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081018\
Data File : BF108085.D
Acq On : 10 Aug 2018 15:49
Operator : JU/SJ
Sample : J4409-05 5X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10C

Manual Integrations
APPROVED

Sohil
8/13/2018 3:51:54 PM

Quant Time: Aug 10 17:33:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration

