




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081018\
Data File : BF108104.D
Acq On : 11 Aug 2018 00:59
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 13 01:31:31 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration



TIC: BF108104.D

(69) Pentachlorophenol (C)




11.360min (+0.006) 26294.82ng

response 62744

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	67.23
264.00	61.90	61.60
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081018\
 Data File : BF108104.D
 Acq On : 11 Aug 2018 00:59
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 13 01:31:31 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration



TIC: BF108104.D

(76) Benzidine

12.907min (+0.006) 10491.19ng

response 599724

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	15.07
183.00	11.60	12.72
0.00	0.00	0.00