

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081018\
 Data File : BF108082.D
 Acq On : 10 Aug 2018 14:28
 Operator : JU/SJ
 Sample : J4409-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-10B

Manual Integrations
 APPROVED

Sohil
 8/13/2018 3:51:45 PM

Quant Time: Aug 10 16:23:08 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.02	152	283122	20.00	ng	0.00
21) Naphthalene-d8	8.30	136	1038177	20.00	ng	0.00
38) Acenaphthene-d10	10.07	164	502764	20.00	ng	0.00
63) Phenanthrene-d10	11.56	188	879868	20.00	ng	0.00
75) Chrysene-d12	14.22	240	623690	20.00	ng	0.00
86) Perylene-d12	15.79	264	655094	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.64	112	1767018	112.99	ng	0.01
7) Phenol-d6	6.64	99	2348949	113.03	ng	0.00
23) Nitrobenzene-d5	7.58	82	1578788	84.86	ng	0.00
41) 2,4,6-Tribromophenol	10.86	330	585739	116.80	ng	0.00
44) 2-Fluorobiphenyl	9.38	172	2533546	80.90	ng	0.00
78) Terphenyl-d14	13.15	244	2143625	87.39	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.64	94	109955	5.115	ng	73
31) Naphthalene	8.32	128	255936	5.421	ng	99
48) Acenaphthylene	9.92	152	203050	4.384	ng	99
49) Dimethylphthalate	9.77	163	447755	12.785	ng	99
51) Acenaphthene	10.10	154	318941	11.242	ng	99
54) Dibenzofuran	10.27	168	90669	2.206	ng	97
57) Fluorene	10.61	166	258890	7.957	ng	99
70) Phenanthrene	11.59	178	2425007	58.434	ng	95
71) Anthracene	11.64	178	996243	23.422	ng	99
72) Carbazole	11.78	167	90344	2.391	ng	98
74) Fluoranthene	12.79	202	3068368	67.746	ng	94
77) Pyrene	13.02	202	3163490m	80.117	ng	
80) Benzo(a)anthracene	14.21	228	1432865m	40.032	ng	
82) Chrysene	14.25	228	1183916	36.078	ng	98
85) Indeno(1,2,3-cd)pyrene	17.41	276	652285	22.941	ng	# 87
87) Benzo(b)fluoranthene	15.32	252	1568795m	40.077	ng	
88) Benzo(k)fluoranthene	15.34	252	523984m	14.562	ng	
89) Benzo(a)pyrene	15.72	252	1331255	37.978	ng	96
90) Dibenzo(a,h)anthracene	17.42	278	155667	5.367	ng	95
91) Benzo(g,h,i)perylene	17.91	276	606513	21.237	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081018\
Data File : BF108082.D
Acq On : 10 Aug 2018 14:28
Operator : JU/SJ
Sample : J4409-03
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Manual Integrations
APPROVED

Sohil
8/13/2018 3:51:45 PM

Quant Time: Aug 10 16:23:08 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.02 min Scan# 796
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	283122		
150	156.8	124.6	186.8	
115	57.4	50.1	75.1	

Manual Integrations
APPROVED




Sohil
8/13/2018 3:51:45 PM




#5

2-Fluorophenol
Concen: 112.994 ng
RT: 5.64 min Scan# 562
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Resp	Lower	Upper
112	100	1767018		
64	73.5	47.9	71.9#	
63	39.3	23.7	35.5#	





#7

Phenol-d6

Concen: 113.033 ng

RT: 6.64 min Scan# 731

Delta R.T. 0.01 min

Lab File: BF108082.D

Acq: 10 Aug 2018 14:28

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-10B

Tgt Ion: 99 Resp: 2348949

Ion Ratio Lower Upper

99 100

42 24.6 14.5 21.7#





71 39.0 25.4 38.0#

Manual Integrations

APPROVED

Sohil

8/13/2018 3:51:45 PM



#10

Phenol

Concen: 5.115 ng

RT: 6.64 min Scan# 732

Delta R.T. 0.00 min

Lab File: BF108082.D

Acq: 10 Aug 2018 14:28




Tgt Ion: 94 Resp: 109955


Ion Ratio Lower Upper

94 100

65 37.1 7.9 47.9

66 56.1 16.2 56.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.30 min Scan# 1014
Delta R.T. 0.00 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B




Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.9	13.3
54	9.5	6.6	9.8
68	6.0	5.0	7.4

Manual Integrations
APPROVED

Sohil
8/13/2018 3:51:45 PM




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1






#23
Nitrobenzene-d5
Concen: 84.859 ng
RT: 7.58 min Scan# 891
Delta R.T. 0.00 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Lower	Upper
82	100		
128	43.2	36.1	54.1
54	61.1	41.4	62.2



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF1









#31
Naphthalene
Concen: 5.421 ng
RT: 8.32 min Scan# 1017
Delta R.T. -0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	255936		
129	10.8		8.8	13.2
127	13.3		10.9	16.3




Manual Integrations
APPROVED


Sohil
8/13/2018 3:51:45 PM



#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.07 min Scan# 1314
Delta R.T. 0.00 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Resp	Lower	Upper
164	100	502764		
162	103.6		86.1	129.1
160	47.0		38.3	57.5






#41
 2,4,6-Tribromophenol
 Concen: 116.799 ng
 RT: 10.86 min Scan# 1449
 Delta R.T. 0.00 min
 Lab File: BF108082.D
 Acq: 10 Aug 2018 14:28

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-10B




Tgt Ion:	330	Resp:	585739
Ion Ratio	Lower	Upper	
330	100		
332	95.0	77.0	115.6
141	33.4	29.9	44.9

Manual Integrations
 APPROVED

Sohil
 8/13/2018 3:51:45 PM




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 80.903 ng
 RT: 9.38 min Scan# 1197
 Delta R.T. 0.00 min
 Lab File: BF108082.D
 Acq: 10 Aug 2018 14:28

Tgt Ion:	172	Resp:	2533546
Ion Ratio	Lower	Upper	
172	100		
171	40.0	29.7	44.5
170	27.3	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E








#48
Acenaphthylene
Concen: 4.384 ng
RT: 9.92 min Scan# 1290
Delta R.T. -0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Resp	Lower	Upper
152	100	203050		
151	20.5	16.3	24.5	
153	12.9	10.7	16.1	




Manual Integrations
APPROVED


Sohil
8/13/2018 3:51:45 PM



#49
Dimethylphthalate
Concen: 12.785 ng
RT: 9.77 min Scan# 1263
Delta R.T. -0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	447755		
194	4.0	2.7	4.1	
164	9.9	8.2	12.2	





#51

Acenaphthene

Concen: 11.242 ng

RT: 10.10 min Scan# 1319

Delta R.T. -0.01 min

Lab File: BF108082.D

Acq: 10 Aug 2018 14:28

Instrument :

BNA_F


ClientSampleId :

2018-NYSEG-CLARK-AREA-10B




Tgt Ion:	154	Resp:	318941
Ion Ratio	Lower	Upper	
154	100		
153	115.4	91.2	136.8
152	54.0	43.8	65.8

Manual Integrations
APPROVED

Sohil
8/13/2018 3:51:45 PM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 2.206 ng


RT: 10.27 min Scan# 1348

Delta R.T. -0.01 min



Lab File: BF108082.D


Acq: 10 Aug 2018 14:28

Tgt Ion:	168	Resp:	90669
Ion Ratio	Lower	Upper	
168	100		
139	39.1	30.1	45.1
169	15.4	10.9	16.3



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#57
Fluorene
 Concen: 7.957 ng
 RT: 10.61 min Scan# 1407
 Delta R.T. -0.01 min
 Lab File: BF108082.D
 Acq: 10 Aug 2018 14:28

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Resp	Lower	Upper
166	100	258890		
165	100.7	79.8	119.8	
167	14.0	10.6	16.0	




Manual Integrations
 APPROVED


Sohil
 8/13/2018 3:51:45 PM



#63
Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.56 min Scan# 1568
 Delta R.T. 0.00 min
 Lab File: BF108082.D
 Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	879868		
94	9.5	8.5	12.7	
80	10.7	9.2	13.8	









#70
Phenanthrene
Concen: 58.434 ng
RT: 11.59 min Scan# 1573
Delta R.T. 0.00 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Lower	Upper
178	100		
176	22.9	15.8	23.8
179	17.8	13.0	19.6




Manual Integrations
APPROVED


Sohil
8/13/2018 3:51:45 PM



#71
Anthracene
Concen: 23.422 ng
RT: 11.64 min Scan# 1581
Delta R.T. 0.00 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.4	23.2
179	16.1	12.6	19.0









#72
 Carbazole
 Concen: 2.391 ng
 RT: 11.78 min Scan# 1606
 Delta R.T. -0.01 min
 Lab File: BF108082.D
 Acq: 10 Aug 2018 14:28

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Lower	Upper
167	100		
166	22.5	17.6	26.4
139	12.6	10.9	16.3




Manual Integrations
 APPROVED


Sohil
 8/13/2018 3:51:45 PM



#74
 Fluoranthene
 Concen: 67.746 ng
 RT: 12.79 min Scan# 1777
 Delta R.T. 0.01 min
 Lab File: BF108082.D
 Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.3	0.0	34.1
203	22.5	0.0	38.1





#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.22 min Scan# 2020

Delta R.T. 0.00 min

Lab File: BF108082.D

Acq: 10 Aug 2018 14:28

Instrument :

BNA_F


ClientSampleId :

2018-NYSEG-CLARK-AREA-10B




Tgt Ion:	240	Resp:	623690
Ion Ratio	Lower	Upper	
240	100		
120	10.8	9.5	14.3
236	26.4	20.7	31.1

Manual Integrations
APPROVED

Sohil
8/13/2018 3:51:45 PM



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 80.117 ng m


RT: 13.02 min Scan# 1817

Delta R.T. 0.01 min



Lab File: BF108082.D


Acq: 10 Aug 2018 14:28

Tgt Ion:	202	Resp:	3163490
Ion Ratio	Lower	Upper	
202	100		
200	26.7	17.4	26.2#
203	22.0	14.3	21.5#



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 87.388 ng

RT: 13.15 min Scan# 1839

Delta R.T. 0.01 min

Lab File: BF108082.D

Acq: 10 Aug 2018 14:28

Instrument :

BNA_F

ClientSampleId :


2018-NYSEG-CLARK-AREA-10B

Tgt Ion:244 Resp: 2143625

Ion	Ratio	Lower	Upper
244	100		
212	8.1	5.4	8.0#
122	10.3	11.0	16.4#

Manual Integrations
APPROVED




Sohil
8/13/2018 3:51:45 PM



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 40.032 ng m

RT: 14.21 min Scan# 2019

Delta R.T. 0.01 min

Lab File: BF108082.D

Acq: 10 Aug 2018 14:28




Tgt Ion:228 Resp: 1432865


Ion	Ratio	Lower	Upper
228	100		
226	48.2	22.6	33.8#
229	21.1	15.8	23.6

Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#82
Chrysene
Concen: 36.078 ng
RT: 14.25 min Scan# 2025
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Lower	Upper
228	100		
226	32.0	25.0	37.6
229	22.3	16.2	24.4




Manual Integrations
APPROVED


Sohil
8/13/2018 3:51:45 PM



#85
Indeno(1,2,3-cd)pyrene
Concen: 22.941 ng
RT: 17.41 min Scan# 2563
Delta R.T. 0.02 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Lower	Upper
276	100		
138	18.6	25.0	37.6#
277	24.5	20.1	30.1









#86
Perylene-d12
Concen: 20.000 ng
RT: 15.79 min Scan# 2287
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.7	19.2	28.8
265	22.1	17.5	26.3




Manual Integrations
APPROVED


Sohil
8/13/2018 3:51:45 PM



#87
Benzo(b)fluoranthene
Concen: 40.077 ng m
RT: 15.32 min Scan# 2207
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.0	17.0	25.6
125	10.1	9.0	13.6









#88
Benzo(k)fluoranthene
Concen: 14.562 ng m
RT: 15.34 min Scan# 2211
Delta R.T. 0.00 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.1	17.9	26.9
125	9.8	10.2	15.2




Manual Integrations
APPROVED


Sohil
8/13/2018 3:51:45 PM



#89
Benzo(a)pyrene
Concen: 37.978 ng
RT: 15.72 min Scan# 2276
Delta R.T. 0.02 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.3	17.1	25.7
125	10.3	9.8	14.6






#90
Dibenzo(a,h)anthracene
Concen: 5.367 ng
RT: 17.42 min Scan# 2565
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-10B




Tgt Ion:	278	Resp:	155667
Ion Ratio	Lower	Upper	
278	100		
139	16.7	15.9	23.9
279	25.7	19.0	28.4

Manual Integrations
APPROVED

Sohil
8/13/2018 3:51:45 PM




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#91
Benzo(g,h,i)perylene
Concen: 21.237 ng
RT: 17.91 min Scan# 2647
Delta R.T. 0.01 min
Lab File: BF108082.D
Acq: 10 Aug 2018 14:28

Tgt Ion:	276	Resp:	606513
Ion Ratio	Lower	Upper	
276	100		
277	23.7	17.9	26.9
138	20.3	21.8	32.8#



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

