

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081023\
 Data File : BF134712.D
 Acq On : 10 Aug 2023 16:49
 Operator : CG\JU
 Sample : 03932-01DL 100X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB-01-Z18-20DL

Quant Time: Aug 11 03:55:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	151894	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	604945	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	292702	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	474150	20.000	ng	# 0.00
76) Chrysene-d12	13.969	240	318757	20.000	ng	# 0.00
86) Perylene-d12	15.439	264	298958	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	3207	0.321	ng	0.00
7) Phenol-d6	6.434	99	3158	0.247	ng	0.00
23) Nitrobenzene-d5	7.263	82	652	0.067	ng	-0.11
42) 2,4,6-Tribromophenol	10.622	330	984	0.330	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	11074	0.629	ng	-0.01
79) Terphenyl-d14	12.910	244	10890	0.602	ng	-0.01
Target Compounds						
9) Benzaldehyde	6.357	77	3726	Below Cal		Qvalue # 1
18) Hexachloroethane	7.328	117	10827	2.951	ng	# 1
26) 2-Nitrophenol	7.710	139	318	4.857	ng	# 1
31) Naphthalene	8.104	128	2355742	77.483	ng	98
32) Benzoic acid	7.863	122	316	7.910	ng	# 1
33) 4-Chloroaniline	8.104	127	312709	23.083	ng	# 32
37) 2-Methylnaphthalene	8.792	142	791547	39.278	ng	100
38) 1-Methylnaphthalene	8.892	142	507509	27.082	ng	98
46) 1,1'-Biphenyl	9.251	154	96230	4.228	ng	96
48) 2-Nitroaniline	9.351	65	3333	3.752	ng	# 31
50) Dimethylphthalate	9.833	163	61857	3.101	ng	# 1
51) 2,6-Dinitrotoluene	9.528	165	1884	2.932	ng	# 78
52) Acenaphthene	9.869	154	321095	18.482	ng	98
54) 2,4-Dinitrophenol	9.922	184	61	6.830	ng	# 1
57) 2,4-Dinitrotoluene	10.039	165	2488	3.325	ng	# 61
58) Fluorene	10.381	166	148482	8.242	ng	93
65) 4,6-Dinitro-2-methylph...	10.445	198	72	8.730	ng	# 1
71) Phenanthrene	11.345	178	612555	24.322	ng	99
72) Anthracene	11.398	178	179199	6.966	ng	99
75) Fluoranthene	12.539	202	191837	7.211	ng	98
78) Pyrene	12.769	202	290686	10.539	ng	99
81) Benzo(a)anthracene	13.998	228	93905	4.120	ng	99
83) Chrysene	13.998	228	93905	4.229	ng	97
88) Benzo(b)fluoranthene	15.010	252	62800	3.444	ng	# 94
89) Benzo(k)fluoranthene	15.010	252	62800	3.458	ng	# 94
90) Benzo(a)pyrene	15.374	252	55577	3.269	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081023\
Data File : BF134712.D
Acq On : 10 Aug 2023 16:49
Operator : CG\JU
Sample : 03932-01DL 100X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL

Quant Time: Aug 11 03:55:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.798 min Scan# 800
 Delta R.T. 0.000 min
 Lab File: BF134712.D
 Acq: 10 Aug 2023 16:49




Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB-01-Z18-20DL


Tgt Ion:152 Resp: 151894
 Ion Ratio Lower Upper
 152 100
 150 155.5 128.3 192.5
 115 64.1 52.4 78.6



#5
 2-Fluorophenol
 Concen: 0.321 ng
 RT: 5.428 min Scan# 568
 Delta R.T. 0.000 min
 Lab File: BF134712.D
 Acq: 10 Aug 2023 16:49

Tgt Ion:112 Resp: 3207
 Ion Ratio Lower Upper
 112 100
 64 61.7 44.9 67.3
 63 21.1 25.4 38.0#





#7

Phenol-d6

Concen: 0.247 ng

RT: 6.434 min Scan# 71

Delta R.T. -0.006 min

Lab File: BF134712.D


Acq: 10 Aug 2023 16:49

Instrument :




BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL



Tgt Ion:	99	Resp:	3158
Ion	Ratio	Lower	Upper
99	100		
42	25.3	18.7	28.1
71	45.8	36.4	54.6



#18

Hexachloroethane

Concen: 2.951 ng




RT: 7.328 min Scan# 891


Delta R.T. 0.018 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Tgt Ion:	117	Resp:	10827
Ion	Ratio	Lower	Upper
117	100		
119	671.3	75.8	113.8#
201	0.0	73.6	110.4#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.081 min Scan# 1018

Delta R.T. 0.000 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL

Tgt Ion:136 Resp: 604945





Ion Ratio Lower Upper

136 100

137 11.2 8.4 12.6

54 8.1 5.2 7.8#

68 6.2 2.9 4.3#



#23

Nitrobenzene-d5

Concen: 0.067 ng

RT: 7.263 min Scan# 880

Delta R.T. -0.106 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49




Tgt Ion: 82 Resp: 652


Ion Ratio Lower Upper

82 100

128 50.2 34.3 51.5

54 45.1 43.8 65.6









#26
2-Nitrophenol
Concen: 4.857 ng
RT: 7.710 min Scan# 91
Delta R.T. 0.012 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:139 Resp: 318
Ion Ratio Lower Upper
139 100
109 378.0 29.8 44.6#
65 363.2 45.4 68.2#



#31
Naphthalene
Concen: 77.483 ng
RT: 8.104 min Scan# 1023
Delta R.T. 0.000 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:128 Resp: 2355742
Ion Ratio Lower Upper
128 100
129 11.6 8.4 12.6
127 13.6 10.7 16.1









#32
Benzoic acid
Concen: 7.910 ng
RT: 7.863 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:122 Resp: 316
Ion Ratio Lower Upper
122 100
105 84.8 87.5 127.5#
77 1542.0 71.4 111.4#



#33
4-Chloroaniline
Concen: 23.083 ng
RT: 8.104 min Scan# 1023
Delta R.T. -0.047 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:127 Resp: 312709
Ion Ratio Lower Upper
127 100
129 85.1 25.6 38.4#
65 0.0 25.4 38.0#
92 0.1 16.2 24.4#









#37
2-Methylnaphthalene
Concen: 39.278 ng
RT: 8.792 min Scan# 1139
Delta R.T. 0.000 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:142 Resp: 791547
Ion Ratio Lower Upper
142 100
141 88.7 70.8 106.2



#38
1-Methylnaphthalene
Concen: 27.082 ng
RT: 8.892 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:142 Resp: 507509
Ion Ratio Lower Upper
142 100
141 91.2 71.4 107.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.833 min Scan# 1

Delta R.T. 0.000 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL





Tgt Ion:164 Resp: 292702

Ion Ratio Lower Upper

164 100

162 101.7 82.6 123.8

160 46.2 37.2 55.8



#42

2,4,6-Tribromophenol

Concen: 0.330 ng

RT: 10.622 min Scan# 1451

Delta R.T. -0.006 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49




Tgt Ion:330 Resp: 984


Ion Ratio Lower Upper

330 100

332 102.2 78.8 118.2

141 86.9 25.4 38.2#









#45
2-Fluorobiphenyl
Concen: 0.629 ng
RT: 9.151 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:172 Resp: 11074
Ion Ratio Lower Upper
172 100
171 34.3 28.8 43.2
170 21.8 18.7 28.1



#46
1,1'-Biphenyl
Concen: 4.228 ng
RT: 9.251 min Scan# 1218
Delta R.T. -0.012 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:154 Resp: 96230
Ion Ratio Lower Upper
154 100
153 38.9 20.8 60.8
76 13.0 0.0 30.2





#48

2-Nitroaniline

Concen: 3.752 ng

RT: 9.351 min Scan# 1317

Delta R.T. -0.029 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL





Tgt Ion: 65 Resp: 3333

Ion Ratio Lower Upper

65 100

92 11.3 50.7 76.1#

138 28.8 78.2 117.4#



#50

Dimethylphthalate

Concen: 3.101 ng

RT: 9.833 min Scan# 1317

Delta R.T. 0.265 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49




Tgt Ion:163 Resp: 61857


Ion Ratio Lower Upper

163 100

194 0.0 3.3 4.9#

164 490.2 7.5 11.3#









#51
2,6-Dinitrotoluene
Concen: 2.932 ng
RT: 9.528 min Scan# 11
Delta R.T. -0.100 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:165 Resp: 1884
Ion Ratio Lower Upper
165 100
63 83.1 57.8 86.6
89 40.3 51.8 77.8#



#52
Acenaphthene
Concen: 18.482 ng
RT: 9.869 min Scan# 1323
Delta R.T. -0.006 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:154 Resp: 321095
Ion Ratio Lower Upper
154 100
153 113.9 90.0 135.0
152 50.6 42.4 63.6









#54
2,4-Dinitrophenol
Concen: 6.830 ng
RT: 9.922 min Scan# 11
Delta R.T. 0.024 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:184 Resp: 61
Ion Ratio Lower Upper
184 100
63 279.3 57.1 85.7#
154 796.0 59.1 88.7#



#57
2,4-Dinitrotoluene
Concen: 3.325 ng
RT: 10.039 min Scan# 1352
Delta R.T. 0.012 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:165 Resp: 2488
Ion Ratio Lower Upper
165 100
63 105.8 48.9 73.3#
89 57.8 61.7 92.5#
182 19.9 2.0 3.0#









#58
 Fluorene
 Concen: 8.242 ng
 RT: 10.381 min Scan# 1410
 Delta R.T. -0.006 min
 Lab File: BF134712.D
 Acq: 10 Aug 2023 16:49




Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB-01-Z18-20DL


Tgt Ion:166 Resp: 148482
 Ion Ratio Lower Upper
 166 100
 165 100.5 74.5 111.7
 167 15.0 10.6 16.0



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 11.322 min Scan# 1570
 Delta R.T. 0.000 min
 Lab File: BF134712.D
 Acq: 10 Aug 2023 16:49

Tgt Ion:188 Resp: 474150
 Ion Ratio Lower Upper
 188 100
 94 10.5 5.1 7.7#
 80 11.4 6.6 9.8#









#65
4,6-Dinitro-2-methylphenol
Concen: 8.730 ng
RT: 10.445 min Scan# 1417
Delta R.T. 0.006 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:198 Resp: 72
Ion Ratio Lower Upper
198 100
51 972.9 38.6 78.6#
105 560.6 27.9 67.9#



#71
Phenanthrene
Concen: 24.322 ng
RT: 11.345 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Tgt Ion:178 Resp: 612555
Ion Ratio Lower Upper
178 100
176 18.9 15.8 23.6
179 15.5 12.7 19.1









#72
 Anthracene
 Concen: 6.966 ng
 RT: 11.398 min Scan# 1582
 Delta R.T. -0.006 min
 Lab File: BF134712.D
 Acq: 10 Aug 2023 16:49




Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB-01-Z18-20DL


Tgt Ion:178 Resp: 179199
 Ion Ratio Lower Upper
 178 100
 176 17.9 14.6 21.8
 179 15.7 12.3 18.5



#75
 Fluoranthene
 Concen: 7.211 ng
 RT: 12.539 min Scan# 1777
 Delta R.T. 0.000 min
 Lab File: BF134712.D
 Acq: 10 Aug 2023 16:49

Tgt Ion:202 Resp: 191837
 Ion Ratio Lower Upper
 202 100
 101 10.4 0.0 28.5
 203 17.3 0.0 37.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.969 min Scan# 20

Delta R.T. -0.006 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL





Tgt Ion:240 Resp: 318757

Ion Ratio Lower Upper

240 100

120 11.7 6.6 10.0#

236 26.0 20.2 30.2



#78

Pyrene

Concen: 10.539 ng

RT: 12.769 min Scan# 1816

Delta R.T. 0.000 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49




Tgt Ion:202 Resp: 290686


Ion Ratio Lower Upper

202 100

200 20.2 15.8 23.8

203 17.1 13.5 20.3





#79

Terphenyl-d14

Concen: 0.602 ng

RT: 12.910 min Scan# 1840

Delta R.T. -0.012 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL





Tgt Ion:244 Resp: 10890

Ion Ratio Lower Upper

244 100

212 6.0 5.6 8.4

122 10.8 8.0 12.0



#81

Benzo(a)anthracene

Concen: 4.120 ng

RT: 13.998 min Scan# 2025

Delta R.T. 0.035 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49




Tgt Ion:228 Resp: 93905


Ion Ratio Lower Upper

228 100

226 28.4 22.3 33.5

229 20.0 15.8 23.6





#83

Chrysene

Concen: 4.229 ng

RT: 13.998 min Scan# 20

Delta R.T. 0.000 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB-01-Z18-20DL





Tgt Ion:228 Resp: 93905

Ion Ratio Lower Upper

228 100

226 28.4 24.5 36.7

229 20.0 15.8 23.8



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.439 min Scan# 2270

Delta R.T. 0.000 min

Lab File: BF134712.D

Acq: 10 Aug 2023 16:49




Tgt Ion:264 Resp: 298958


Ion Ratio Lower Upper

264 100

260 24.0 18.4 27.6

265 21.6 15.8 23.8









#88
Benzo(b)fluoranthene
Concen: 3.444 ng
RT: 15.010 min Scan# 2197
Delta R.T. 0.000 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL


Tgt Ion:252 Resp: 62800
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 15.1 6.0 9.0#



#89
Benzo(k)fluoranthene
Concen: 3.458 ng
RT: 15.010 min Scan# 2197
Delta R.T. -0.029 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49


Tgt Ion:252 Resp: 62800
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 15.1 6.3 9.5#





#90
Benzo(a)pyrene
Concen: 3.269 ng
RT: 15.374 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF134712.D
Acq: 10 Aug 2023 16:49

Instrument :
BNA_F
ClientSampleId :
NEVINS-SB-01-Z18-20DL



Tgt Ion:252 Resp: 55577
Ion Ratio Lower Upper
252 100
253 22.9 17.0 25.4
125 12.4 6.2 9.4#

