

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081023\
 Data File : BF134714.D
 Acq On : 10 Aug 2023 17:55
 Operator : CG\JU
 Sample : 03954-19 2X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-21

Quant Time: Aug 11 03:56:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	157190	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	597884	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	292335	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	487688	20.000	ng	# 0.00
76) Chrysene-d12	13.974	240	309050	20.000	ng	# 0.00
86) Perylene-d12	15.445	264	299978	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	446039	43.134	ng	0.00
7) Phenol-d6	6.434	99	542744	41.059	ng	0.00
23) Nitrobenzene-d5	7.357	82	334201	34.932	ng	-0.01
42) 2,4,6-Tribromophenol	10.622	330	136162	45.691	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	669124	38.057	ng	-0.01
79) Terphenyl-d14	12.916	244	621421	35.447	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.357	77	1513	Below Cal	#	1
19) n-Nitroso-di-n-propyla...	7.357	70	44795	5.660	ng	# 73
25) Isophorone	7.357	82	332065	15.913	ng	# 65
31) Naphthalene	8.098	128	80782	2.688	ng	# 98
32) Benzoic acid	7.863	122	557	7.944	ng	# 22
48) 2-Nitroaniline	9.351	65	226	3.206	ng	# 8
50) Dimethylphthalate	9.833	163	61081	3.066	ng	# 1
51) 2,6-Dinitrotoluene	9.833	165	39551	11.680	ng	# 17
52) Acenaphthene	9.863	154	78759	4.539	ng	# 97
55) Dibenzofuran	10.039	168	85238	3.451	ng	# 90
56) 4-Nitrophenol	10.033	139	29714	8.181	ng	# 1
57) 2,4-Dinitrotoluene	9.992	165	1317	3.102	ng	# 43
58) Fluorene	10.381	166	99893	5.552	ng	# 94
65) 4,6-Dinitro-2-methylph...	10.622	198	183	8.775	ng	# 1
71) Phenanthrene	11.351	178	1103981	42.618	ng	# 99
72) Anthracene	11.398	178	290820	10.991	ng	# 99
73) Carbazole	11.551	167	128877	5.455	ng	# 99
75) Fluoranthene	12.539	202	1226136	44.810	ng	# 95
78) Pyrene	12.769	202	1028749	38.469	ng	# 98
81) Benzo(a)anthracene	13.963	228	549804	24.880	ng	# 99
83) Chrysene	13.998	228	450911	20.944	ng	# 96
87) Indeno(1,2,3-cd)pyrene	16.933	276	188058	10.659	ng	# 95
88) Benzo(b)fluoranthene	15.015	252	707443	38.668	ng	# 95
89) Benzo(k)fluoranthene	15.015	252	707443	38.817	ng	# 96
90) Benzo(a)pyrene	15.380	252	354591	20.787	ng	# 94
91) Dibenzo(a,h)anthracene	16.945	278	49614	3.480	ng	# 88
92) Benzo(g,h,i)perylene	17.380	276	163827	11.264	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081023\
Data File : BF134714.D
Acq On : 10 Aug 2023 17:55
Operator : CG\JU
Sample : 03954-19 2X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-21

Quant Time: Aug 11 03:56:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.798 min Scan# 800
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:152 Resp: 157190
Ion Ratio Lower Upper
152 100
150 157.9 128.3 192.5
115 62.1 52.4 78.6



#5
2-Fluorophenol
Concen: 43.134 ng
RT: 5.428 min Scan# 568
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:112 Resp: 446039
Ion Ratio Lower Upper
112 100
64 56.8 44.9 67.3
63 28.9 25.4 38.0









#7
Phenol-d6
Concen: 41.059 ng
RT: 6.434 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion: 99 Resp: 542744
Ion Ratio Lower Upper
99 100
42 17.0 18.7 28.1#
71 31.5 36.4 54.6#



#19
n-Nitroso-di-n-propylamine
Concen: 5.660 ng
RT: 7.357 min Scan# 896
Delta R.T. 0.135 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion: 70 Resp: 44795
Ion Ratio Lower Upper
70 100
42 48.1 54.7 82.1#
101 0.0 8.2 12.2#
130 2.6 15.1 22.7#





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.075 min Scan# 1018

Delta R.T. -0.006 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55

Instrument :

BNA_F

ClientSampleId :

WC-21

Tgt Ion:136 Resp: 597884





Ion Ratio Lower Upper

136 100

137 11.3 8.4 12.6

54 9.0 5.2 7.8#

68 7.2 2.9 4.3#



#23

Nitrobenzene-d5

Concen: 34.932 ng

RT: 7.357 min Scan# 896

Delta R.T. -0.012 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55




Tgt Ion: 82 Resp: 334201


Ion Ratio Lower Upper

82 100

128 46.1 34.3 51.5

54 52.0 43.8 65.6









#25
Isophorone
Concen: 15.913 ng
RT: 7.357 min Scan# 896
Delta R.T. -0.270 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion: 82 Resp: 332065
Ion Ratio Lower Upper
82 100
95 0.0 6.0 9.0#
138 0.0 14.6 22.0#



#31
Naphthalene
Concen: 2.688 ng
RT: 8.098 min Scan# 1022
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion: 128 Resp: 80782
Ion Ratio Lower Upper
128 100
129 11.0 8.4 12.6
127 12.6 10.7 16.1









#32
Benzoic acid
Concen: 7.944 ng
RT: 7.863 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:122 Resp: 557
Ion Ratio Lower Upper
122 100
105 0.0 87.5 127.5#
77 46.3 71.4 111.4#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.833 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:164 Resp: 292335
Ion Ratio Lower Upper
164 100
162 102.2 82.6 123.8
160 45.6 37.2 55.8





#42

2,4,6-Tribromophenol

Concen: 45.691 ng

RT: 10.622 min Scan# 1451

Delta R.T. -0.006 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55

Instrument :

BNA_F

ClientSampleId :

WC-21





Tgt Ion:330 Resp: 136162

Ion Ratio Lower Upper

330 100

332 95.8 78.8 118.2

141 30.4 25.4 38.2



#45

2-Fluorobiphenyl

Concen: 38.057 ng

RT: 9.151 min Scan# 1201

Delta R.T. -0.012 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55




Tgt Ion:172 Resp: 669124


Ion Ratio Lower Upper

172 100

171 34.9 28.8 43.2

170 23.3 18.7 28.1









#48
2-Nitroaniline
Concen: 3.206 ng
RT: 9.351 min Scan# 11
Delta R.T. -0.029 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion: 65 Resp: 226
Ion Ratio Lower Upper
65 100
92 0.0 50.7 76.1#
138 0.0 78.2 117.4#



#50
Dimethylphthalate
Concen: 3.066 ng
RT: 9.833 min Scan# 1317
Delta R.T. 0.265 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion: 163 Resp: 61081
Ion Ratio Lower Upper
163 100
194 0.0 3.3 4.9#
164 470.4 7.5 11.3#









#51
2,6-Dinitrotoluene
Concen: 11.680 ng
RT: 9.833 min Scan# 1317
Delta R.T. 0.206 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:165 Resp: 39551
Ion Ratio Lower Upper
165 100
63 1.6 57.8 86.6#
89 1.7 51.8 77.8#



#52
Acenaphthene
Concen: 4.539 ng
RT: 9.863 min Scan# 1322
Delta R.T. -0.012 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:154 Resp: 78759
Ion Ratio Lower Upper
154 100
153 114.4 90.0 135.0
152 48.4 42.4 63.6









#55
Dibenzenofuran
Concen: 3.451 ng
RT: 10.039 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:168 Resp: 85238
Ion Ratio Lower Upper
168 100
139 34.5 33.4 50.2
169 15.2 10.6 15.8



#56
4-Nitrophenol
Concen: 8.181 ng
RT: 10.033 min Scan# 1351
Delta R.T. 0.082 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:139 Resp: 29714
Ion Ratio Lower Upper
139 100
109 0.9 76.9 116.9#
65 2.1 98.5 138.5#









#57
2,4-Dinitrotoluene
Concen: 3.102 ng
RT: 9.992 min Scan# 11
Delta R.T. -0.035 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:165 Resp: 1317
Ion Ratio Lower Upper
165 100
63 30.1 48.9 73.3#
89 14.6 61.7 92.5#
182 0.0 2.0 3.0#



#58
Fluorene
Concen: 5.552 ng
RT: 10.381 min Scan# 1410
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:166 Resp: 99893
Ion Ratio Lower Upper
166 100
165 99.4 74.5 111.7
167 14.3 10.6 16.0





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.322 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55

Instrument :

BNA_F

ClientSampleId :

WC-21





Tgt Ion:188 Resp: 487688

Ion Ratio Lower Upper

188 100

94 10.4 5.1 7.7#

80 11.0 6.6 9.8#



#65

4,6-Dinitro-2-methylphenol

Concen: 8.775 ng

RT: 10.622 min Scan# 1451

Delta R.T. 0.182 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55




Tgt Ion:198 Resp: 183


Ion Ratio Lower Upper

198 100

51 596.3 38.6 78.6#

105 313.9 27.9 67.9#









#71
Phenanthrene
Concen: 42.618 ng
RT: 11.351 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:178 Resp: 1103981
Ion Ratio Lower Upper
178 100
176 19.7 15.8 23.6
179 15.4 12.7 19.1



#72
Anthracene
Concen: 10.991 ng
RT: 11.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:178 Resp: 290820
Ion Ratio Lower Upper
178 100
176 18.9 14.6 21.8
179 15.9 12.3 18.5









#73
Carbazole
Concen: 5.455 ng
RT: 11.551 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:167 Resp: 128877
Ion Ratio Lower Upper
167 100
166 21.9 17.5 26.3
139 12.8 11.0 16.4



#75
Fluoranthene
Concen: 44.810 ng
RT: 12.539 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:202 Resp: 1226136
Ion Ratio Lower Upper
202 100
101 12.4 0.0 28.5
203 18.3 0.0 37.5





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.974 min Scan# 20

Delta R.T. 0.000 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55

Instrument :

BNA_F

ClientSampleId :

WC-21





Tgt Ion:240 Resp: 309050

Ion Ratio Lower Upper

240 100

120 10.8 6.6 10.0#

236 25.5 20.2 30.2



#78

Pyrene

Concen: 38.469 ng

RT: 12.769 min Scan# 1816

Delta R.T. 0.000 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55




Tgt Ion:202 Resp: 1028749


Ion Ratio Lower Upper

202 100

200 20.7 15.8 23.8

203 17.4 13.5 20.3









#79
 Terphenyl-d14
 Concen: 35.447 ng
 RT: 12.916 min Scan# 1840
 Delta R.T. -0.006 min
 Lab File: BF134714.D
 Acq: 10 Aug 2023 17:55




Instrument :
 BNA_F
 ClientSampleId :
 WC-21


Tgt Ion:244 Resp: 621421
 Ion Ratio Lower Upper
 244 100
 212 5.9 5.6 8.4
 122 10.8 8.0 12.0



#81
 Benzo(a)anthracene
 Concen: 24.880 ng
 RT: 13.963 min Scan# 2019
 Delta R.T. 0.000 min
 Lab File: BF134714.D
 Acq: 10 Aug 2023 17:55

Tgt Ion:228 Resp: 549804
 Ion Ratio Lower Upper
 228 100
 226 27.6 22.3 33.5
 229 20.6 15.8 23.6









#83
Chrysene
Concen: 20.944 ng
RT: 13.998 min Scan# 20
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:228 Resp: 450911
Ion Ratio Lower Upper
228 100
226 29.6 24.5 36.7
229 22.6 15.8 23.8



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.445 min Scan# 2271
Delta R.T. 0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:264 Resp: 299978
Ion Ratio Lower Upper
264 100
260 22.9 18.4 27.6
265 23.2 15.8 23.8





#87

Indeno(1,2,3-cd)pyrene

Concen: 10.659 ng

RT: 16.933 min Scan# 2198

Delta R.T. 0.006 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55

Instrument :

BNA_F

ClientSampleId :

WC-21





Tgt Ion:276 Resp: 188058

Ion Ratio Lower Upper

276 100

138 22.7 14.2 21.4#

277 24.5 19.9 29.9



#88

Benzo(b)fluoranthene

Concen: 38.668 ng

RT: 15.015 min Scan# 2198

Delta R.T. 0.006 min

Lab File: BF134714.D

Acq: 10 Aug 2023 17:55




Tgt Ion:252 Resp: 707443


Ion Ratio Lower Upper

252 100

253 23.1 17.3 25.9

125 11.7 6.0 9.0#









#89
Benzo(k)fluoranthene
Concen: 38.817 ng
RT: 15.015 min Scan# 2199
Delta R.T. -0.023 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55




Instrument :
BNA_F
ClientSampleId :
WC-21


Tgt Ion:252 Resp: 707443
Ion Ratio Lower Upper
252 100
253 23.1 17.9 26.9
125 11.7 6.3 9.5#



#90
Benzo(a)pyrene
Concen: 20.787 ng
RT: 15.380 min Scan# 2260
Delta R.T. 0.000 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:252 Resp: 354591
Ion Ratio Lower Upper
252 100
253 23.3 17.0 25.4
125 11.6 6.2 9.4#









#91
Dibenzo(a,h)anthracene
Concen: 3.480 ng
RT: 16.945 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Instrument :
BNA_F
ClientSampleId :
WC-21

Tgt Ion:278 Resp: 49614
Ion Ratio Lower Upper
278 100
139 18.3 10.0 15.0#
279 29.6 19.7 29.5#



#92
Benzo(g,h,i)perylene
Concen: 11.264 ng
RT: 17.380 min Scan# 2600
Delta R.T. 0.006 min
Lab File: BF134714.D
Acq: 10 Aug 2023 17:55

Tgt Ion:276 Resp: 163827
Ion Ratio Lower Upper
276 100
277 22.9 19.0 28.6
138 23.6 12.2 18.2#

