

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
 Data File : BF134739.D
 Acq On : 11 Aug 2023 18:42
 Operator : CG\JU
 Sample : 03974-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EO-01-081023

Quant Time: Aug 11 22:23:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	194885	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	721385	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	304030	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	478458	20.000	ng	# 0.00
76) Chrysene-d12	13.980	240	328763	20.000	ng	# 0.00
86) Perylene-d12	15.451	264	258819	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	474604	37.019	ng	0.00
7) Phenol-d6	6.428	99	565129	34.483	ng	-0.01
23) Nitrobenzene-d5	7.357	82	337316	29.221	ng	-0.01
42) 2,4,6-Tribromophenol	10.622	330	116214	37.497	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	630309	34.471	ng	-0.01
79) Terphenyl-d14	12.916	244	583594	31.293	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.357	77	217	Below Cal	#	26
19) n-Nitroso-di-n-propyla...	7.357	70	43970	4.481	ng	# 72
25) Isophorone	7.357	82	333925	13.263	ng	# 65
26) 2-Nitrophenol	7.692	139	267	4.838	ng	# 1
32) Benzoic acid	7.869	122	523	7.927	ng	# 53
48) 2-Nitroaniline	9.404	65	1992	3.503	ng	# 57
49) Acenaphthylene	9.692	152	93555	3.360	ng	98
50) Dimethylphthalate	9.833	163	65581	3.165	ng	# 1
51) 2,6-Dinitrotoluene	9.675	165	1627	2.859	ng	# 44
52) Acenaphthene	9.863	154	88672	4.914	ng	98
54) 2,4-Dinitrophenol	9.998	184	54	6.823	ng	# 1
55) Dibenzofuran	10.039	168	93594	3.643	ng	# 92
57) 2,4-Dinitrotoluene	10.039	165	1988	3.215	ng	# 1
58) Fluorene	10.380	166	167406	8.947	ng	98
65) 4,6-Dinitro-2-methylph...	10.480	198	100	8.741	ng	# 1
71) Phenanthrene	11.351	178	1285038	50.565	ng	100
72) Anthracene	11.398	178	468178	18.035	ng	99
73) Carbazole	11.551	167	78493	3.387	ng	97
75) Fluoranthene	12.545	202	2376075	88.511	ng	96
78) Pyrene	12.774	202	2142986	75.329	ng	96
81) Benzo(a)anthracene	13.968	228	1239163	52.712	ng	98
83) Chrysene	14.004	228	983390	42.937	ng	97
87) Indeno(1,2,3-cd)pyrene	16.945	276	458616	30.129	ng	# 94
88) Benzo(b)fluoranthene	15.021	252	1538022	97.435	ng	# 95
89) Benzo(k)fluoranthene	15.021	252	1538022	97.810	ng	# 97
90) Benzo(a)pyrene	15.392	252	780170	53.009	ng	# 96
91) Dibenzo(a,h)anthracene	16.956	278	111068	9.029	ng	# 88
92) Benzo(g,h,i)perylene	17.398	276	406222	32.372	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
Data File : BF134739.D
Acq On : 11 Aug 2023 18:42
Operator : CG\JU
Sample : 03974-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EO-01-081023

Quant Time: Aug 11 22:23:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.798 min Scan# 800
Delta R.T. 0.000 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:152 Resp: 194885
Ion Ratio Lower Upper
152 100
150 157.3 128.3 192.5
115 64.2 52.4 78.6



#5
2-Fluorophenol
Concen: 37.019 ng
RT: 5.422 min Scan# 567
Delta R.T. -0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:112 Resp: 474604
Ion Ratio Lower Upper
112 100
64 62.3 44.9 67.3
63 31.6 25.4 38.0









#7
Phenol-d6
Concen: 34.483 ng
RT: 6.428 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion: 99 Resp: 565129
Ion Ratio Lower Upper
99 100
42 18.6 18.7 28.1#
71 32.5 36.4 54.6#



#19
n-Nitroso-di-n-propylamine
Concen: 4.481 ng
RT: 7.357 min Scan# 896
Delta R.T. 0.135 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion: 70 Resp: 43970
Ion Ratio Lower Upper
70 100
42 47.8 54.7 82.1#
101 0.0 8.2 12.2#
130 2.1 15.1 22.7#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.075 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:136 Resp: 721385
Ion Ratio Lower Upper
136 100
137 11.2 8.4 12.6
54 9.1 5.2 7.8#
68 7.1 2.9 4.3#



#23
Nitrobenzene-d5
Concen: 29.221 ng
RT: 7.357 min Scan# 896
Delta R.T. -0.012 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion: 82 Resp: 337316
Ion Ratio Lower Upper
82 100
128 45.6 34.3 51.5
54 52.5 43.8 65.6









#25
Isophorone
Concen: 13.263 ng
RT: 7.357 min Scan# 896
Delta R.T. -0.271 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion: 82 Resp: 333925
Ion Ratio Lower Upper
82 100
95 0.2 6.0 9.0#
138 0.0 14.6 22.0#



#26
2-Nitrophenol
Concen: 4.838 ng
RT: 7.692 min Scan# 953
Delta R.T. -0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:139 Resp: 267
Ion Ratio Lower Upper
139 100
109 315.5 29.8 44.6#
65 75.3 45.4 68.2#









#32
Benzoic acid
Concen: 7.927 ng
RT: 7.869 min Scan# 91
Delta R.T. 0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:122 Resp: 523
Ion Ratio Lower Upper
122 100
105 40.5 87.5 127.5#
77 117.0 71.4 111.4#



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.833 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:164 Resp: 304030
Ion Ratio Lower Upper
164 100
162 103.0 82.6 123.8
160 46.1 37.2 55.8





#42

2,4,6-Tribromophenol

Concen: 37.497 ng

RT: 10.622 min Scan# 1451

Delta R.T. -0.006 min

Lab File: BF134739.D


Acq: 11 Aug 2023 18:42

Instrument :

BNA_F

ClientSampleId :

EO-01-081023






Tgt Ion:330 Resp: 116214

Ion Ratio Lower Upper

330 100

332 96.0 78.8 118.2

141 29.5 25.4 38.2



#45

2-Fluorobiphenyl

Concen: 34.471 ng

RT: 9.151 min Scan# 1201

Delta R.T. -0.012 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42




Tgt Ion:172 Resp: 630309


Ion Ratio Lower Upper

172 100

171 35.8 28.8 43.2

170 23.4 18.7 28.1





#48

2-Nitroaniline

Concen: 3.503 ng

RT: 9.404 min Scan# 1244

Delta R.T. 0.024 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42





Instrument :

BNA_F

ClientSampleId :

EO-01-081023

Tgt Ion:	65	Resp:	1992
Ion	Ratio	Lower	Upper
65	100		
92	68.0	50.7	76.1
138	32.1	78.2	117.4#



#49

Acenaphthylene

Concen: 3.360 ng




RT: 9.692 min Scan# 1293


Delta R.T. -0.006 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42

Tgt Ion:	152	Resp:	93555
Ion	Ratio	Lower	Upper
152	100		
151	19.9	16.4	24.6
153	14.9	10.9	16.3









#50
Dimethylphthalate
Concen: 3.165 ng
RT: 9.833 min Scan# 11
Delta R.T. 0.265 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:163 Resp: 65581
Ion Ratio Lower Upper
163 100
194 0.3 3.3 4.9#
164 482.3 7.5 11.3#



#51
2,6-Dinitrotoluene
Concen: 2.859 ng
RT: 9.675 min Scan# 1290
Delta R.T. 0.047 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:165 Resp: 1627
Ion Ratio Lower Upper
165 100
63 29.2 57.8 86.6#
89 16.5 51.8 77.8#









#52
Acenaphthene
Concen: 4.914 ng
RT: 9.863 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:154 Resp: 88672
Ion Ratio Lower Upper
154 100
153 112.1 90.0 135.0
152 49.1 42.4 63.6



#54
2,4-Dinitrophenol
Concen: 6.823 ng
RT: 9.998 min Scan# 1345
Delta R.T. 0.100 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:184 Resp: 54
Ion Ratio Lower Upper
184 100
63 392.2 57.1 85.7#
154 388.9 59.1 88.7#









#55
Dibenzenofuran
Concen: 3.643 ng
RT: 10.039 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:168 Resp: 93594
Ion Ratio Lower Upper
168 100
139 36.6 33.4 50.2
169 16.2 10.6 15.8#



#57
2,4-Dinitrotoluene
Concen: 3.215 ng
RT: 10.039 min Scan# 1352
Delta R.T. 0.012 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:165 Resp: 1988
Ion Ratio Lower Upper
165 100
63 241.5 48.9 73.3#
89 134.4 61.7 92.5#
182 15.7 2.0 3.0#









#58
Fluorene
Concen: 8.947 ng
RT: 10.380 min Scan# 1410
Delta R.T. -0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:166 Resp: 167406
Ion Ratio Lower Upper
166 100
165 95.0 74.5 111.7
167 13.8 10.6 16.0



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.322 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:188 Resp: 478458
Ion Ratio Lower Upper
188 100
94 10.3 5.1 7.7#
80 11.1 6.6 9.8#









#65
4,6-Dinitro-2-methylphenol
Concen: 8.741 ng
RT: 10.480 min Scan# 1417
Delta R.T. 0.041 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:198 Resp: 100
Ion Ratio Lower Upper
198 100
51 324.4 38.6 78.6#
105 992.2 27.9 67.9#



#71
Phenanthrene
Concen: 50.565 ng
RT: 11.351 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:178 Resp: 1285038
Ion Ratio Lower Upper
178 100
176 19.7 15.8 23.6
179 15.8 12.7 19.1









#72
 Anthracene
 Concen: 18.035 ng
 RT: 11.398 min Scan# 1582
 Delta R.T. -0.006 min
 Lab File: BF134739.D
 Acq: 11 Aug 2023 18:42




Instrument :
 BNA_F
 ClientSampleId :
 EO-01-081023


Tgt Ion:178 Resp: 468178
 Ion Ratio Lower Upper
 178 100
 176 18.7 14.6 21.8
 179 16.0 12.3 18.5



#73
 Carbazole
 Concen: 3.387 ng
 RT: 11.551 min Scan# 1609
 Delta R.T. -0.006 min
 Lab File: BF134739.D
 Acq: 11 Aug 2023 18:42

Tgt Ion:167 Resp: 78493
 Ion Ratio Lower Upper
 167 100
 166 23.5 17.5 26.3
 139 14.8 11.0 16.4









#75
Fluoranthene
Concen: 88.511 ng
RT: 12.545 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:202 Resp: 2376075
Ion Ratio Lower Upper
202 100
101 11.2 0.0 28.5
203 18.7 0.0 37.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.980 min Scan# 2022
Delta R.T. 0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:240 Resp: 328763
Ion Ratio Lower Upper
240 100
120 10.2 6.6 10.0#
236 25.6 20.2 30.2





#78

Pyrene

Concen: 75.329 ng

RT: 12.774 min Scan# 1817

Delta R.T. 0.006 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42

Instrument :

BNA_F

ClientSampleId :

EO-01-081023





Tgt Ion:202 Resp: 2142986

Ion Ratio Lower Upper

202 100

200 22.2 15.8 23.8

203 18.2 13.5 20.3



#79

Terphenyl-d14

Concen: 31.293 ng

RT: 12.916 min Scan# 1841

Delta R.T. -0.006 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42




Tgt Ion:244 Resp: 583594


Ion Ratio Lower Upper

244 100

212 6.2 5.6 8.4

122 10.5 8.0 12.0









#81
Benzo(a)anthracene
Concen: 52.712 ng
RT: 13.968 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:228 Resp: 1239163
Ion Ratio Lower Upper
228 100
226 29.0 22.3 33.5
229 20.6 15.8 23.6



#83
Chrysene
Concen: 42.937 ng
RT: 14.004 min Scan# 2026
Delta R.T. 0.006 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:228 Resp: 983390
Ion Ratio Lower Upper
228 100
226 30.8 24.5 36.7
229 22.9 15.8 23.8





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.451 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42

Instrument :

BNA_F

ClientSampleId :

EO-01-081023





Tgt Ion:264 Resp: 258819

Ion Ratio Lower Upper

264 100

260 23.9 18.4 27.6

265 23.6 15.8 23.8



#87

Indeno(1,2,3-cd)pyrene

Concen: 30.129 ng

RT: 16.945 min Scan# 2526

Delta R.T. 0.018 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42




Tgt Ion:276 Resp: 458616


Ion Ratio Lower Upper

276 100

138 21.7 14.2 21.4#

277 23.0 19.9 29.9









#88
Benzo(b)fluoranthene
Concen: 97.435 ng
RT: 15.021 min Scan# 2199
Delta R.T. 0.012 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42




Instrument :
BNA_F
ClientSampleId :
EO-01-081023


Tgt Ion:252 Resp: 1538022
Ion Ratio Lower Upper
252 100
253 22.7 17.3 25.9
125 11.5 6.0 9.0#



#89
Benzo(k)fluoranthene
Concen: 97.810 ng
RT: 15.021 min Scan# 2199
Delta R.T. -0.018 min
Lab File: BF134739.D
Acq: 11 Aug 2023 18:42

Tgt Ion:252 Resp: 1538022
Ion Ratio Lower Upper
252 100
253 22.7 17.9 26.9
125 11.5 6.3 9.5#





#90

Benzo(a)pyrene

Concen: 53.009 ng

RT: 15.392 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42

Instrument :

BNA_F

ClientSampleId :

EO-01-081023





Tgt Ion:252 Resp: 780170

Ion Ratio Lower Upper

252 100

253 22.8 17.0 25.4

125 10.2 6.2 9.4#



#91

Dibenzo(a,h)anthracene

Concen: 9.029 ng

RT: 16.956 min Scan# 2528

Delta R.T. 0.006 min

Lab File: BF134739.D

Acq: 11 Aug 2023 18:42




Tgt Ion:278 Resp: 111068


Ion Ratio Lower Upper

278 100

139 18.2 10.0 15.0#

279 29.8 19.7 29.5#





#92

Benzo(g,h,i)perylene

Concen: 32.372 ng

RT: 17.398 min Scan# 20

Delta R.T. 0.024 min

Lab File: BF134739.D


Acq: 11 Aug 2023 18:42

Instrument :

BNA_F

ClientSampleId :

EO-01-081023



Tgt Ion:276 Resp: 406222

Ion Ratio Lower Upper

276 100

277 24.6 19.0 28.6

138 22.8 12.2 18.2#

