

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
 Data File : BF134723.D
 Acq On : 11 Aug 2023 09:52
 Operator : CG\JU
 Sample : 03954-16
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-20

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/12/2023
 Supervised By :mohammad ahmed 08/12/2023

Quant Time: Aug 11 22:19:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	173735	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	672850	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	308042	20.000	ng	0.00
64) Phenanthrene-d10	11.321	188	484957	20.000	ng	# 0.00
76) Chrysene-d12	13.968	240	338861	20.000	ng	# 0.00
86) Perylene-d12	15.439	264	283892	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	960076	84.002	ng	0.00
7) Phenol-d6	6.439	99	1209212	82.766	ng	0.00
23) Nitrobenzene-d5	7.363	82	727789	67.595	ng	0.00
42) 2,4,6-Tribromophenol	10.621	330	271354	86.413	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1303486	70.357	ng	0.00
79) Terphenyl-d14	12.915	244	1169798	60.857	ng	0.00
Target Compounds						Qvalue
75) Fluoranthene	12.539	202	73044	2.685	ng	98
78) Pyrene	12.762	202	64159	2.188	ng	96
88) Benzo(b)fluoranthene	15.009	252	46039m	2.659	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
Data File : BF134723.D
Acq On : 11 Aug 2023 09:52
Operator : CG\JU
Sample : 03954-16
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-20

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/12/2023
Supervised By :mohammad ahmed 08/12/2023

Quant Time: Aug 11 22:19:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

