

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
 Data File : BF134746.D
 Acq On : 11 Aug 2023 22:28
 Operator : CG\JU
 Sample : 03971-05
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-PP041A

Quant Time: Aug 11 23:59:00 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	210984	20.000	ng	0.00
21) Naphthalene-d8	8.080	136	794902	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	340397	20.000	ng	0.00
64) Phenanthrene-d10	11.321	188	504127	20.000	ng	# 0.00
76) Chrysene-d12	13.974	240	368314	20.000	ng	0.00
86) Perylene-d12	15.445	264	271460	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	627210	45.189	ng	0.00
7) Phenol-d6	6.434	99	1076222	60.658	ng	0.00
23) Nitrobenzene-d5	7.363	82	712009	55.976	ng	0.00
42) 2,4,6-Tribromophenol	10.621	330	97725	28.163	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1253225	61.215	ng	0.00
79) Terphenyl-d14	12.915	244	1134116	54.282	ng	0.00
Target Compounds						
31) Naphthalene	8.098	128	251657	6.299	ng	98
37) 2-Methylnaphthalene	8.792	142	61605	2.326	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
Data File : BF134746.D
Acq On : 11 Aug 2023 22:28
Operator : CG\JU
Sample : 03971-05
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-PP041A

Quant Time: Aug 11 23:59:00 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

