

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/11/2023	BNA_F	<u>BF134719.D</u>
Compound Name	Response	Retention Time
DDT	773889	13.492
DDD	28715	13.192
DDE	1452	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
30167	804056	3.75