

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/11/2023	BNA_F	<u>BF134729.D</u>
Compound Name	Response	Retention Time
DDT	680703	13.492
DDD	22862	13.192
DDE	996	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
23858	704561	3.39