



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
Data File : BF134719.D
Acq On : 11 Aug 2023 07:36
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Aug 04 09:29:15 2023






AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.7	106693	PASS
68	69	0.00	2	2.0	1931	PASS
69	198	0.00	100	32.6	97325	PASS
70	69	0.00	2	0.5	532	PASS
127	198	10	80	43.3	129597	PASS
197	198	0.00	2	0.3	840	PASS
198	198	100	100	100.0	298965	PASS
199	198	5	9	6.7	20126	PASS
275	198	10	60	29.6	88507	PASS
365	198	1	100	4.0	11997	PASS
441	198	0.01	100	18.8	56059	PASS
442	442	50	100	100.0	361429	PASS
443	442	15	24	19.3	69685	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
 Data File : BF134719.D
 Acq On : 11 Aug 2023 07:36
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Quant Time: Aug 12 00:46:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration



TIC: BF134719.D\data.ms

(70) Pentachlorophenol (C)

11.127min (+ 0.000) 528589.61 ng




response 363197

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.10	63.80
264.00	59.70	62.02
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081123\
Data File : BF134719.D
Acq On : 11 Aug 2023 07:36
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 12 00:46:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration



TIC: BF134719.D\data.ms

(77) Benzidine

12.674min (+ 0.012) 194871.19 ng

response 1810543

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.09
183.00	10.80	12.12
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/11/2023	BNA_F	<u>BF134719.D</u>
Compound Name	Response	Retention Time
DDT	773889	13.492
DDD	28715	13.192
DDE	1452	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
30167	804056	3.75

Instrument :
BNA_F
ClientSampleId :
DFTPP