

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF081319\
 Data File : BF116220.D
 Acq On : 13 Aug 2019 9:56
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 13 23:15:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 11:13:18 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.604	0.609	-0.8	114	-0.04
3	Pyridine	1.686	1.657	1.7	110	-0.03
4	n-Nitrosodimethylamine	0.665	0.659	0.9	111	-0.04
5 S	2-Fluorophenol	1.195	1.294	-8.3	118	-0.01
6	Aniline	2.159	2.136	1.1	108	-0.01
7 S	Phenol-d6	1.507	1.575	-4.5	115	0.00
8	2-Chlorophenol	1.321	1.428	-8.1	118	-0.01
9	Benzaldehyde	1.040	1.008	3.1	106	0.00
10 C	Phenol	1.775	1.806	-1.7	112	0.00
11	bis(2-Chloroethyl)ether	1.305	1.414	-8.4	119	-0.01
12	1,3-Dichlorobenzene	1.527	1.582	-3.6	114	0.00
13 C	1,4-Dichlorobenzene	1.529	1.593	-4.2	114	0.00
14	1,2-Dichlorobenzene	1.408	1.452	-3.1	111	-0.01
15	Benzyl Alcohol	1.144	1.170	-2.3	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.780	1.856	-4.3	113	-0.01
17	2-Methylphenol	1.119	1.184	-5.8	118	0.00
18	Hexachloroethane	0.563	0.598	-6.2	116	-0.01
19 P	n-Nitroso-di-n-propylamine	0.934	0.982	-5.1	117	-0.01
20	3+4-Methylphenols	1.324	1.395	-5.4	113	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	115	-0.01
22	Acetophenone	0.439	0.443	-0.9	115	-0.01
23 S	Nitrobenzene-d5	0.346	0.350	-1.2	115	0.00
24	Nitrobenzene	0.374	0.386	-3.2	118	0.00
25	Isophorone	0.663	0.703	-6.0	122	-0.01
26 C	2-Nitrophenol	0.176	0.194	-10.2	124	0.00
27	2,4-Dimethylphenol	0.265	0.268	-1.1	115	-0.01
28	bis(2-Chloroethoxy)methane	0.411	0.440	-7.1	123	-0.01
29 C	2,4-Dichlorophenol	0.280	0.300	-7.1	120	0.00
30	1,2,4-Trichlorobenzene	0.319	0.329	-3.1	117	0.00
31	Naphthalene	0.941	0.970	-3.1	115	-0.01
32	Benzoic acid	0.187	0.178	4.8	119	0.00
33	4-Chloroaniline	0.421	0.443	-5.2	119	0.00
34 C	Hexachlorobutadiene	0.184	0.188	-2.2	116	-0.01
35	Caprolactam	0.091	0.095	-4.4	119	0.00
36 C	4-Chloro-3-methylphenol	0.291	0.315	-8.2	123	0.00
37	2-Methylnaphthalene	0.620	0.640	-3.2	116	0.00
38	1-Methylnaphthalene	0.586	0.604	-3.1	116	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	114	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.535	0.556	-3.9	117	0.00
41 P	Hexachlorocyclopentadiene	0.209	0.180	13.9	95	0.00
42 S	2,4,6-Tribromophenol	0.194	0.195	-0.5	113	0.00
43 C	2,4,6-Trichlorophenol	0.368	0.364	1.1	112	0.00
44	2,4,5-Trichlorophenol	0.380	0.379	0.3	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF081319\
 Data File : BF116220.D
 Acq On : 13 Aug 2019 9:56
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 13 23:15:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 07 11:13:18 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.068	1.041	2.5	109	0.00
46	1,1'-Biphenyl	1.412	1.431	-1.3	114	0.00
47	2-Chloronaphthalene	1.175	1.199	-2.0	115	0.00
48	2-Nitroaniline	0.388	0.400	-3.1	117	0.00
49	Acenaphthylene	1.715	1.726	-0.6	113	0.00
50	Dimethylphthalate	1.362	1.420	-4.3	118	0.00
51	2,6-Dinitrotoluene	0.296	0.312	-5.4	119	0.00
52 C	Acenaphthene	1.052	1.069	-1.6	114	-0.01
53	3-Nitroaniline	0.366	0.371	-1.4	115	0.00
54 P	2,4-Dinitrophenol	0.126	0.123	2.4	108	0.00
55	Dibenzofuran	1.530	1.469	4.0	107	0.00
56 P	4-Nitrophenol	0.234	0.216	7.7	104	0.00
57	2,4-Dinitrotoluene	0.394	0.372	5.6	105	0.00
58	Fluorene	1.177	1.227	-4.2	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.320	0.308	3.8	109	0.00
60	Diethylphthalate	1.271	1.322	-4.0	116	0.00
61	4-Chlorophenyl-phenylether	0.596	0.621	-4.2	115	0.00
62	4-Nitroaniline	0.378	0.382	-1.1	114	0.00
63	Azobenzene	1.308	1.368	-4.6	118	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	0.106	0.116	-9.4	117	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.643	-6.8	117	-0.01
67	4-Bromophenyl-phenylether	0.214	0.232	-8.4	117	0.00
68	Hexachlorobenzene	0.248	0.259	-4.4	113	0.00
69	Atrazine	0.198	0.179	9.6	99	0.00
70 C	Pentachlorophenol	0.120	0.106	11.7	93	0.00
71	Phenanthrene	1.022	1.046	-2.3	111	0.00
72	Anthracene	1.035	1.077	-4.1	113	0.00
73	Carbazole	0.939	1.000	-6.5	115	0.00
74	Di-n-butylphthalate	1.095	1.181	-7.9	114	0.00
75 C	Fluoranthene	1.070	1.103	-3.1	112	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzidine	0.676	0.755	-11.7	107	0.00
78	Pyrene	1.508	1.594	-5.7	110	0.00
79 S	Terphenyl-d14	0.949	0.985	-3.8	108	0.00
80	Butylbenzylphthalate	0.638	0.727	-13.9	115	-0.01
81	Benzo(a)anthracene	1.278	1.372	-7.4	110	0.00
82	3,3'-Dichlorobenzidine	0.444	0.477	-7.4	107	0.00
83	Chrysene	1.241	1.298	-4.6	107	0.00
84	Bis(2-ethylhexyl)phthalate	0.829	0.975	-17.6	118	0.00
85 c	Di-n-octyl phthalate	1.316	1.529	-16.2	115	0.00
86	Indeno(1,2,3-cd)pyrene	1.042	1.085	-4.1	111	0.01
87 I	Perylene-d12	1.000	1.000	0.0	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF081319\
Data File : BF116220.D
Acq On : 13 Aug 2019 9:56
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 13 23:15:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 07 11:13:18 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.156	1.216	-5.2	104	0.00
89	Benzo(k)fluoranthene	1.126	1.161	-3.1	111	0.00
90 C	Benzo(a)pyrene	1.034	1.087	-5.1	109	0.00
91	Dibenzo(a,h)anthracene	0.895	0.927	-3.6	111	0.00
92	Benzo(q,h,i)perylene	0.879	0.926	-5.3	116	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0