

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081418\
 Data File : BF108138.D
 Acq On : 14 Aug 2018 12:19
 Operator : JU/SJ
 Sample : J4409-13RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-12C

Quant Time: Aug 14 13:00:09 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 08 12:24:24 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.02	152	189183	20.00	ng	0.00
21) Naphthalene-d8	8.30	136	669557	20.00	ng	0.00
38) Acenaphthene-d10	10.07	164	296937	20.00	ng	0.00
63) Phenanthrene-d10	11.57	188	518929	20.00	ng	0.00
75) Chrysene-d12	14.22	240	447560	20.00	ng	0.00
86) Perylene-d12	15.79	264	342609	20.00	ng	0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.65	112	1009191	96.58	ng	0.02
7) Phenol-d6	6.64	99	1340122	96.51	ng	0.01
23) Nitrobenzene-d5	7.58	82	862456	71.88	ng	0.00
41) 2,4,6-Tribromophenol	10.86	330	289024	97.58	ng	0.00
44) 2-Fluorobiphenyl	9.38	172	1330200	71.92	ng	0.00
78) Terphenyl-d14	13.15	244	1244846	70.72	ng	0.00


Target Compounds				Qvalue		
10) Phenol	6.65	94	66464	4.627	ng	77
19) n-Nitroso-di-n-propylamine	7.58	70	117821	12.547	ng	# 85
25) Isophorone	7.58	82	860144	37.609	ng	# 64
31) Naphthalene	8.32	128	148697	4.883	ng	98
32) Benzoic acid	8.07	122	392	6.033	ng	# 60
49) Dimethylphthalate	9.77	163	213078	10.302	ng	99
51) Acenaphthene	10.10	154	197879	11.810	ng	98
53) 2,4-Dinitrophenol	10.30	184	287	7.182	ng	# 1
57) Fluorene	10.61	166	203681	10.599	ng	98
64) 4,6-Dinitro-2-methylphenol	10.86	198	777	5.713	ng	# 1
66) 4-Bromophenyl-phenylether	10.86	248	19543	3.465	ng	# 1
70) Phenanthrene	11.59	178	935330	38.214	ng	99
71) Anthracene	11.64	178	545530	21.746	ng	99
74) Fluoranthene	12.81	202	1661339	62.193	ng	95
77) Pyrene	13.03	202	1444060	50.963	ng	98
80) Benzo(a)anthracene	14.21	228	717672	27.941	ng	97
82) Chrysene	14.25	228	507915	21.569	ng	97
85) Indeno(1,2,3-cd)pyrene	17.41	276	160339	7.858	ng	# 88
87) Benzo(b)fluoranthene	15.32	252	715216	34.936	ng	97
88) Benzo(k)fluoranthene	15.32	252	715216	38.005	ng	97
89) Benzo(a)pyrene	15.72	252	389486	21.246	ng	97
90) Dibenzo(a,h)anthracene	17.61	278	40343	2.660	ng	96
91) Benzo(g,h,i)perylene	17.90	276	128921	8.632	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF081418\
Data File : BF108138.D
Acq On : 14 Aug 2018 12:19
Operator : JU/SJ
Sample : J4409-13RE
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-12C

Quant Time: Aug 14 13:00:09 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 08 12:24:24 2018
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.02 min Scan# 796
Delta R.T. 0.01 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19

Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-12C




Tgt Ion:152 Resp: 189183
Ion Ratio Lower Upper
152 100
150 157.0 124.6 186.8
115 61.3 50.1 75.1



#5

2-Fluorophenol
Concen: 96.578 ng
RT: 5.65 min Scan# 563
Delta R.T. 0.02 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19

Tgt Ion:112 Resp: 1009191
Ion Ratio Lower Upper
112 100
64 71.5 47.9 71.9
63 37.7 23.7 35.5#





#7

Phenol-d6

Concen: 96.509 ng

RT: 6.64 min Scan# 732

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :


BNA_F

ClientSampleId :




2018-NYSEG-CLARK-AREA-12C

Tgt Ion: 99 Resp: 1340122

Ion	Ratio	Lower	Upper
99	100		
42	21.9	14.5	21.7#
71	35.7	25.4	38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#10

Phenol

Concen: 4.627 ng

RT: 6.65 min Scan# 733


Delta R.T. 0.01 min

Lab File: BF108138.D



Acq: 14 Aug 2018 12:19


Tgt Ion: 94 Resp: 66464

Ion	Ratio	Lower	Upper
94	100		
65	35.3	7.9	47.9
66	53.6	16.2	56.2



Abundance Ion 94.00 (93.70 to 94.70): BF1
Ion 65.00 (64.70 to 65.70): BF1
Ion 66.00 (65.70 to 66.70): BF1









#19
n-Nitroso-di-n-propylamine
Concen: 12.547 ng
RT: 7.58 min Scan# 891
Delta R.T. 0.15 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19




Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-12C


Tgt Ion: 70 Resp: 117821
Ion Ratio Lower Upper
70 100
42 56.0 47.5 71.3
101 0.0 7.4 11.2#
130 1.9 16.6 25.0#



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.30 min Scan# 1014
Delta R.T. 0.00 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19

Tgt Ion: 136 Resp: 669557
Ion Ratio Lower Upper
136 100
137 11.3 8.9 13.3
54 10.5 6.6 9.8#
68 6.4 5.0 7.4









#23
 Nitrobenzene-d5
 Concen: 71.878 ng
 RT: 7.58 min Scan# 891
 Delta R.T. 0.00 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19




Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-12C


Tgt Ion	Ratio	Lower	Upper
82	100		
128	41.0	36.1	54.1
54	60.1	41.4	62.2



#25
 Isophorone
 Concen: 37.609 ng
 RT: 7.58 min Scan# 891
 Delta R.T. -0.26 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.1	5.2	7.8#
138	0.0	14.9	22.3#





#31

Naphthalene

Concen: 4.883 ng

RT: 8.32 min Scan# 1018

Delta R.T. 0.00 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C





Tgt Ion:128 Resp: 148697

Ion Ratio Lower Upper

128 100

129 10.7 8.8 13.2

127 12.3 10.9 16.3



#32

Benzoic acid

Concen: 6.033 ng

RT: 8.07 min Scan# 975

Delta R.T. 0.03 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19




Tgt Ion:122 Resp: 392


Ion Ratio Lower Upper

122 100

105 73.8 101.1 141.1#

77 125.6 70.7 110.7#









#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 10.07 min Scan# 1314
 Delta R.T. 0.00 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19

Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-12C

Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.0	86.1	129.1
160	47.3	38.3	57.5




Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E






#41
 2,4,6-Tribromophenol
 Concen: 97.582 ng
 RT: 10.86 min Scan# 1449
 Delta R.T. 0.00 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.0	115.6
141	33.8	29.9	44.9



Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E





#44

2-Fluorobiphenyl

Concen: 71.921 ng

RT: 9.38 min Scan# 1197

Delta R.T. 0.00 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion:172 Resp: 1330200

Ion Ratio Lower Upper

172 100

171 36.4 29.7 44.5




170 24.5 19.6 29.4



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#49

Dimethylphthalate

Concen: 10.302 ng

RT: 9.77 min Scan# 1263

Delta R.T. -0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19


Tgt Ion:163 Resp: 213078

Ion Ratio Lower Upper

163 100

194 3.9 2.7 4.1



164 9.8 8.2 12.2




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E





#51

Acenaphthene

Concen: 11.810 ng

RT: 10.10 min Scan# 1320

Delta R.T. 0.00 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion:154 Resp: 197879

Ion Ratio Lower Upper

154 100

153 115.9 91.2 136.8




152 57.2 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#53

2,4-Dinitrophenol

Concen: 7.182 ng

RT: 10.30 min Scan# 1353

Delta R.T. 0.18 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19


Tgt Ion:184 Resp: 287

Ion Ratio Lower Upper

184 100

63 234.6 54.2 81.2#



154 383.8 184.0 276.0#




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 154.00 (153.70 to 154.70): E





#57

Fluorene

Concen: 10.599 ng

RT: 10.61 min Scan# 1407

Delta R.T. -0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion:166 Resp: 203681

Ion Ratio Lower Upper

166 100

165 101.2 79.8 119.8




167 14.7 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.57 min Scan# 1569

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19


Tgt Ion:188 Resp: 518929

Ion Ratio Lower Upper

188 100

94 8.6 8.5 12.7



80 9.2 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#64
 4,6-Dinitro-2-methylphenol
 Concen: 5.713 ng
 RT: 10.86 min Scan# 1449
 Delta R.T. 0.21 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19




Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-12C


Tgt Ion:198 Resp: 777
 Ion Ratio Lower Upper
 198 100
 51 559.1 37.7 77.7#
 105 230.1 36.3 76.3#



#66
 4-Bromophenyl-phenylether
 Concen: 3.465 ng
 RT: 10.86 min Scan# 1449
 Delta R.T. -0.24 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19

Tgt Ion:248 Resp: 19543
 Ion Ratio Lower Upper
 248 100
 250 197.7 76.9 115.3#
 141 484.9 74.4 111.6#









#70
Phenanthrene
Concen: 38.214 ng
RT: 11.59 min Scan# 1573
Delta R.T. 0.00 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19




Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-12C


Tgt Ion:178 Resp: 935330
Ion Ratio Lower Upper
178 100
176 20.3 15.8 23.8
179 16.3 13.0 19.6



#71
Anthracene
Concen: 21.746 ng
RT: 11.64 min Scan# 1582
Delta R.T. 0.01 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19

Tgt Ion:178 Resp: 545530
Ion Ratio Lower Upper
178 100
176 18.9 15.4 23.2
179 15.9 12.6 19.0





#74

Fluoranthene

Concen: 62.193 ng

RT: 12.81 min Scan# 1780

Delta R.T. 0.02 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion:202 Resp: 1661339

Ion Ratio Lower Upper

202 100

101 10.6 0.0 34.1




203 18.9 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.22 min Scan# 2021

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19


Tgt Ion:240 Resp: 447560

Ion Ratio Lower Upper

240 100

120 9.4 9.5 14.3#



236 26.9 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Pyrene

Concen: 50.963 ng

RT: 13.03 min Scan# 1818

Delta R.T. 0.02 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion:202 Resp: 1444060

Ion Ratio Lower Upper

202 100

200 22.4 17.4 26.2




203 19.0 14.3 21.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 70.719 ng

RT: 13.15 min Scan# 1839

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19


Tgt Ion:244 Resp: 1244846

Ion Ratio Lower Upper

244 100

212 7.4 5.4 8.0



122 10.0 11.0 16.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#80

Benzo(a)anthracene

Concen: 27.941 ng

RT: 14.21 min Scan# 2019

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion:228 Resp: 717672

Ion Ratio Lower Upper




228 100

226 30.1 22.6 33.8

229 21.0 15.8 23.6



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E



#82

Chrysene

Concen: 21.569 ng

RT: 14.25 min Scan# 2025

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19


Tgt Ion:228 Resp: 507915

Ion Ratio Lower Upper



228 100


226 31.5 25.0 37.6

229 23.5 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 7.858 ng
 RT: 17.41 min Scan# 2562
 Delta R.T. 0.01 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19




Instrument :
 BNA_F
 ClientSampleId :
 2018-NYSEG-CLARK-AREA-12C


Tgt Ion	Ratio	Lower	Upper
276	100		
138	19.9	25.0	37.6#
277	24.9	20.1	30.1



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.79 min Scan# 2287
 Delta R.T. 0.01 min
 Lab File: BF108138.D
 Acq: 14 Aug 2018 12:19

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.2	28.8
265	23.0	17.5	26.3





#87

Benzo(b)fluoranthene

Concen: 34.936 ng

RT: 15.32 min Scan# 2207

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C





Tgt Ion:252 Resp: 715216

Ion Ratio Lower Upper

252 100

253 22.9 17.0 25.6

125 10.3 9.0 13.6



#88

Benzo(k)fluoranthene

Concen: 38.005 ng

RT: 15.32 min Scan# 2207

Delta R.T. -0.02 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19




Tgt Ion:252 Resp: 715216


Ion Ratio Lower Upper

252 100

253 22.9 17.9 26.9

125 10.3 10.2 15.2









#89
Benzo(a)pyrene
Concen: 21.246 ng
RT: 15.72 min Scan# 2275
Delta R.T. 0.01 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19




Instrument :
BNA_F
ClientSampleId :
2018-NYSEG-CLARK-AREA-12C


Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.0	17.1	25.7
125	11.7	9.8	14.6



#90
Dibenzo(a,h)anthracene
Concen: 2.660 ng
RT: 17.61 min Scan# 2596
Delta R.T. 0.19 min
Lab File: BF108138.D
Acq: 14 Aug 2018 12:19

Tgt Ion	Ratio	Lower	Upper
278	100		
139	21.0	15.9	23.9
279	26.5	19.0	28.4





#91

Benzo(g,h,i)perylene

Concen: 8.632 ng

RT: 17.90 min Scan# 2646

Delta R.T. 0.01 min

Lab File: BF108138.D

Acq: 14 Aug 2018 12:19

Instrument :

BNA_F

ClientSampleId :

2018-NYSEG-CLARK-AREA-12C


Tgt Ion: 276 Resp: 128921

Ion Ratio Lower Upper

276 100

277 23.8 17.9 26.9

138 20.7 21.8 32.8#



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

