


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081418\
 Data File : BF108134.D
 Acq On : 14 Aug 2018 9:48
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF080818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Aug 08 12:24:24 2018



AutoFind: Scans 1607, 1608, 1609; Background Corrected with Scan 1602

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.3	63781	PASS
68	69	0.00	2	1.6	865	PASS
69	198	0.00	100	32.3	55274	PASS
70	69	0.00	2	0.7	372	PASS
127	198	10	80	42.5	72624	PASS
197	198	0.00	2	0.3	501	PASS
198	198	100	100	100.0	170890	PASS
199	198	5	9	6.4	10969	PASS
275	198	10	60	26.6	45461	PASS
365	198	1	100	3.5	5920	PASS
441	198	0.01	100	13.4	22976	PASS
442	198	50	100	83.3	142381	PASS
443	442	15	24	18.6	26501	PASS