

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081423\
 Data File : BF134764.D
 Acq On : 14 Aug 2023 17:41
 Operator : CG\JU
 Sample : 03989-02
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-PP24A

Quant Time: Aug 15 00:32:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	156441	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	595919	20.000	ng	# 0.00
39) Acenaphthene-d10	9.834	164	271663	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	477666	20.000	ng	# 0.00
76) Chrysene-d12	13.969	240	313475	20.000	ng	# 0.00
86) Perylene-d12	15.439	264	304702	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	670481	65.149	ng	0.00
7) Phenol-d6	6.434	99	806432	61.299	ng	0.00
23) Nitrobenzene-d5	7.357	82	469184	49.202	ng	-0.01
42) 2,4,6-Tribromophenol	10.622	330	193755	69.964	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	856959	52.450	ng	-0.01
79) Terphenyl-d14	12.916	244	873995	49.150	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081423\
Data File : BF134764.D
Acq On : 14 Aug 2023 17:41
Operator : CG\JU
Sample : 03989-02
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-PP24A

Quant Time: Aug 15 00:32:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

