

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081523\
 Data File : BF134787.D
 Acq On : 15 Aug 2023 16:38
 Operator : CG\JU
 Sample : 04014-04 10X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB06-Z84-85

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/16/2023
 Supervised By :Sohil Jodhani 08/16/2023

Quant Time: Aug 16 03:11:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	182837	20.000	ng	0.00
21) Naphthalene-d8	8.087	136	514052	20.000	ng	# 0.00
39) Acenaphthene-d10	9.839	164	333661	20.000	ng	0.00
64) Phenanthrene-d10	11.328	188	493045	20.000	ng	# 0.00
76) Chrysene-d12	13.980	240	341568	20.000	ng	# 0.00
86) Perylene-d12	15.451	264	336765	20.000	ng	# 0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	66560	5.534	ng	-0.01
7) Phenol-d6	6.422	99	85251	5.545	ng	-0.02
23) Nitrobenzene-d5	7.357	82	47843	5.816	ng	-0.01
42) 2,4,6-Tribromophenol	10.628	330	15541	4.569	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	108520	5.408	ng	-0.01
79) Terphenyl-d14	12.916	244	89069	4.597	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.140	128	10576332	409.375	ng	95
37) 2-Methylnaphthalene	8.816	142	3682898	215.065	ng	100
38) 1-Methylnaphthalene	8.916	142	3664725	230.138	ng	98
46) 1,1'-Biphenyl	9.263	154	1357975	52.338	ng	98
49) Acenaphthylene	9.698	152	686647	22.469	ng	# 85
52) Acenaphthene	9.892	154	2769467	139.840	ng	99
55) Dibenzofuran	10.045	168	408632	14.494	ng	# 89
58) Fluorene	10.398	166	2139295	104.177	ng	92
71) Phenanthrene	11.380	178	5777793	220.623	ng	98
72) Anthracene	11.422	178	2478931	92.669	ng	97
73) Carbazole	11.557	167	105542	4.419	ng	# 88
75) Fluoranthene	12.557	202	2529729	91.447	ng	96
78) Pyrene	12.792	202	3466697	117.291	ng	94
81) Benzo(a)anthracene	13.974	228	1718488m	70.361	ng	
83) Chrysene	14.010	228	1563727	65.716	ng	96
87) Indeno(1,2,3-cd)pyrene	16.933	276	274648	13.867	ng	# 95
88) Benzo(b)fluoranthene	15.021	252	985728m	47.993	ng	
89) Benzo(k)fluoranthene	15.045	252	264882m	12.946	ng	
90) Benzo(a)pyrene	15.392	252	1016094	53.059	ng	# 97
91) Dibenzo(a,h)anthracene	16.945	278	87807	5.486	ng	# 92
92) Benzo(g,h,i)perylene	17.386	276	291825	17.873	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081523\
Data File : BF134787.D
Acq On : 15 Aug 2023 16:38
Operator : CG\JU
Sample : 04014-04 10X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85

Quant Time: Aug 16 03:11:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.798 min Scan# 800
Delta R.T. 0.000 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85



Tgt Ion:152 Resp: 18283
Ion Ratio Lower Upper
152 100
150 156.9 128.3 192.5
115 62.3 52.4 78.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#5
2-Fluorophenol
Concen: 5.534 ng
RT: 5.416 min Scan# 566
Delta R.T. -0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38


Tgt Ion:112 Resp: 66560
Ion Ratio Lower Upper
112 100
64 64.9 44.9 67.3
63 42.4 25.4 38.0#





#7
Phenol-d6
Concen: 5.545 ng
RT: 6.422 min Scan# 71
Delta R.T. -0.017 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38



Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85




Tgt Ion: 99 Resp: 8525
Ion Ratio Lower Upper
99 100
42 19.5 18.7 28.1
71 32.4 36.4 54.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.087 min Scan# 1020
Delta R.T. 0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Tgt Ion:136 Resp: 514052
Ion Ratio Lower Upper
136 100
137 11.3 8.4 12.6
54 8.7 5.2 7.8#
68 6.6 2.9 4.3#





#23
Nitrobenzene-d5
Concen: 5.816 ng
RT: 7.357 min Scan# 896
Delta R.T. -0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85



Tgt Ion: 82 Resp: 4784
Ion Ratio Lower Upper
82 100
128 54.9 34.3 51.5
54 52.1 43.8 65.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#31
Naphthalene
Concen: 409.375 ng
RT: 8.140 min Scan# 1029
Delta R.T. 0.035 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38

Tgt Ion:128 Resp:10576332
Ion Ratio Lower Upper
128 100
129 12.4 8.4 12.6
127 15.1 10.7 16.1





#37

2-Methylnaphthalene

Concen: 215.065 ng

RT: 8.816 min Scan# 1144

Delta R.T. 0.024 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38

Instrument :

BNA_F

ClientSampleId :

NEVINS-SB06-Z84-85

Tgt Ion:142 Resp: 368289

Ion Ratio Lower Upper

142 100





141 88.3 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/16/2023

Supervised By :Sohil Jodhani 08/16/2023



#38

1-Methylnaphthalene

Concen: 230.138 ng

RT: 8.916 min Scan# 1161

Delta R.T. 0.024 min

Lab File: BF134787.D




Acq: 15 Aug 2023 16:38


Tgt Ion:142 Resp: 3664725

Ion Ratio Lower Upper

142 100


141 91.2 71.4 107.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.839 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85



Tgt Ion:164 Resp: 33366
Ion Ratio Lower Upper
164 100
162 100.6 82.6 123.8
160 46.3 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#42
2,4,6-Tribromophenol
Concen: 4.569 ng
RT: 10.628 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38

Tgt Ion:330 Resp: 15541
Ion Ratio Lower Upper
330 100
332 94.1 78.8 118.2
141 79.5 25.4 38.2#





#45

2-Fluorobiphenyl

Concen: 5.408 ng

RT: 9.151 min Scan# 1201

Delta R.T. -0.012 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38

Instrument :

BNA_F

ClientSampleId :





NEVINS-SB06-Z84-85

Tgt Ion:	172	Resp:	108520
Ion Ratio	Lower	Upper	
172	100		
171	34.2	28.8	43.2
170	24.1	18.7	28.1

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#46

1,1'-Biphenyl

Concen: 52.338 ng




RT: 9.263 min Scan# 1220


Delta R.T. 0.000 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38

Tgt Ion:	154	Resp:	1357975
Ion Ratio	Lower	Upper	
154	100		
153	41.9	20.8	60.8
76	11.4	0.0	30.2





#49

Acenaphthylene

Concen: 22.469 ng

RT: 9.698 min Scan# 1293

Delta R.T. 0.000 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38

Instrument :

BNA_F

ClientSampleId :





NEVINS-SB06-Z84-85

Tgt Ion:152 Resp: 68664
 Ion Ratio Lower Upper
 152 100
 151 21.7 16.4 24.6
 153 27.6 10.9 16.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/16/2023
 Supervised By :Sohil Jodhani 08/16/2023



#52

Acenaphthene

Concen: 139.840 ng




RT: 9.892 min Scan# 1327


Delta R.T. 0.018 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38

Tgt Ion:154 Resp: 2769467
 Ion Ratio Lower Upper
 154 100
 153 110.9 90.0 135.0
 152 52.2 42.4 63.6









#55
Dibenzenofuran
Concen: 14.494 ng
RT: 10.045 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38

Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85

Tgt Ion:168 Resp: 40863
Ion Ratio Lower Upper
168 100
139 42.0 33.4 50.2
169 31.9 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#58
Fluorene
Concen: 104.177 ng
RT: 10.398 min Scan# 1413
Delta R.T. 0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38

Tgt Ion:166 Resp: 2139295
Ion Ratio Lower Upper
166 100
165 100.6 74.5 111.7
167 15.8 10.6 16.0









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.328 min Scan# 1571
Delta R.T. 0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38

Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85

Tgt Ion:188 Resp: 493045
Ion Ratio Lower Upper
188 100
94 10.5 5.1 7.7
80 11.4 6.6 9.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#71
Phenanthrene
Concen: 220.623 ng
RT: 11.380 min Scan# 1580
Delta R.T. 0.030 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38


Tgt Ion:178 Resp: 5777793
Ion Ratio Lower Upper
178 100
176 20.6 15.8 23.6
179 16.7 12.7 19.1





#72
 Anthracene
 Concen: 92.669 ng
 RT: 11.422 min Scan# 11
 Delta R.T. 0.018 min
 Lab File: BF134787.D
 Acq: 15 Aug 2023 16:38




Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB06-Z84-85




Tgt Ion:178 Resp: 247893
 Ion Ratio Lower Upper
 178 100
 176 19.9 14.6 21.8
 179 16.6 12.3 18.5

Manual Integrations
 APPROVED



Reviewed By :Yogesh Patel 08/16/2023
 Supervised By :Sohil Jodhani 08/16/2023




#73
 Carbazole
 Concen: 4.419 ng
 RT: 11.557 min Scan# 1610
 Delta R.T. 0.000 min
 Lab File: BF134787.D
 Acq: 15 Aug 2023 16:38




Tgt Ion:167 Resp: 105542
 Ion Ratio Lower Upper
 167 100
 166 28.8 17.5 26.3#
 139 17.2 11.0 16.4#





#75
Fluoranthene
Concen: 91.447 ng
RT: 12.557 min Scan# 11
Delta R.T. 0.018 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85




Tgt Ion:202 Resp: 2529729
Ion Ratio Lower Upper
202 100
101 10.0 0.0 28.5
203 19.1 0.0 37.5

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023




#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.980 min Scan# 2022
Delta R.T. 0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Tgt Ion:240 Resp: 341568
Ion Ratio Lower Upper
240 100
120 13.1 6.6 10.0#
236 26.5 20.2 30.2





#78
Pyrene
Concen: 117.291 ng
RT: 12.792 min Scan# 1820
Delta R.T. 0.024 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85




Tgt Ion:202 Resp: 346669
Ion Ratio Lower Upper
202 100
200 22.4 15.8 23.8
203 19.8 13.5 20.3

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023




#79
Terphenyl-d14
Concen: 4.597 ng
RT: 12.916 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Tgt Ion:244 Resp: 89069
Ion Ratio Lower Upper
244 100
212 7.6 5.6 8.4
122 13.1 8.0 12.0#





#81
Benzo(a)anthracene
Concen: 70.361 ng m
RT: 13.974 min Scan# 2017
Delta R.T. 0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85



Tgt Ion:228 Resp: 171848
Ion Ratio Lower Upper
228 100
226 41.5 22.3 33.5
229 21.7 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#83
Chrysene
Concen: 65.716 ng
RT: 14.010 min Scan# 2027
Delta R.T. 0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38

Tgt Ion:228 Resp: 1563727
Ion Ratio Lower Upper
228 100
226 32.2 24.5 36.7
229 22.4 15.8 23.8





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.451 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38

Instrument :

BNA_F

ClientSampleId :





NEVINS-SB06-Z84-85

Tgt Ion:264 Resp: 336768
 Ion Ratio Lower Upper
 264 100
 260 23.8 18.4 27.6
 265 26.3 15.8 23.8

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/16/2023
 Supervised By :Sohil Jodhani 08/16/2023



#87

Indeno(1,2,3-cd)pyrene

Concen: 13.867 ng




RT: 16.933 min Scan# 2524


Delta R.T. 0.006 min

Lab File: BF134787.D

Acq: 15 Aug 2023 16:38


Tgt Ion:276 Resp: 274648
 Ion Ratio Lower Upper
 276 100
 138 22.6 14.2 21.4
 277 24.6 19.9 29.9





#88
Benzo(b)fluoranthene
Concen: 47.993 ng m
RT: 15.021 min Scan# 2194
Delta R.T. 0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85



Tgt Ion:252 Resp: 985728
Ion Ratio Lower Upper
252 100
253 23.3 17.3 25.9
125 11.9 6.0 9.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023



#89
Benzo(k)fluoranthene
Concen: 12.946 ng m
RT: 15.045 min Scan# 2203
Delta R.T. 0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38


Tgt Ion:252 Resp: 264882
Ion Ratio Lower Upper
252 100
253 23.7 17.9 26.9
125 10.1 6.3 9.5





#90
Benzo(a)pyrene
Concen: 53.059 ng
RT: 15.392 min Scan# 21
Delta R.T. 0.012 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Instrument :
BNA_F
ClientSampleId :
NEVINS-SB06-Z84-85




Tgt Ion:252 Resp: 1016094
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 10.3 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/16/2023
Supervised By :Sohil Jodhani 08/16/2023




#91
Dibenzo(a,h)anthracene
Concen: 5.486 ng
RT: 16.945 min Scan# 2526
Delta R.T. -0.006 min
Lab File: BF134787.D
Acq: 15 Aug 2023 16:38




Tgt Ion:278 Resp: 87807
Ion Ratio Lower Upper
278 100
139 18.1 10.0 15.0#
279 27.3 19.7 29.5





#92
 Benzo(g,h,i)perylene
 Concen: 17.873 ng
 RT: 17.386 min Scan# 20
 Delta R.T. 0.012 min
 Lab File: BF134787.D
 Acq: 15 Aug 2023 16:38

Instrument :
 BNA_F
 ClientSampleId :
 NEVINS-SB06-Z84-85



Tgt Ion: 276 Resp: 291825

Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.0	28.6
138	22.4	12.2	18.2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/16/2023
 Supervised By :Sohil Jodhani 08/16/2023

