

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081623\
 Data File : BF134798.D
 Acq On : 16 Aug 2023 09:25
 Operator : CG\JU
 Sample : 04017-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-03-081423-B

Quant Time: Aug 17 01:29:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	188024	20.000	ng	0.00
21) Naphthalene-d8	8.075	136	714779	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	347174	20.000	ng	0.00
64) Phenanthrene-d10	11.322	188	532944	20.000	ng	# 0.00
76) Chrysene-d12	13.968	240	398892	20.000	ng	# 0.00
86) Perylene-d12	15.445	264	302428	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	530263	42.870	ng	0.00
7) Phenol-d6	6.428	99	659802	41.729	ng	-0.01
23) Nitrobenzene-d5	7.357	82	392133	34.284	ng	-0.01
42) 2,4,6-Tribromophenol	10.622	330	152787	43.171	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	740050	35.443	ng	-0.01
79) Terphenyl-d14	12.916	244	652640	28.843	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081623\
Data File : BF134798.D
Acq On : 16 Aug 2023 09:25
Operator : CG\JU
Sample : 04017-03
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-03-081423-B

Quant Time: Aug 17 01:29:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

