

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081623\
 Data File : BF134810.D
 Acq On : 16 Aug 2023 15:54
 Operator : CG\JU
 Sample : 04022-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-24

Quant Time: Aug 17 01:34:18 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.798	152	164666	20.000	ng	0.00
21) Naphthalene-d8	8.081	136	614263	20.000	ng	# 0.00
39) Acenaphthene-d10	9.833	164	261897	20.000	ng	0.00
64) Phenanthrene-d10	11.327	188	401063	20.000	ng	# 0.00
76) Chrysene-d12	13.974	240	289684	20.000	ng	# 0.00
86) Perylene-d12	15.451	264	213288	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	859281	79.324	ng	0.00
7) Phenol-d6	6.434	99	1051559	75.939	ng	0.00
23) Nitrobenzene-d5	7.363	82	631715	64.268	ng	0.00
42) 2,4,6-Tribromophenol	10.627	330	212397	79.556	ng	0.00
45) 2-Fluorobiphenyl	9.157	172	1129427	71.703	ng	0.00
79) Terphenyl-d14	12.921	244	1055663	64.242	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081623\
Data File : BF134810.D
Acq On : 16 Aug 2023 15:54
Operator : CG\JU
Sample : 04022-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-24

Quant Time: Aug 17 01:34:18 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

