




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081723\
Data File : BF134816.D
Acq On : 17 Aug 2023 09:25
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 09:50:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration



TIC: BF134816.D\data.ms

(70) Pentachlorophenol (C)




11.128min (+ 0.000) 695153.37 ng

response 293552

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.10	63.59
264.00	59.70	64.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081723\
Data File : BF134816.D
Acq On : 17 Aug 2023 09:25
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 09:50:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration



TIC: BF134816.D\data.ms

(77) Benzidine

12.674min (+ 0.012) 6793.80 ng m

response 1374061

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	13.98
183.00	10.80	11.70
0.00	0.00	0.00