

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081823\
 Data File : BF134850.D
 Acq On : 18 Aug 2023 20:21
 Operator : CG\JU
 Sample : 04074-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MAPLE-SHADE-COMP

Quant Time: Aug 19 01:46:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 04 09:29:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.792	152	138779	20.000	ng	0.00
21) Naphthalene-d8	8.069	136	534366	20.000	ng	#-0.01
39) Acenaphthene-d10	9.827	164	239559	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	339482	20.000	ng	# 0.00
76) Chrysene-d12	13.962	240	284637	20.000	ng	#-0.01
86) Perylene-d12	15.433	264	192544	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	592813	64.933	ng	0.00
7) Phenol-d6	6.422	99	462277	39.611	ng	-0.02
23) Nitrobenzene-d5	7.357	82	866496	101.334	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	290522	118.965	ng	-0.01
45) 2-Fluorobiphenyl	9.151	172	1436755	99.720	ng	-0.01
79) Terphenyl-d14	12.910	244	1210148	74.949	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF081823\
Data File : BF134850.D
Acq On : 18 Aug 2023 20:21
Operator : CG\JU
Sample : 04074-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MAPLE-SHADE-COMP

Quant Time: Aug 19 01:46:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration

