

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082019\
 Data File : BF116447.D
 Acq On : 20 Aug 2019 23:07
 Operator : HP/JU
 Sample : K4404-06MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Quant Time: Aug 21 11:07:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 06:03:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	85314	20.00	ng	-0.01
21) Naphthalene-d8	8.08	136	330566	20.00	ng	-0.02
39) Acenaphthene-d10	9.83	164	181910	20.00	ng	-0.02
64) Phenanthrene-d10	11.32	188	339701	20.00	ng	-0.02
76) Chrysene-d12	13.97	240	294616	20.00	ng	-0.01
87) Perylene-d12	15.42	264	304507	20.00	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.45	112	721313	141.54	ng	0.00
7) Phenol-d6	6.46	99	852725	132.62	ng	0.00
23) Nitrobenzene-d5	7.37	82	618670	108.03	ng	-0.01
42) 2,4,6-Tribromophenol	10.63	330	278861	158.11	ng	-0.01
45) 2-Fluorobiphenyl	9.16	172	1090232	112.26	ng	-0.01
79) Terphenyl-d14	12.90	244	1408756	100.76	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.62	88	101986	39.613	ng	99
3) Pyridine	3.40	79	179647	24.979	ng	97
4) n-Nitrosodimethylamine	3.29	42	115673	40.756	ng	99
6) Aniline	6.49	93	93620	10.167	ng	# 48
8) 2-Chlorophenol	6.60	128	284968	50.567	ng	95
9) Benzaldehyde	6.37	77	89755	20.231	ng	98
10) Phenol	6.47	94	307334	40.590	ng	81
11) bis(2-Chloroethyl)ether	6.54	93	280041	50.305	ng	98
12) 1,3-Dichlorobenzene	6.75	146	299024	45.913	ng	99
13) 1,4-Dichlorobenzene	6.82	146	314662	48.253	ng	96
14) 1,2-Dichlorobenzene	6.97	146	288474	48.043	ng	99
15) Benzyl Alcohol	6.96	79	220621	45.213	ng	97
16) 2,2'-oxybis(1-Chloropropan	7.08	45	360316	47.453	ng	# 51
17) 2-Methylphenol	7.07	107	234520	49.147	ng	# 88
18) Hexachloroethane	7.31	117	110449	46.003	ng	95
19) n-Nitroso-di-n-propylamine	7.22	70	210288	52.778	ng	98
20) 3+4-Methylphenols	7.24	107	316401	56.032	ng	# 61
22) Acetophenone	7.22	105	413155	56.989	ng	# 93
24) Nitrobenzene	7.39	77	326648	52.825	ng	99
25) Isophorone	7.63	82	583604	53.295	ng	99
26) 2-Nitrophenol	7.71	139	155308	53.282	ng	96
27) 2,4-Dimethylphenol	7.75	122	262613	59.885	ng	98
28) bis(2-Chloroethoxy)methane	7.83	93	354892	52.288	ng	99
29) 2,4-Dichlorophenol	7.96	162	252132	54.453	ng	99
30) 1,2,4-Trichlorobenzene	8.03	180	263722	49.965	ng	100
31) Naphthalene	8.10	128	848875	54.570	ng	99
32) Benzoic acid	7.92	122	125730	40.666	ng	97
33) 4-Chloroaniline	8.19	127	66244	9.531	ng	100
34) Hexachlorobutadiene	8.22	225	147659	48.570	ng	99
35) Caprolactam	8.54	113	61343m	40.962	ng	
36) 4-Chloro-3-methylphenol	8.66	107	267400	55.512	ng	99
37) 2-Methylnaphthalene	8.80	142	566893	55.335	ng	98
38) 1-Methylnaphthalene	8.89	142	529100	54.592	ng	98
40) 1,2,4,5-Tetrachlorobenzene	8.96	216	258935	53.244	ng	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082019\
 Data File : BF116447.D
 Acq On : 20 Aug 2019 23:07
 Operator : HP/JU
 Sample : K4404-06MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Quant Time: Aug 21 11:07:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 19 06:03:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	8.95	237	80486	40.059	ng	99
43) 2,4,6-Trichlorophenol	9.09	196	162589	48.572	ng	98
44) 2,4,5-Trichlorophenol	9.14	196	169245	48.912	ng	99
46) 1,1'-Biphenyl	9.26	154	700920	54.577	ng	98
47) 2-Chloronaphthalene	9.29	162	548309	51.297	ng	98
48) 2-Nitroaniline	9.39	65	173651	49.150	ng	98
49) Acenaphthylene	9.70	152	829783	53.211	ng	99
50) Dimethylphthalate	9.56	163	680271	54.922	ng	100
51) 2,6-Dinitrotoluene	9.63	165	145341	53.996	ng	# 88
52) Acenaphthene	9.87	154	509556	53.263	ng	100
53) 3-Nitroaniline	9.82	138	41573	12.500	ng	97
54) 2,4-Dinitrophenol	9.93	184	115709	84.077	ng	99
55) Dibenzofuran	10.05	168	731854	52.603	ng	100
56) 4-Nitrophenol	9.99	139	120911	56.750	ng	96
57) 2,4-Dinitrotoluene	10.05	165	187393	52.289	ng	95
58) Fluorene	10.39	166	610744	57.057	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.17	232	160011	54.965	ng	95
60) Diethylphthalate	10.26	149	645216	55.796	ng	99
61) 4-Chlorophenyl-phenylether	10.37	204	306057	56.457	ng	96
62) 4-Nitroaniline	10.42	138	129448	37.661	ng	98
63) Azobenzene	10.53	77	658562	55.368	ng	99
65) 4,6-Dinitro-2-methylphenol	10.46	198	95452	53.023	ng	99
66) n-Nitrosodiphenylamine	10.49	169	481608	47.103	ng	99
67) 4-Bromophenyl-phenylether	10.86	248	186047	51.161	ng	96
68) Hexachlorobenzene	10.94	284	206857	49.193	ng	98
69) Atrazine	11.02	200	116932	34.763	ng	99
70) Pentachlorophenol	11.15	266	161635	79.174	ng	98
71) Phenanthrene	11.35	178	920386	52.999	ng	100
72) Anthracene	11.40	178	964649	54.886	ng	99
73) Carbazole	11.56	167	812475	50.954	ng	99
74) Di-n-butylphthalate	11.87	149	1087265	58.452	ng	99
75) Fluoranthene	12.54	202	1042495	57.341	ng	98
78) Pyrene	12.77	202	1056830	47.587	ng	99
80) Butylbenzylphthalate	13.37	149	501873	53.423	ng	98
81) Benzo(a)anthracene	13.96	228	958628	50.907	ng	100
82) 3,3'-Dichlorobenzidine	13.92	252	32791	5.012	ng	100
83) Chrysene	13.99	228	964620	52.764	ng	99
84) Bis(2-ethylhexyl)phthalate	13.92	149	732364	59.991	ng	99
85) Di-n-octyl phthalate	14.53	149	1241216	64.012	ng	99
86) Indeno(1,2,3-cd)pyrene	16.89	276	947138	61.684	ng	100
88) Benzo(b)fluoranthene	15.00	252	921529	52.339	ng	100
89) Benzo(k)fluoranthene	15.03	252	901015m	52.539	ng	
90) Benzo(a)pyrene	15.36	252	872602	55.441	ng	99
91) Dibenzo(a,h)anthracene	16.89	278	811521	59.565	ng	98
92) Benzo(g,h,i)perylene	17.32	276	771419	57.654	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 21 11:07:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF073019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 19 06:03:31 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.80 min Scan# 802
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD

Tot Ion:152 Resp: 85314
Ion Ratio Lower Upper
152 100
150 158.9 126.2 189.2
115 62.2 49.9 74.9




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 39.613 ng
RT: 2.62 min Scan# 90
Delta R.T. 0.07 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion: 88 Resp: 101986
Ion Ratio Lower Upper
88 100
58 67.0 53.7 80.5
43 25.0 21.0 31.4



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 24.979 ng

RT: 3.40 min Scan# 224

Delta R.T. 0.11 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion: 79 Resp: 179647

Ion Ratio Lower Upper




79 100

52 59.4 49.8 74.8

51 30.9 25.2 37.8



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 40.756 ng

RT: 3.29 min Scan# 204

Delta R.T. 0.03 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion: 42 Resp: 115673

Ion Ratio Lower Upper



42 100


74 144.1 116.0 174.0

44 6.4 4.6 7.0



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 141.539 ng

RT: 5.45 min Scan# 571

Delta R.T. 0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:112 Resp: 721313

Ion Ratio Lower Upper

112 100

64 57.7 45.4 68.2




63 29.6 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): B

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 10.167 ng

RT: 6.49 min Scan# 748

Delta R.T. 0.00 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion: 93 Resp: 93620

Ion Ratio Lower Upper

93 100

66 71.5 32.2 48.4#



65 47.2 17.0 25.6#



Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 132.621 ng

RT: 6.46 min Scan# 743

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion: 99 Resp: 852725

Ion Ratio Lower Upper

99 100

42 16.5 13.8 20.8




71 32.7 27.3 40.9



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 50.567 ng

RT: 6.60 min Scan# 767

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion: 128 Resp: 284968

Ion Ratio Lower Upper

128 100

130 32.3 12.9 52.9



64 49.1 24.3 64.3



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 20.231 ng

RT: 6.37 min Scan# 729

Delta R.T. 0.00 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD





Tot Ion: 77 Resp: 89755

Ion Ratio Lower Upper

77 100

105 101.0 79.2 119.2

106 97.0 74.3 114.3



#10

Phenol

Concen: 40.590 ng

RT: 6.47 min Scan# 745

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07




Tgt Ion: 94 Resp: 307334

Ion Ratio Lower Upper

94 100

65 28.0 16.8 56.8

66 38.2 34.2 74.2









#11
bis(2-Chloroethyl)ether
Concen: 50.305 ng
RT: 6.54 min Scan# 757
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD


Tot Ion: 93 Resp: 280041
Ion Ratio Lower Upper
93 100
63 74.6 51.9 91.9
95 32.6 13.0 53.0



#12
1,3-Dichlorobenzene
Concen: 45.913 ng
RT: 6.75 min Scan# 792
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion: 146 Resp: 299024
Ion Ratio Lower Upper
146 100
148 64.1 51.3 76.9
75 33.1 27.4 41.2









#13
 1,4-Dichlorobenzene
 Concen: 48.253 ng
 RT: 6.82 min Scan# 805
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion:146 Resp: 314662
 Ion Ratio Lower Upper
 146 100
 148 61.9 51.3 76.9
 111 39.3 34.0 51.0



#14
 1,2-Dichlorobenzene
 Concen: 48.043 ng
 RT: 6.97 min Scan# 831
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:146 Resp: 288474
 Ion Ratio Lower Upper
 146 100
 148 64.4 51.2 76.8
 111 42.4 33.5 50.3





#15

Benzyl Alcohol

Concen: 45.213 ng

RT: 6.96 min Scan# 829

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion: 79 Resp: 220621

Ion Ratio Lower Upper

79 100

108 76.3 62.7 94.1




77 67.8 52.2 78.2



Abundance Ion 79.00 (78.70 to 79.70): BF1

Ion 108.00 (107.70 to 108.70): BF1

Ion 77.00 (76.70 to 77.70): BF1



#16

2,2'-oxybis(1-Chloropropane)

Concen: 47.453 ng

RT: 7.08 min Scan# 849

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion: 45 Resp: 360316

Ion Ratio Lower Upper

45 100

77 41.6 0.0 40.0#



79 37.9 0.0 35.9#




Abundance Ion 45.00 (44.70 to 45.70): BF1

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1





#17

2-Methylphenol

Concen: 49.147 ng

RT: 7.07 min Scan# 848

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD

Tot Ion:107 Resp: 234520


Ion Ratio Lower Upper

107 100

108 111.0 91.2 136.8

77 66.4 40.3 60.5#

79 60.6 37.6 56.4#






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1



#18

Hexachloroethane

Concen: 46.003 ng

RT: 7.31 min Scan# 888

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:117 Resp: 110449

Ion Ratio Lower Upper

117 100

119 97.1 76.1 114.1

201 104.7 77.0 115.4







Abundance

Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 52.778 ng
RT: 7.22 min Scan# 873
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07


Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD

Tot Ion:	70	Resp:	210288
Ion	Ratio	Lower	Upper
70	100		
42	52.1	40.9	61.3
101	10.1	7.9	11.9
130	22.2	16.9	25.3



#20
3+4-Methylphenols
Concen: 56.032 ng
RT: 7.24 min Scan# 876
Delta R.T. 0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:	107	Resp:	316401
Ion	Ratio	Lower	Upper
107	100		
108	89.8	67.6	107.6
77	41.5	104.4	144.4#
79	27.2	7.4	47.4









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.08 min Scan# 1019
Delta R.T. -0.02 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD

Tot Ion:136	Resp:	330566
Ion Ratio	Lower	Upper
136	100	
137	11.1	9.0 13.4
54	7.4	5.8 8.6
68	6.6	5.2 7.8



#22
Acetophenone
Concen: 56.989 ng
RT: 7.22 min Scan# 872
Delta R.T. -0.02 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:105	Resp:	413155
Ion Ratio	Lower	Upper
105	100	
71	8.0	5.9 8.9
51	24.6	25.3 37.9#
120	22.0	17.9 26.9









#23
 Nitrobenzene-d5
 Concen: 108.032 ng
 RT: 7.37 min Scan# 899
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Tot Ion: 82 Resp: 618670
 Ion Ratio Lower Upper
 82 100
 128 47.1 36.5 54.7
 54 44.1 36.0 54.0



#24
 Nitrobenzene
 Concen: 52.825 ng
 RT: 7.39 min Scan# 902
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion: 77 Resp: 326648
 Ion Ratio Lower Upper
 77 100
 123 46.5 36.7 55.1
 65 12.8 10.6 16.0





#25

Isophorone

Concen: 53.295 ng

RT: 7.63 min Scan# 942

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion: 82 Resp: 583604

Ion Ratio Lower Upper




82 100

95 6.7 5.4 8.0

138 17.3 14.5 21.7



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 138.00 (137.70 to 138.70): BF1



#26

2-Nitrophenol

Concen: 53.282 ng

RT: 7.71 min Scan# 956

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion: 139 Resp: 155308

Ion Ratio Lower Upper



139 100


109 28.5 23.4 35.2

65 47.9 41.1 61.7



Abundance Ion 139.00 (138.70 to 139.70): BF1
Ion 109.00 (108.70 to 109.70): BF1
Ion 65.00 (64.70 to 65.70): BF1









#27
2,4-Dimethylphenol
Concen: 59.885 ng
RT: 7.75 min Scan# 962
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD

Tot Ion:122 Resp: 262613
Ion Ratio Lower Upper
122 100
107 118.0 93.0 139.6
121 58.0 47.6 71.4




Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E





#28
bis(2-Chloroethoxy)methane
Concen: 52.288 ng
RT: 7.83 min Scan# 977
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion: 93 Resp: 354892
Ion Ratio Lower Upper
93 100
95 33.3 26.2 39.4
123 11.5 9.2 13.8



Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 123.00 (122.70 to 123.70): E








#29
 2,4-Dichlorophenol
 Concen: 54.453 ng
 RT: 7.96 min Scan# 999
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion	Ratio	Lower	Upper
162	100		
164	64.3	45.1	85.1
98	34.9	15.8	55.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF1






#30
 1,2,4-Trichlorobenzene
 Concen: 49.965 ng
 RT: 8.03 min Scan# 1010
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.8	75.7	113.5
145	29.9	24.3	36.5



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 54.570 ng

RT: 8.10 min Scan# 1023

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:128 Resp: 848875

Ion Ratio Lower Upper

128 100

129 11.6 9.3 13.9




127 13.3 11.4 17.0



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 40.666 ng

RT: 7.92 min Scan# 991

Delta R.T. 0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:122 Resp: 125730

Ion Ratio Lower Upper

122 100

105 121.9 106.1 146.1



77 94.4 76.7 116.7




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1









#33
4-Chloroaniline
Concen: 9.531 ng
RT: 8.19 min Scan# 1037
Delta R.T. 0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD


Tot Ion:127 Resp: 66244
Ion Ratio Lower Upper
127 100
129 32.9 26.1 39.1
65 31.0 24.9 37.3
92 19.3 15.7 23.5



#34
Hexachlorobutadiene
Concen: 48.570 ng
RT: 8.22 min Scan# 1042
Delta R.T. -0.02 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:225 Resp: 147659
Ion Ratio Lower Upper
225 100
223 64.1 50.0 75.0
227 63.8 51.0 76.4









#35
Caprolactam
Concen: 40.962 ng/ml
RT: 8.54 min Scan# 1097
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD


Tot Ion:113 Resp: 61343
Ion Ratio Lower Upper
113 100
55 155.1 144.3 184.3
56 114.3 97.1 137.1



#36
4-Chloro-3-methylphenol
Concen: 55.512 ng
RT: 8.66 min Scan# 1117
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:107 Resp: 267400
Ion Ratio Lower Upper
107 100
144 25.0 20.2 30.2
142 76.3 61.7 92.5









#37
2-Methylnaphthalene
Concen: 55.335 ng
RT: 8.80 min Scan# 1141
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD


Tot Ion	Ratio	Lower	Upper
142	100		
141	88.8	70.9	106.3
115	33.3	29.0	43.6



#38
1-Methylnaphthalene
Concen: 54.592 ng
RT: 8.89 min Scan# 1157
Delta R.T. -0.02 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.7	74.4	111.6
116	3.7	3.1	4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.83 min Scan# 1317

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:164 Resp: 181910

Ion Ratio Lower Upper

164 100

162 102.0 82.6 124.0




160 44.6 37.6 56.4



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 53.244 ng

RT: 8.96 min Scan# 1169

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Tgt Ion:216 Resp: 258935


Ion Ratio Lower Upper

216 100

214 79.9 63.6 95.4

179 21.7 17.8 26.8

108 20.2 16.5 24.7





Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 40.059 ng

RT: 8.95 min Scan# 1166

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD



Tot Ion:237 Resp: 80486

Ion Ratio Lower Upper



237 100

235 62.6 43.4 83.4

272 13.0 0.0 33.2



Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 158.114 ng

RT: 10.63 min Scan# 1453

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07



Tgt Ion:330 Resp: 278861

Ion Ratio Lower Upper


330 100

332 94.4 76.8 115.2

141 32.3 26.3 39.5



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 48.572 ng
 RT: 9.09 min Scan# 1190
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion:196	Resp:	162589
Ion Ratio	Lower	Upper
196	100	
198	101.0	79.2 118.8
200	32.0	25.2 37.8




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#44
 2,4,5-Trichlorophenol
 Concen: 48.912 ng
 RT: 9.14 min Scan# 1199
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:196	Resp:	169245
Ion Ratio	Lower	Upper
196	100	
198	100.1	78.8 118.2
97	52.1	42.2 63.4
132	30.7	24.0 36.0



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E






#45
 2-Fluorobiphenyl
 Concen: 112.258 ng
 RT: 9.16 min Scan# 1202
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion:172 Resp: 1090232

Ion	Ratio	Lower	Upper
172	100		
171	35.9	29.6	44.4
170	24.9	20.1	30.1




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#46
 1,1'-Biphenyl
 Concen: 54.577 ng
 RT: 9.26 min Scan# 1219
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:154 Resp: 700920

Ion	Ratio	Lower	Upper
154	100		
153	40.7	21.4	61.4
76	13.5	0.0	35.2



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF1








#47
 2-Chloronaphthalene
 Concen: 51.297 ng
 RT: 9.29 min Scan# 1224
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion	Ratio	Lower	Upper
162	100		
127	39.3	32.8	49.2
164	33.1	26.6	39.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 2-Nitroaniline
 Concen: 49.150 ng
 RT: 9.39 min Scan# 1242
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
65	100		
92	72.1	57.8	86.8
138	114.4	94.0	141.0



Abundance Ion 65.00 (64.70 to 65.70): BF1
 Ion 92.00 (91.70 to 92.70): BF1
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 53.211 ng

RT: 9.70 min Scan# 1294

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:152 Resp: 829783

Ion Ratio Lower Upper

152 100

151 20.8 17.2 25.8




153 14.3 11.8 17.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 54.922 ng

RT: 9.56 min Scan# 1270

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:163 Resp: 680271

Ion Ratio Lower Upper

163 100

194 3.8 3.0 4.6



164 10.4 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 53.996 ng
 RT: 9.63 min Scan# 1282
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion	Ratio	Lower	Upper
165	100		
63	59.2	38.2	57.2#
89	58.3	42.9	64.3




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF1
 Ion 89.00 (88.70 to 89.70): BF1






#52
 Acenaphthene
 Concen: 53.263 ng
 RT: 9.87 min Scan# 1323
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
154	100		
153	112.1	89.4	134.0
152	53.1	42.3	63.5



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 152.00 (151.70 to 152.70): E





#53

3-Nitroaniline

Concen: 12.500 ng

RT: 9.82 min Scan# 1314

Delta R.T. 0.00 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:138 Resp: 41573

Ion Ratio Lower Upper

138 100

108 10.6 8.0 12.0




92 110.2 91.0 136.6



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF1



#54

2,4-Dinitrophenol

Concen: 84.077 ng

RT: 9.93 min Scan# 1333

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:184 Resp: 115709

Ion Ratio Lower Upper

184 100

63 66.9 53.0 79.4



154 62.8 50.1 75.1




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 154.00 (153.70 to 154.70): E





#55

Dibenzofuran

Concen: 52.603 ng

RT: 10.05 min Scan# 1353

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:168 Resp: 731854

Ion Ratio Lower Upper

168 100

139 39.0 31.4 47.2




169 13.3 10.9 16.3



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 56.750 ng

RT: 9.99 min Scan# 1344

Delta R.T. 0.00 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:139 Resp: 120911

Ion Ratio Lower Upper

139 100

109 68.4 54.4 94.4



65 102.0 80.6 120.6




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1





#57

2,4-Dinitrotoluene

Concen: 52.289 ng

RT: 10.05 min Scan# 1353

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD

Tot Ion:165 Resp: 187393



Ion Ratio Lower Upper

165 100

63 57.7 39.6 59.4

89 79.0 62.2 93.2

182 2.7 2.1 3.1





Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#58

Fluorene

Concen: 57.057 ng

RT: 10.39 min Scan# 1411

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07



Tgt Ion:166 Resp: 610744

Ion Ratio Lower Upper

166 100

165 98.7 78.4 117.6


167 13.8 11.0 16.4



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
 2,3,4,6-Tetrachlorophenol
 Concen: 54.965 ng
 RT: 10.17 min Scan# 1375
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Tot Ion:232	Resp:	160011
Ion	Ratio	Lower Upper
232	100	
131	41.4	36.2 54.2
130	2.3	1.9 2.9
166	26.8	23.0 34.4



#60
 Diethylphthalate
 Concen: 55.796 ng
 RT: 10.26 min Scan# 1389
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:149	Resp:	645216
Ion	Ratio	Lower Upper
149	100	
177	23.8	18.8 28.2
150	12.5	10.2 15.4









#61
 4-Chlorophenyl-phenylether
 Concen: 56.457 ng
 RT: 10.37 min Scan# 1408
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Tot Ion:	204	Resp:	306057
Ion	Ratio	Lower	Upper
204	100		
206	32.3	26.4	39.6
141	62.3	46.4	69.6



#62
 4-Nitroaniline
 Concen: 37.661 ng
 RT: 10.42 min Scan# 1417
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:	138	Resp:	129448
Ion	Ratio	Lower	Upper
138	100		
92	50.2	29.0	69.0
108	73.2	54.7	94.7





#63

Azobenzene

Concen: 55.368 ng

RT: 10.53 min Scan# 1436

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD

Tot Ion: 77 Resp: 658562


Ion Ratio Lower Upper

77 100

182 24.2 4.0 44.0

105 22.2 2.0 42.0

51 26.2 7.5 47.5






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.32 min Scan# 1570

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion: 188 Resp: 339701

Ion Ratio Lower Upper

188 100

94 10.6 8.1 12.1



80 11.4 8.7 13.1




Abundance Ion 188.00 (187.70 to 188.70): BF1

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#65

4,6-Dinitro-2-methylphenol

Concen: 53.023 ng

RT: 10.46 min Scan# 1424

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:198 Resp: 95452

Ion Ratio Lower Upper

198 100

51 36.9 17.0 57.0




105 40.4 21.6 61.6



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BF1

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 47.103 ng

RT: 10.49 min Scan# 1429

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:169 Resp: 481608

Ion Ratio Lower Upper

169 100

168 65.8 53.8 80.6



167 37.1 29.6 44.4



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
 4-Bromophenyl-phenylether
 Concen: 51.161 ng
 RT: 10.86 min Scan# 1492
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion:248	Resp:	186047
Ion	Ratio	Lower Upper
248	100	
250	95.7	77.4 116.2
141	64.4	56.8 85.2




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#68
 Hexachlorobenzene
 Concen: 49.193 ng
 RT: 10.94 min Scan# 1505
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:284	Resp:	206857
Ion	Ratio	Lower Upper
284	100	
142	36.8	28.4 42.6
249	31.5	24.5 36.7



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E









#69
Atrazine
Concen: 34.763 ng
RT: 11.02 min Scan# 1519
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD

Tot Ion:200 Resp: 116932
Ion Ratio Lower Upper
200 100
173 25.6 6.8 46.8
215 47.5 27.5 67.5




Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E






#70
Pentachlorophenol
Concen: 79.174 ng
RT: 11.15 min Scan# 1540
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:266 Resp: 161635
Ion Ratio Lower Upper
266 100
268 63.3 48.8 73.2
264 61.9 50.6 76.0



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E





#71

Phenanthrene

Concen: 52.999 ng

RT: 11.35 min Scan# 1575

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:178 Resp: 920386

Ion Ratio Lower Upper

178 100

176 20.8 16.6 24.8




179 15.9 12.5 18.7



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#72

Anthracene

Concen: 54.886 ng

RT: 11.40 min Scan# 1583

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:178 Resp: 964649

Ion Ratio Lower Upper

178 100

176 20.3 15.7 23.5



179 16.2 13.0 19.6



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#73
 Carbazole
 Concen: 50.954 ng
 RT: 11.56 min Scan# 1610
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Tot Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.8	26.6
139	13.9	11.4	17.0



#74
 Di-n-butylphthalate
 Concen: 58.452 ng
 RT: 11.87 min Scan# 1664
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.6	8.0	12.0
104	5.0	4.5	6.7









#75
 Fluoranthene
 Concen: 57.341 ng
 RT: 12.54 min Scan# 1777
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion:202 Resp: 1042495
 Ion Ratio Lower Upper
 202 100
 101 12.4 0.0 31.2
 203 18.1 0.0 37.8




Abundance Ion 202.00 (201.70 to 202.70): E
 1500000 Ion 101.00 (100.70 to 101.70): E
 Ion 203.00 (202.70 to 203.70): E






#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 13.97 min Scan# 2020
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion:240 Resp: 294616
 Ion Ratio Lower Upper
 240 100
 120 11.4 10.7 16.1
 236 27.0 22.2 33.2



Abundance Ion 240.00 (239.70 to 240.70): E
 500000 Ion 120.00 (119.70 to 120.70): E
 Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 47.587 ng
RT: 12.77 min Scan# 1816
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD

Tot Ion:202 Resp: 1056830
Ion Ratio Lower Upper
202 100
200 23.0 18.0 27.0
203 18.3 14.1 21.1




Abundance Ion 202.00 (201.70 to 202.70): E
1500000 Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E






#79
Terphenyl-d14
Concen: 100.758 ng
RT: 12.90 min Scan# 1839
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:244 Resp: 1408756
Ion Ratio Lower Upper
244 100
212 7.2 5.8 8.8
122 10.9 9.0 13.4



Abundance Ion 244.00 (243.70 to 244.70): E
1500000 Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E









#80
Butylbenzylphthalate
Concen: 53.423 ng
RT: 13.37 min Scan# 1918
Delta R.T. -0.02 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD


Tot Ion	Ratio	Lower	Upper
149	100		
91	71.7	56.3	84.5
206	21.3	18.0	27.0



#81
Benzo(a)anthracene
Concen: 50.907 ng
RT: 13.96 min Scan# 2018
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.4	22.7	34.1
229	20.3	16.2	24.4









#82
 3,3'-Dichlorobenzidine
 Concen: 5.012 ng
 RT: 13.92 min Scan# 2011
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Tot Ion	Ratio	Lower	Upper
252	100		
254	63.7	50.9	76.3
126	12.2	9.8	14.8



#83
 Chrysene
 Concen: 52.764 ng
 RT: 13.99 min Scan# 2024
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.0	25.5	38.3
229	20.5	16.6	24.8





#84

Bis(2-ethylhexyl)phthalate

Concen: 59.991 ng

RT: 13.92 min Scan# 2012

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:149 Resp: 732364

Ion Ratio Lower Upper

149 100

167 27.4 21.6 32.4




279 3.8 3.0 4.4



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#85

Di-n-octyl phthalate

Concen: 64.012 ng

RT: 14.53 min Scan# 2116

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:149 Resp: 1241216

Ion Ratio Lower Upper

149 100

167 1.6 1.4 2.0



43 9.0 7.5 11.3




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1









#86
 Indeno(1,2,3-cd)pyrene
 Concen: 61.684 ng
 RT: 16.89 min Scan# 2517
 Delta R.T. -0.01 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07




Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD


Tot Ion	Ratio	Lower	Upper
276	100		
138	25.7	20.4	30.6
277	25.2	20.0	30.0



#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.42 min Scan# 2267
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	19.7	29.5
265	21.1	17.4	26.0









#88
Benzo(b)fluoranthene
Concen: 52.339 ng
RT: 15.00 min Scan# 2196
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07




Instrument :
BNA_F
ClientSampleId :
T1-L(0-2.5)MSD


Tot Ion:252 Resp: 921529
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.8
125 10.7 8.4 12.6



#89
Benzo(k)fluoranthene
Concen: 52.539 ng m
RT: 15.03 min Scan# 2200
Delta R.T. -0.01 min
Lab File: BF116447.D
Acq: 20 Aug 2019 23:07

Tgt Ion:252 Resp: 901015
Ion Ratio Lower Upper
252 100
253 23.0 17.8 26.8
125 11.7 7.7 11.5#





#90

Benzo(a)pyrene

Concen: 55.441 ng

RT: 15.36 min Scan# 2257

Delta R.T. -0.02 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07

Instrument :

BNA_F

ClientSampleId :

T1-L(0-2.5)MSD


Tot Ion:252 Resp: 872602

Ion Ratio Lower Upper

252 100

253 22.6 18.0 27.0




125 12.0 8.9 13.3



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Dibenzo(a,h)anthracene

Concen: 59.565 ng

RT: 16.89 min Scan# 2516

Delta R.T. -0.01 min

Lab File: BF116447.D

Acq: 20 Aug 2019 23:07


Tgt Ion:278 Resp: 811521

Ion Ratio Lower Upper

278 100

139 16.6 12.8 19.2



279 22.7 18.8 28.2




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E






#92
 Benzo(a,h,i)perylene
 Concen: 57.654 ng
 RT: 17.32 min Scan# 2590
 Delta R.T. -0.02 min
 Lab File: BF116447.D
 Acq: 20 Aug 2019 23:07

Instrument :
 BNA_F
 ClientSampleId :
 T1-L(0-2.5)MSD

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.4	27.6
138	22.3	17.0	25.6



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

