

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082118\
 Data File : BF108369.D
 Acq On : 22 Aug 2018 5:48
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 22 06:11:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 21 15:20:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	56	0.00
2	1,4-Dioxane	0.568	0.546	3.9	53	0.00
3	Pyridine	1.462	1.410	3.6	53	0.00
4	n-Nitrosodimethylamine	0.823	0.735	10.7	49#	0.00
5 S	2-Fluorophenol	1.105	1.137	-2.9	57	0.00
6	Aniline	1.997	1.994	0.2	56	0.00
7 S	Phenol-d6	1.468	1.482	-1.0	56	0.00
8	2-Chlorophenol	1.244	1.268	-1.9	56	0.00
9	Benzaldehyde	1.138	1.094	3.9	53	0.00
10 C	Phenol	1.518	1.517	0.1	56	0.00
11	bis(2-Chloroethyl)ether	1.228	1.238	-0.8	56	0.00
12	1,3-Dichlorobenzene	1.439	1.464	-1.7	57	0.00
13 C	1,4-Dichlorobenzene	1.445	1.469	-1.7	56	0.00
14	1,2-Dichlorobenzene	1.344	1.358	-1.0	56	0.00
15	Benzyl Alcohol	1.236	1.189	3.8	52	0.00
16	2,2'-oxybis(1-Chloropropane	2.529	2.363	6.6	51	0.00
17	2-Methylphenol	1.067	1.056	1.0	54	0.00
18	Hexachloroethane	0.515	0.472	8.3	50#	0.00
19 P	n-Nitroso-di-n-propylamine	0.993	0.941	5.2	53	0.00
20	3+4-Methylphenols	1.367	1.332	2.6	54	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	53	0.00
22	Acetophenone	0.468	0.475	-1.5	53	0.00
23 S	Nitrobenzene-d5	0.358	0.377	-5.3	55	0.00
24	Nitrobenzene	0.362	0.377	-4.1	53	0.00
25	Isophorone	0.683	0.680	0.4	52	0.00
26 C	2-Nitrophenol	0.137	0.154	-12.4	58	0.00
27	2,4-Dimethylphenol	0.303	0.310	-2.3	53	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.407	-2.0	54	0.00
29 C	2,4-Dichlorophenol	0.279	0.288	-3.2	53	0.00
30	1,2,4-Trichlorobenzene	0.308	0.310	-0.6	53	0.00
31	Naphthalene	0.910	0.947	-4.1	55	0.00
32	Benzoic acid	0.161	0.090	44.1#	29#	0.00
33	4-Chloroaniline	0.396	0.400	-1.0	52	0.00
34 C	Hexachlorobutadiene	0.195	0.206	-5.6	55	0.00
35	Caprolactam	0.087	0.084	3.4	48#	0.00
36 C	4-Chloro-3-methylphenol	0.301	0.295	2.0	50	0.00
37	2-Methylnaphthalene	0.644	0.632	1.9	52	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	45#	0.00
39	1,2,4,5-Tetrachlorobenzene	0.614	0.700	-14.0	51	0.00
40 P	Hexachlorocyclopentadiene	0.397	0.058	85.4#	6#	0.00
41 S	2,4,6-Tribromophenol	0.199	0.195	2.0	42#	0.00
42 C	2,4,6-Trichlorophenol	0.381	0.443	-16.3	51	0.00
43	2,4,5-Trichlorophenol	0.420	0.467	-11.2	48#	0.00
44 S	2-Fluorobiphenyl	1.246	1.453	-16.6	54	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082118\
 Data File : BF108369.D
 Acq On : 22 Aug 2018 5:48
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 22 06:11:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 21 15:20:21 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.475	1.664	-12.8	51	0.00
46	2-Chloronaphthalene	1.165	1.277	-9.6	50#	0.00
47	2-Nitroaniline	0.377	0.423	-12.2	48#	0.00
48	Acenaphthylene	1.843	1.945	-5.5	48#	0.00
49	Dimethylphthalate	1.393	1.537	-10.3	50	0.00
50	2,6-Dinitrotoluene	0.291	0.313	-7.6	47#	0.00
51 C	Acenaphthene	1.129	1.144	-1.3	46#	0.00
52	3-Nitroaniline	0.319	0.339	-6.3	46#	0.00
53 P	2,4-Dinitrophenol	0.092	0.032#	65.2#	15#	0.00
54	Dibenzofuran	1.635	1.673	-2.3	47#	0.00
55 P	4-Nitrophenol	0.241	0.209	13.3	38#	0.00
56	2,4-Dinitrotoluene	0.375	0.383	-2.1	43#	0.00
57	Fluorene	1.294	1.284	0.8	46#	0.00
58	2,3,4,6-Tetrachlorophenol	0.351	0.356	-1.4	43#	0.00
59	Diethylphthalate	1.349	1.458	-8.1	49#	0.00
60	4-Chlorophenyl-phenylether	0.674	0.692	-2.7	46#	0.00
61	4-Nitroaniline	0.315	0.312	1.0	43#	0.00
62	Azobenzene	1.393	1.371	1.6	44#	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	37#	0.00
64	4,6-Dinitro-2-methylphenol	0.069	0.034	50.7#	17#	0.00
65 c	n-Nitrosodiphenylamine	0.591	0.727	-23.0#	45#	0.00
66	4-Bromophenyl-phenylether	0.217	0.257	-18.4	43#	0.00
67	Hexachlorobenzene	0.229	0.247	-7.9	40#	0.00
68	Atrazine	0.198	0.228	-15.2	42#	0.00
69 C	Pentachlorophenol	0.109	0.096	11.9	31#	0.00
70	Phenanthrene	0.943	0.993	-5.3	40#	0.00
71	Anthracene	0.967	1.032	-6.7	40#	0.00
72	Carbazole	0.859	0.868	-1.0	37#	0.00
73	Di-n-butylphthalate	0.973	1.199	-23.2	45#	0.00
74 C	Fluoranthene	1.030	0.997	3.2	36#	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	36#	0.00
76	Benzidine	0.747	0.734	1.7	34#	0.00
77	Pyrene	1.266	1.237	2.3	36#	0.00
78 S	Terphenyl-d14	0.787	0.812	-3.2	38#	0.00
79	Butylbenzylphthalate	0.503	0.541	-7.6	37#	0.00
80	Benzo(a)anthracene	1.148	1.186	-3.3	37#	0.00
81	3,3'-Dichlorobenzidine	0.411	0.464	-12.9	39#	0.00
82	Chrysene	1.052	1.119	-6.4	39#	0.00
83	Bis(2-ethylhexyl)phthalate	0.681	0.740	-8.7	38#	0.00
84 c	Di-n-octyl phthalate	1.038	1.314	-26.6#	43#	0.00
85	Indeno(1,2,3-cd)pyrene	0.912	0.814	10.7	33#	0.00
86 I	Perylene-d12	1.000	1.000	0.0	37#	0.00
87	Benzo(b)fluoranthene	1.195	1.294	-8.3	38#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082118\
Data File : BF108369.D
Acq On : 22 Aug 2018 5:48
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 22 06:11:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF080818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 21 15:20:21 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.099	1.143	-4.0	40#	0.00
89 C	Benzo(a)pyrene	1.070	1.102	-3.0	37#	0.00
90	Dibenzo(a,h)anthracene	0.885	0.819	7.5	34#	0.00
91	Benzo(a,h,i)perylene	0.872	0.748	14.2	32#	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 2