

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082120\  
 Data File : BF121383.D  
 Acq On : 21 Aug 2020 23:23  
 Operator : JU/CG  
 Sample : L3734-02 2X  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A

Quant Time: Aug 22 00:35:51 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF081920.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Aug 20 04:56:10 2020  
 Response via : Initial Calibration


| Internal Standards             | R.T.  | QIon | Response | Conc      | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-----------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 6.70  | 152  | 159424   | 20.00     | ng    | 0.00     |
| 21) Naphthalene-d8             | 7.98  | 136  | 592758   | 20.00     | ng    | 0.00     |
| 39) Acenaphthene-d10           | 9.73  | 164  | 271708   | 20.00     | ng    | 0.00     |
| 64) Phenanthrene-d10           | 11.22 | 188  | 436879   | 20.00     | ng    | 0.00     |
| 76) Chrysene-d12               | 13.85 | 240  | 337500   | 20.00     | ng    | 0.00     |
| 86) Perylene-d12               | 15.27 | 264  | 349276   | 20.00     | ng    | 0.02     |
| System Monitoring Compounds    |       |      |          |           |       |          |
| 5) 2-Fluorophenol              | 5.30  | 112  | 178448   | 18.35     | ng    | 0.00     |
| 7) Phenol-d6                   | 6.34  | 99   | 479472   | 38.63     | ng    | 0.00     |
| 23) Nitrobenzene-d5            | 7.26  | 82   | 335839   | 32.04     | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol       | 10.52 | 330  | 21685    | 6.85      | ng    | 0.00     |
| 45) 2-Fluorobiphenyl           | 9.05  | 172  | 618843   | 36.32     | ng    | 0.00     |
| 79) Terphenyl-d14              | 12.80 | 244  | 456123   | 26.41     | ng    | 0.00     |
| Target Compounds               |       |      |          |           |       |          |
| 19) n-Nitroso-di-n-propylamine | 7.26  | 70   | 43766    | 6.557     | ng    | # 78     |
| 20) 3+4-Methylphenols          | 7.14  | 107  | 451      | Below Cal |       | # 44     |
| 25) Isophorone                 | 7.26  | 82   | 330719   | 17.339    | ng    | # 65     |
| 50) Dimethylphthalate          | 9.45  | 163  | 57299    | 2.899     | ng    | # 98     |
| 51) 2,6-Dinitrotoluene         | 9.73  | 165  | 36522    | 8.651     | ng    | # 30     |
| 55) Dibenzofuran               | 9.93  | 168  | 5570     | Below Cal |       | # 74     |
| 57) 2,4-Dinitrotoluene         | 9.73  | 165  | 36522    | 7.091     | ng    | # 25     |
| 58) Fluorene                   | 10.27 | 166  | 8977     | Below Cal |       | # 98     |
| 61) 4-Chlorophenyl-phenylether | 10.00 | 204  | 120      | Below Cal |       | # 30     |
| 71) Phenanthrene               | 11.23 | 178  | 86383    | 3.820     | ng    | # 99     |
| 75) Fluoranthene               | 12.42 | 202  | 146089   | 5.953     | ng    | # 97     |
| 78) Pylene                     | 12.65 | 202  | 130391   | 5.143     | ng    | # 99     |
| 81) Benzo(a)anthracene         | 13.84 | 228  | 68287    | 3.258     | ng    | # 92     |
| 83) Chrysene                   | 13.88 | 228  | 72889    | 3.400     | ng    | # 95     |
| 84) Bis(2-ethylhexyl)phthalate | 13.84 | 149  | 101603   | 7.931     | ng    | # 96     |
| 88) Benzo(b)fluoranthene       | 14.87 | 252  | 115011   | 5.097     | ng    | # 85     |
| 89) Benzo(k)fluoranthene       | 14.87 | 252  | 115011   | 5.790     | ng    | # 84     |
| 90) Benzo(a)pyrene             | 15.21 | 252  | 56297    | 2.888     | ng    | # 78     |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082120\  
Data File : BF121383.D  
Acq On : 21 Aug 2020 23:23  
Operator : JU/CG  
Sample : L3734-02 2X  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
EROC-BF-1A

Quant Time: Aug 22 00:35:51 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF081920.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Aug 20 04:56:10 2020  
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4  
 Concen: 20.000 ng  
 RT: 6.70 min Scan# 784  
 Delta R.T. 0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 152     | 100   |       |       |
| 150     | 155.4 | 124.8 | 187.2 |
| 115     | 56.7  | 46.1  | 69.1  |




Abundance Ion 152.00 (151.70 to 152.70): E  
 Ion 150.00 (149.70 to 150.70): E  
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol  
 Concen: 18.354 ng  
 RT: 5.30 min Scan# 547  
 Delta R.T. 0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 112     | 100   |       |       |
| 64      | 55.1  | 43.5  | 65.3  |
| 63      | 27.6  | 22.0  | 33.0  |



Abundance Ion 112.00 (111.70 to 112.70): E  
 Ion 64.00 (63.70 to 64.70): BF1  
 Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 38.631 ng

RT: 6.34 min Scan# 723

Delta R.T. 0.00 min

Lab File: BF121383.D

Acq: 21 Aug 2020 23:23

Instrument :

BNA\_F

ClientSampleId :

EROC-BF-1A





Tot Ion: 99 Resp: 479472

Ion Ratio Lower Upper

99 100

42 17.1 13.5 20.3

71 30.3 25.2 37.8



#19

n-Nitroso-di-n-propylamine

Concen: 6.557 ng

RT: 7.26 min Scan# 880

Delta R.T. 0.15 min

Lab File: BF121383.D

Acq: 21 Aug 2020 23:23

Tgt Ion: 70 Resp: 43766




Ion Ratio Lower Upper


70 100

42 46.4 45.2 67.8

101 0.0 7.7 11.5#

130 2.3 17.6 26.4#









#20  
3+4-Methylphenols  
Concen: Below Cal  
RT: 7.14 min Scan# 859  
Delta R.T. 0.02 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23




Instrument :  
BNA\_F  
ClientSampleId :  
EROC-BF-1A


Tot Ion:107 Resp: 451  
Ion Ratio Lower Upper  
107 100  
108 70.2 72.6 112.6#  
77 0.0 63.8 103.8#  
79 0.0 9.1 49.1#



#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 7.98 min Scan# 1002  
Delta R.T. 0.01 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23

Tgt Ion:136 Resp: 592758  
Ion Ratio Lower Upper  
136 100  
137 11.2 8.9 13.3  
54 7.3 6.1 9.1  
68 6.0 5.0 7.4









#23  
 Nitrobenzene-d5  
 Concen: 32.043 ng  
 RT: 7.26 min Scan# 880  
 Delta R.T. 0.00 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A

| Tot Ion   | 82    | Resp  | 335839 |
|-----------|-------|-------|--------|
| Ion Ratio | Lower | Upper |        |
| 82        | 100   |       |        |
| 128       | 49.4  | 37.9  | 56.9   |
| 54        | 47.0  | 39.0  | 58.4   |




Abundance Ion 82.00 (81.70 to 82.70): BF1  
 Ion 128.00 (127.70 to 128.70): BF1  
 Ion 54.00 (53.70 to 54.70): BF1






#25  
 Isophorone  
 Concen: 17.339 ng  
 RT: 7.26 min Scan# 880  
 Delta R.T. -0.26 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

| Tgt Ion   | 82    | Resp  | 330719 |
|-----------|-------|-------|--------|
| Ion Ratio | Lower | Upper |        |
| 82        | 100   |       |        |
| 95        | 0.1   | 5.3   | 7.9#   |
| 138       | 0.0   | 14.6  | 21.8#  |



Abundance Ion 82.00 (81.70 to 82.70): BF1  
 Ion 95.00 (94.70 to 95.70): BF1  
 Ion 138.00 (137.70 to 138.70): BF1









#39  
 Acenaphthene-d10  
 Concen: 20.000 ng  
 RT: 9.73 min Scan# 1299  
 Delta R.T. 0.00 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23




Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 164     | 100   |       |       |
| 162     | 102.1 | 81.7  | 122.5 |
| 160     | 45.8  | 37.0  | 55.4  |



#42  
 2,4,6-Tribromophenol  
 Concen: 6.849 ng  
 RT: 10.52 min Scan# 1434  
 Delta R.T. 0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 330     | 100   |       |       |
| 332     | 97.0  | 76.9  | 115.3 |
| 141     | 36.8  | 22.8  | 34.2  |









#45  
 2-Fluorobiphenyl  
 Concen: 36.319 ng  
 RT: 9.05 min Scan# 1184  
 Delta R.T. 0.00 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 172     | 100   |       |       |
| 171     | 35.1  | 29.0  | 43.6  |
| 170     | 23.7  | 19.4  | 29.2  |




Abundance Ion 172.00 (171.70 to 172.70): E  
 Ion 171.00 (170.70 to 171.70): E  
 Ion 170.00 (169.70 to 170.70): E






#50  
 Dimethylphthalate  
 Concen: 2.899 ng  
 RT: 9.45 min Scan# 1252  
 Delta R.T. -0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 163     | 100   |       |       |
| 194     | 4.1   | 3.0   | 4.6   |
| 164     | 9.4   | 8.4   | 12.6  |



Abundance Ion 163.00 (162.70 to 163.70): E  
 Ion 194.00 (193.70 to 194.70): E  
 Ion 164.00 (163.70 to 164.70): E





#51

2,6-Dinitrotoluene

Concen: 8.651 ng

RT: 9.73 min Scan# 1299

Delta R.T. 0.21 min

Lab File: BF121383.D

Acq: 21 Aug 2020 23:23

Instrument :

BNA\_F

ClientSampleId :

EROC-BF-1A


Tot Ion:165 Resp: 36522

Ion Ratio Lower Upper

165 100

63 1.4 38.5 57.7#




89 1.6 41.2 61.8#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1



#55

Dibenzofuran

Concen: Below Cal

RT: 9.93 min Scan# 1334

Delta R.T. 0.00 min

Lab File: BF121383.D

Acq: 21 Aug 2020 23:23


Tgt Ion:168 Resp: 5570

Ion Ratio Lower Upper

168 100

139 46.4 29.1 43.7#



169 34.6 11.0 16.6#




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E









#57  
 2,4-Dinitrotoluene  
 Concen: 7.091 ng  
 RT: 9.73 min Scan# 1299  
 Delta R.T. -0.20 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23




Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A


|          |       |       |       |
|----------|-------|-------|-------|
| Tot Ion: | 165   | Resp: | 36522 |
| Ion      | Ratio | Lower | Upper |
| 165      | 100   |       |       |
| 63       | 1.4   | 37.5  | 56.3# |
| 89       | 1.6   | 54.2  | 81.4# |
| 182      | 0.0   | 2.2   | 3.4#  |



#58  
 Fluorene  
 Concen: Below Cal  
 RT: 10.27 min Scan# 1392  
 Delta R.T. 0.00 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 166   | Resp: | 8977  |
| Ion      | Ratio | Lower | Upper |
| 166      | 100   |       |       |
| 165      | 94.9  | 75.8  | 113.8 |
| 167      | 19.1  | 11.0  | 16.4# |









#61  
4-Chlorophenyl-phenylether  
Concen: Below Cal  
RT: 10.00 min Scan# 1345  
Delta R.T. -0.27 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23




Instrument :  
BNA\_F  
ClientSampleId :  
EROC-BF-1A


Tot Ion:204 Resp: 120  
Ion Ratio Lower Upper  
204 100  
206 0.0 25.9 38.9#  
141 0.0 45.5 68.3#



#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 11.22 min Scan# 1552  
Delta R.T. 0.01 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23

Tgt Ion:188 Resp: 436879  
Ion Ratio Lower Upper  
188 100  
94 8.3 7.2 10.8  
80 8.8 7.4 11.0









#71  
Phenanthrene  
Concen: 3.820 ng  
RT: 11.23 min Scan# 1555  
Delta R.T. 0.00 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23

Instrument :  
BNA\_F  
ClientSampleId :  
EROC-BF-1A

Tot Ion:178 Resp: 86383  
Ion Ratio Lower Upper  
178 100  
176 19.1 16.0 24.0  
179 16.4 12.8 19.2




Abundance Ion 178.00 (177.70 to 178.70): E  
Ion 176.00 (175.70 to 176.70): E  
Ion 179.00 (178.70 to 179.70): E






#75  
Fluoranthene  
Concen: 5.953 ng  
RT: 12.42 min Scan# 1757  
Delta R.T. 0.00 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23

Tgt Ion:202 Resp: 146089  
Ion Ratio Lower Upper  
202 100  
101 12.7 0.0 30.6  
203 17.8 0.0 37.5



Abundance Ion 202.00 (201.70 to 202.70): E  
Ion 101.00 (100.70 to 101.70): E  
Ion 203.00 (202.70 to 203.70): E





#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.85 min Scan# 2000

Delta R.T. 0.01 min

Lab File: BF121383.D

Acq: 21 Aug 2020 23:23

Instrument :

BNA\_F

ClientSampleId :

EROC-BF-1A


Tot Ion:240 Resp: 337500

Ion Ratio Lower Upper

240 100

120 12.5 9.6 14.4




236 25.8 20.6 31.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Pyrene

Concen: 5.143 ng

RT: 12.65 min Scan# 1796

Delta R.T. 0.00 min

Lab File: BF121383.D

Acq: 21 Aug 2020 23:23


Tgt Ion:202 Resp: 130391

Ion Ratio Lower Upper

202 100

200 20.1 16.6 24.8



203 18.1 14.2 21.2




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#79  
 Terphenyl-d14  
 Concen: 26.410 ng  
 RT: 12.80 min Scan# 1821  
 Delta R.T. 0.00 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23




Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A


Tot Ion:244 Resp: 456123  
 Ion Ratio Lower Upper  
 244 100  
 212 6.8 5.4 8.0  
 122 11.3 7.8 11.8



#81  
 Benzo(a)anthracene  
 Concen: 3.258 ng  
 RT: 13.84 min Scan# 1999  
 Delta R.T. 0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

Tgt Ion:228 Resp: 68287  
 Ion Ratio Lower Upper  
 228 100  
 226 31.1 21.8 32.6  
 229 24.3 15.8 23.8#









#83  
 Chrysene  
 Concen: 3.400 ng  
 RT: 13.88 min Scan# 2005  
 Delta R.T. 0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23




Instrument :  
 BNA\_F  
 ClientSampleId :  
 EROC-BF-1A


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 228     | 100   |       |       |
| 226     | 31.3  | 24.6  | 36.8  |
| 229     | 25.0  | 15.9  | 23.9  |



#84  
 Bis(2-ethylhexyl)phthalate  
 Concen: 7.931 ng  
 RT: 13.84 min Scan# 1998  
 Delta R.T. 0.01 min  
 Lab File: BF121383.D  
 Acq: 21 Aug 2020 23:23

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 149     | 100   |       |       |
| 167     | 29.0  | 21.5  | 32.3  |
| 279     | 4.8   | 2.9   | 4.3   |





#86

Pervlene-d12

Concen: 20.000 ng

RT: 15.27 min Scan# 2242

Delta R.T. 0.02 min





Lab File: BF121383.D

Acq: 21 Aug 2020 23:23

Instrument :  
BNA\_F  
ClientSampleId :  
EROC-BF-1A

Tot Ion:264 Resp: 349276

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 264 | 100   |       |       |
| 260 | 23.7  | 19.5  | 29.3  |
| 265 | 22.1  | 17.1  | 25.7  |



#88

Benzo(b)fluoranthene

Concen: 5.097 ng

RT: 14.87 min Scan# 2173




Delta R.T. 0.02 min


Lab File: BF121383.D

Acq: 21 Aug 2020 23:23

Tgt Ion:252 Resp: 115011

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 252 | 100   |       |       |
| 253 | 27.3  | 18.0  | 27.0# |
| 125 | 20.8  | 8.6   | 12.8# |









#89  
Benzo(k)fluoranthene  
Concen: 5.790 ng  
RT: 14.87 min Scan# 2173  
Delta R.T. -0.01 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23

Instrument :  
BNA\_F  
ClientSampleId :  
EROC-BF-1A

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 253     | 27.3  | 17.9  | 26.9# |
| 125     | 20.8  | 8.6   | 12.8# |



#90  
Benzo(a)pyrene  
Concen: 2.888 ng  
RT: 15.21 min Scan# 2231  
Delta R.T. 0.02 min  
Lab File: BF121383.D  
Acq: 21 Aug 2020 23:23

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 252     | 100   |       |       |
| 253     | 28.3  | 18.0  | 27.0# |
| 125     | 27.1  | 9.4   | 14.0# |

