

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082120\
 Data File : BF121382.D
 Acq On : 21 Aug 2020 22:53
 Operator : JU/CG
 Sample : L3734-05 2X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EROC-BF-1D

Manual Integrations
 APPROVED

Sohil
 8/24/2020 1:50:12 PM

Quant Time: Aug 22 02:50:19 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF081920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 20 04:56:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	169496	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	635890	20.00	ng	0.00
39) Acenaphthene-d10	9.73	164	305382	20.00	ng	0.00
64) Phenanthrene-d10	11.21	188	498362	20.00	ng	0.00
76) Chrysene-d12	13.85	240	344898	20.00	ng	0.00
86) Perylene-d12	15.27	264	380841	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	292500	28.30	ng	0.00
7) Phenol-d6	6.34	99	444823	33.71	ng	0.00
23) Nitrobenzene-d5	7.26	82	289180	25.72	ng	0.00
42) 2,4,6-Tribromophenol	10.52	330	42765	12.02	ng	0.00
45) 2-Fluorobiphenyl	9.05	172	548762	25.44	ng	0.00
79) Terphenyl-d14	12.80	244	403890	22.88	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.45	163	53264	2.397	ng	97
71) Phenanthrene	11.23	178	90071	3.492	ng	98
75) Fluoranthene	12.42	202	147765	5.278	ng	98
78) Pyrene	12.65	202	134997	5.211	ng	98
81) Benzo(a)anthracene	13.84	228	69228	3.232	ng	93
83) Chrysene	13.87	228	63956m	2.919	ng	
88) Benzo(b)fluoranthene	14.87	252	90298m	3.670	ng	
90) Benzo(a)pyrene	15.21	252	58074	2.732	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF082120\
Data File : BF121382.D
Acq On : 21 Aug 2020 22:53
Operator : JU/CG
Sample : L3734-05 2X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EROC-BF-1D

Manual Integrations
APPROVED

Sohil
8/24/2020 1:50:12 PM

Quant Time: Aug 22 02:50:19 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF081920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 20 04:56:10 2020
Response via : Initial Calibration

