



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082123\
 Data File : BF134859.D
 Acq On : 21 Aug 2023 10:21
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Aug 04 09:29:15 2023






AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.8	73792	PASS
68	69	0.00	2	1.9	1355	PASS
69	198	0.00	100	32.4	70672	PASS
70	69	0.00	2	0.6	446	PASS
127	198	10	80	43.5	95024	PASS
197	198	0.00	2	0.5	1144	PASS
198	198	100	100	100.0	218349	PASS
199	198	5	9	6.4	13959	PASS
275	198	10	60	29.6	64579	PASS
365	198	1	100	3.9	8546	PASS
441	198	0.01	100	18.6	40603	PASS
442	442	50	100	100.0	268181	PASS
443	442	15	24	19.4	51920	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082123\
Data File : BF134859.D
Acq On : 21 Aug 2023 10:21
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 21 12:03:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration



TIC: BF134859.D\data.ms

(70) Pentachlorophenol (C)

11.122min (-0.006) 259437.45 ng m




response 226540

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.10	64.02
264.00	59.70	61.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF082123\
Data File : BF134859.D
Acq On : 21 Aug 2023 10:21
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 21 12:03:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 04 09:29:15 2023
Response via : Initial Calibration



TIC: BF134859.D\data.ms

(77) Benzidine

12.663min (+ 0.000) 95811.01 ng

response 840261

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.09
183.00	10.80	12.18
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/21/2023	BNA_F	<u>BF134859.D</u>
Compound Name	Response	Retention Time
DDT	398199	13.486
DDD	29218	13.186
DDE	503	12.851
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
29721	427920	6.95

Instrument :
BNA_F
ClientSampleId :
DFTPP